

	Hyperecho Monitoring	Harmonic Dip Detection	
		All Exposures	$I_{\text{SPTA}} = 1910 \text{ W/cm}^2$
True Negatives	6.4% (5 of 78)	32.5% (69 of 212)	20.0% (3 of 15)
True Positives	42.3% (33 of 78)	42.5% (90 of 212)	73.3% (11 of 15)
False Negatives	43.6% (34 of 78)	13.7% (29 of 212)	6.7% (1 of 15)
False Positives	7.7% (6 of 78)	11.3% (24 of 212)	0% (0 of 15)
Incidence of Lesioning	85.9% (67 of 78)	56.1% (119 of 212)	80.0% (12 of 15)
Accuracy	48.7% (38 of 78)	75.0% (159 of 212)	93.3% (14 of 15)
Sensitivity	49.3% (33 of 67)	75.6% (90 of 119)	91.7% (11 of 12)
Specificity	45.5% (5 of 11)	74.2% (69 of 93)	100% (3 of 3)