

Review

Open Access

Jennifer A. Hirst*, Julie H. McLellan, Christopher P. Price, Emma English, Benjamin G. Feakins, Richard J. Stevens and Andrew J. Farmer

Performance of point-of-care HbA_{1c} test devices: implications for use in clinical practice – a systematic review and meta-analysis

DOI 10.1515/cclm-2016-0303

Received April 12, 2016; accepted July 19, 2016

Abstract

Background: Point-of-care (POC) devices could be used to measure hemoglobin A1c (HbA_{1c}) in the doctors' office, allowing immediate feedback of results to patients. Reports have raised concerns about the analytical performance of some of these devices. We carried out a systematic review and meta-analysis using a novel approach to compare the accuracy and precision of POC HbA_{1c} devices.

Methods: Medline, Embase and Web of Science databases were searched in June 2015 for published reports comparing POC HbA_{1c} devices with laboratory methods. Two reviewers screened articles and extracted data on bias, precision and diagnostic accuracy. Mean bias and variability between the POC and laboratory test were combined in a meta-analysis. Study quality was assessed using the QUADAS2 tool.

Results: Two researchers independently reviewed 1739 records for eligibility. Sixty-one studies were included in the meta-analysis of mean bias. Devices evaluated were A1cgear, A1cNow, Afinion, B-analyst, Clover, Cobas b101, DCA 2000/Vantage, HemoCue, Innovastar, Nycocard,

Quo-Lab, Quo-Test and SDA1cCare. Nine devices had a negative mean bias which was significant for three devices. There was substantial variability in bias within devices. There was no difference in bias between clinical or laboratory operators in two devices.

Conclusions: This is the first meta-analysis to directly compare performance of POC HbA_{1c} devices. Use of a device with a mean negative bias compared to a laboratory method may lead to higher levels of glycemia and a lower risk of hypoglycaemia. The implications of this on clinical decision-making and patient outcomes now need to be tested in a randomized trial.

Keywords: diabetes; HbA_{1c}; instrument performance; meta-analysis; point-of-care testing; systematic review.

Introduction

Regular monitoring of glycated hemoglobin subfraction A1c (HbA_{1c}) in people with diabetes and treatment with glucose-lowering medications to improve glycaemic control can reduce the risk of developing complications [1]. In 2011, a World Health Organization consultation concluded that HbA_{1c} at a threshold of 6.5% (48 mmol/mol) can be used as a diagnostic test for diabetes [2]. HbA_{1c} monitoring often requires the patient to attend the health center twice: once to have blood taken and then returning to get test results and receive adjustments to medication.

Point-of-care (POC) analysers are bench-top instruments that use a finger-prick blood sample and are designed for use in a treatment room or at the bed-side. They provide a test result within a few minutes allowing clinical decisions and medication changes to take place immediately. The suitability of many of these devices for the accurate measurement of HbA_{1c} has been questioned, with some POC HbA_{1c} test devices reported not to meet accepted accuracy and precision criteria [3]. Ideal

*Corresponding author: Jennifer A. Hirst, University of Oxford, Nuffield Department of Primary Care Health Sciences, Radcliffe Observatory Quarter Woodstock Road Oxford OX2 6GG, United Kingdom of Great Britain and Northern Ireland, E-mail: jennifer.hirst@phc.ox.ac.uk

Julie H. McLellan, Benjamin G. Feakins, Richard J. Stevens and Andrew J. Farmer: University of Oxford, Nuffield Department of Primary Care Health Sciences, United Kingdom of Great Britain and Northern Ireland

Christopher P. Price: University of Oxford, Nuffield Department of Primary Care Health Sciences, United Kingdom of Great Britain and Northern Ireland

Emma English: University of East Anglia, School of Health Science, Norwich, United Kingdom of Great Britain and Northern Ireland

imprecision goals for HbA_{1c} should be coefficient of variation (CV) of <2% for HbA_{1c} reported in % units (or <3% in SI units, mmol/mol) [4–6].

Most evaluations of POC HbA_{1c} devices have taken place in laboratory settings [7, 8]; fewer studies have assessed device performance in a POC setting or with clinicians performing the tests [9, 10]. The only published review that has attempted to combine data from accuracy studies identified five studies covering three devices and compared correlation coefficients [11]. Systematically reporting and pooling data estimates of bias and precision between POC HbA_{1c} devices and laboratory measurements would enable end users to assess which analysers best meet their analytical performance needs. This may be of particular importance for clinicians in primary care settings where much of the management of diabetes patients takes place. The comparison of accuracy between devices over the entire therapeutic range would need to be carried out by combining data on measurement error (bias) between POC and laboratory tests [12].

The aim of this study was to compare accuracy and precision of POC HbA_{1c} devices with the local laboratory method based on data from published studies and discuss the clinical implications of the findings.

Materials and methods

Data sources and searches

The protocol was registered on the PROSPERO database (Registration number CRD42013006678). Searches of MEDLINE (1950–2015), EMBASE (1980–2015) and Web of Science databases (1900–2015) were carried out in June 2015 using search terms for glycated hemoglobin and POC device names, and Medical Subject Headings for POC systems (Medline search strategy in Table 1, Supplementary Data). Hand searches of original articles, reviews and conference abstracts were also carried out. Reports in all languages published in peer-reviewed journals were considered for inclusion.

Study selection

Title and abstract review of retrieved references and screening of full texts for eligibility were performed by two reviewers (JH and JM).

Included studies performed both a POC HbA_{1c} test (defined as any instrument designed to provide a rapid

quantitative measurement of HbA_{1c} using capillary blood at the point of care) and laboratory-based methods on each sample and reported mean difference between POC and laboratory HbA_{1c} (referred to hereafter as comparator method). The mean difference (or mean bias) from included studies was obtained using methods described by Bland and Altman for assessing agreement between two methods [13]. Studies were excluded if the study did not evaluate the accuracy of HbA_{1c} POC devices, if there was no comparator test, if the article was not peer reviewed, data could not be extracted, patients had hemoglobinopathies or if the POC analysers evaluated were not commercially available on 30 June 2015.

Data extraction and quality assessment

Data were extracted by two reviewers on the following: sample type (capillary or venous blood), operator (clinical or laboratory operator), setting (point of care, laboratory or other), number of participants, mean HbA_{1c} (in either SI, mmol/mol or % units) and mean (\pm standard deviation, SD) bias, defined as mean of the difference between POC and comparator method (POC-comparator) for each HbA_{1c} sample measured. Reviewers were not blinded to any aspects of the studies during data extraction. Disagreements were resolved by discussion or, if consensus could not be reached, discussion with a third reviewer. If duplicate POC tests were performed using different reagent lot numbers, then both comparisons were included in the meta-analysis.

Instrument-specific data were extracted: POC device make and model, make and model of comparator method used and whether the comparator method was carried out in a NGSP (National Glycohemoglobin Standardization Program) or IFCC (International Federation for Clinical Chemistry and Laboratory Medicine) reference laboratory.

Study quality and risk of bias in data collection and methodology were graded using the QUADAS 2 risk of bias table [14]. Funding by device manufacturer and whether the range in HbA_{1c} was sufficiently broad (defined as 5%–11% or 31–97 mmol/mol) were recorded. All studies deemed to be at high risk of bias for at least one criterion were excluded in a sensitivity analysis.

Data on the diagnostic accuracy of devices at a threshold of 6.5% (48 mmol/mol) was requested from study authors. Where data were not available, diagnostic accuracy was estimated from correlation and Bland-Altman graphs by counting the number of points in graph quadrants representative of diagnoses that would be truly negative, falsely positive, truly positive and falsely negative.

Data on imprecision recorded as CV was extracted in % units where reported. We combined both within and between-run precision at low (<6%, 42 mmol/mol), medium (6%–8%, 42–64 mmol/mol), and high (\geq 8%, 64 mmol/mol) levels of HbA_{1c} as well as total CV regardless of whether the samples were control samples or patient samples. The mean imprecision was combined by taking the average for each device at each level of HbA_{1c}. Imprecision data on comparator instruments was also combined where this was reported.

Data synthesis and analysis

All analyses were carried out using Stata 12.0SE (Stata-Corp, College Station, TX, USA). Data on mean bias between POC and laboratory HbA_{1c} and SD mean bias were combined using random effects DerSimonian and Laird meta-analysis [15] as recommended for meta-analyses of method comparison studies with considerable heterogeneity [16]. Data for each POC instrument were separately pooled. The following estimates were made: median difference was approximated to mean bias; if SD or standard error were not reported, then these were imputed averaging the SD from studies of the same device [17]. The SD of the mean bias and the standard error of the SD [18] were pooled in a separate meta-analysis to explore the variability in bias within each device. In a descriptive analysis, we reported bias and SD for each device without pooling. In meta-analyses, we summarized mean bias for each device and SD for each device with 95% confidence intervals. Because performance may vary between different clinical settings, we also show 95% prediction intervals (using the “*rfdist*” command in Stata) to display the estimated uncertainty about the mean bias, and SD, in any future study [19]. Studies in which estimates were made, those at high risk of bias and those in which HbA_{1c} measurement were carried out more than 24 h apart were excluded in sensitivity analyses. Sensitivity analyses including only reports which used NGSP or IFCC reference laboratories for the comparator HbA_{1c} and those published in 2014–2015 were also conducted. Subgroup analyses were performed where possible to explore differences between operator (clinical or laboratory), setting (POC or laboratory), adults or children or blood sample type (venous or capillary) within a single device. Meta-regression was used where there were sufficient studies to establish whether different settings (POC or laboratory) or study publication year significantly affected the mean bias. A post hoc meta-regression was carried out to explore the effect of publication year on SD of the bias in the DCA device.

Where there were sufficient data, a bivariate random-effects method was used to estimate average sensitivity and specificity using the “*metandi*” command in Stata [20]. The average sensitivity and specificity was calculated and hierarchical summary receiver-operating characteristic curves were plotted to display the diagnostic accuracy across studies within a device.

Results

Searches identified 1739 records, a further six records were identified from scanning reference lists. There were 435 duplicates, and 1063 records were excluded after title and abstract review and 181 after full text review, leaving 66 studies; a further six studies were excluded because they included people with hemoglobin variants or data could not be extracted. Sixty studies were included in the analysis of mean bias from 13 devices (Supplementary Data, Figure 1). Forty-five included papers were full text articles, the remainder were conference abstracts, letters or comments. Twenty-five studies were carried out in Europe, 16 in the Americas, nine in Asia, eight in Australia and two unknown. Four articles were non-English and required translation [21–24]. Seven studies were conducted in children [25–31]. Some studies reported results from multiple devices or multiple lot numbers giving 105 comparisons in the meta-analysis. Characteristics of all included studies are shown in Table 1.

Devices compared in the meta-analysis were A1cgear (n=1), A1cNow (n=17), Afinion (n=12), B-analyst (n=5), Clover (n=2), Cobas b101 (n=5), DCA 2000/Vantage (n=39), HemoCue (n=2), Innovastar (n=5), Nycocard (n=6), Quo-Lab (n=3), Quo-Test (n=7), and SDA1cCare (n=1). The majority of studies reported bias in HbA_{1c} in % units; only two studies reporting in mmol/mol [34, 62]. Nineteen studies were carried out in a laboratory setting by laboratory staff, 21 studies were carried out at point of care and in 12 studies, a clinician operated the POC device.

Mean bias and 2SD range of difference in HbA_{1c} between each POC device and comparator method (Figure 1) shows that there is a large variability in bias for most devices. Pooling data on mean bias (Figure 2) in a meta-analysis shows that nine devices have a negative mean bias; this was significant for three devices (Quo-Lab, DCA and InnovaStar). Four devices had a positive mean bias which was significant for two devices (B-analyst and SDA1cCare). Heterogeneity between studies was large for all devices, with *I*² ranging from 67.5% to 99.5%. Although the pooled mean bias for some devices

Table 1: Table of included studies.

Author	Year	POC device(s)	Comparator method	Comparator instrument and manufacturer	POC sample	Comparator sample	Operator	Setting	Number of patients	Mean HbA _{1c} %
Affret et al. [21]	2015	A1cNow	Immunoassay	Vitros 5.1 FS, Ortho Clinical Diagnostics	Capillary	Venous	–	POC	55	7.3
Arsie et al. [32]	2000	DCA 2000	Ion exchange	Diamat, Bio-Rad Laboratories Ltd ^a	Venous	Venous	Laboratory	Laboratory	117	7.2
Azevedo et al. [33]	2011	DCA 2000 Quo-Test	Ns- HPLC	^a	Capillary	Venous	–	–	52	8.1
Berry et al. [34]	2014	Cobas b101, HemoCue, DCA Vantage, Afinion	Ion exchange	Biorad D10, Bio-Rad Laboratories Ltd	–	–	–	–	17–49	–
Boz et al. [22]	1997	DCA 2000	Ns- HPLC	^a	Capillary	Venous	–	POC	100	7.9
Cagliero et al. [35]	1999	DCA 2000	Ns- HPLC	^a	Capillary	Venous	Clinical	POC	264	8.6
Carrera Font et al. [23]	2011	Afinion	Ion exchange	Menarini HA 8160, A Menarini diagnostics Ltd	Capillary	Venous	Clinical	POC	94	7.26
Chan et al. [25]	2014	DCA Vantage	Ion exchange	Biorad Variant II, Bio-Rad Laboratories Ltd	–	–	–	–	72	5.7
Clark et al. [36]	2005	A1cNow	Ion exchange	Tosoh 2.2, Tosoh Bioscience Ltd	Capillary and venous	Venous	Laboratory	Laboratory	499	–
Criel et al. [37]	2015	Cobas b101, Afinion, B-analyst	Ion exchange	Adams Arkray HA8160, Menarini	–	–	–	–	40	–
Dupuy et al. [31]	2013	DCA Vantage	Ion exchange	Menarini HA 8140, A Menarini diagnostics Ltd	Venous	Venous	–	–	124	–
Dupuy et al. [38]	2014	Afinion	Ion exchange	Menarini HA 8140, A Menarini diagnostics Ltd	Venous	Venous	Laboratory	Laboratory	155	–
Ejilemele et al. [39]	2015	A1c gear	Ion exchange	Biorad Variant II, Bio-Rad Laboratories Ltd	Venous	Venous	Laboratory	Laboratory	120	8.1
ElArabi et al. [26]	2013	DCA Vantage	Ion exchange	Menarini HA 8160, A Menarini diagnostics Ltd	Capillary	–	Clinical	POC	120	–
Estevez et al. [40]	2014	DCA Vantage	Ion exchange	Menarini HA 8160, A Menarini diagnostics Ltd	Venous	–	–	–	42	–
Ginde et al. [41]	2008	A1cNow	Ns-HPLC	^a	Capillary	Venous	Laboratory	POC	161	5.8
Greaves et al. [27]	2005	DCA 2000	Ion exchange	Bio-Rad D10, Bio-Rad Laboratories Ltd	Capillary and venous	Capillary and venous	–	POC and laboratory	228	–
Guerci et al. [42]	1997	DCA 2000	Ion exchange	Diamat, Bio-Rad Laboratories Ltd ^a	Capillary	Venous	–	POC	1016	–
Harris et al. [43]	2000	DCA 2000	Immunoassay	LX20 Beckman Coulter	Capillary	Venous	Clinical	POC	102	–
Hawkins [44]	2003	DCA 2000	Ion exchange	Diastat, Bio-Rad Laboratories Ltd ^a	Venous	Venous	Laboratory	Laboratory	110	6.8
Heng et al [45]	2015	Cobas b101	Immunoassay	Roche Tina-quant A1c-gen3, Roche Diagnostics Ltd	Capillary	Venous	–	–	47	–
Holmes et al. [46]	2008	DCA 2000	Ion exchange	Biorad Variant II, Bio-Rad Laboratories Ltd	Venous	Venous	Laboratory	Laboratory	24, 45, 21	–
Jiang et al. [47]	2014	A1cNow	Ion exchange	Biorad Variant II, Bio-Rad Laboratories Ltd	Capillary	Venous	Clinical	POC	1618	7.0
Jou et al. [48]	2013	A1cNow	Ion exchange	Tosoh G8, Tosoh Bioscience Ltd	Venous	Venous	–	–	57	–
Karami and Baradaran [49]	2014	Nycocard	Affinity	Kanuer ^a	Venous	Venous	Laboratory	Laboratory	58	5.8
Knaebel et al [50]	2013	A1cNow	Ion exchange	Tosoh G8, Tosoh Bioscience Ltd	Venous	Venous	–	–	40	–
Leal and Soto-Rowen [51]	2009	A1cNow	Ns	^a	Capillary	Venous	Pharmacist	POC	47	8.1
Leca et al. [52]	2012	DCA Vantage	Ion exchange	Tosoh G8, Tosoh Bioscience Ltd	Venous	Venous	Clinical	POC	100	7.3
Lee et al. [53]	2015	SDA1c Care	Ion exchange	Biorad Variant II, Bio-Rad Laboratories Ltd	Venous	Venous	Laboratory	Laboratory	150	–

Table 1 (continued)

Author	Year	POC device(s)	Comparator method	Comparator instrument and manufacturer	POC sample	Comparator sample	Operator	Setting	Number of patients	Mean HbA _{1c} , %
Lenters-Westra et al. [3]	2010a	DCA Vantage Afinion	Ion exchange	Tosoh G7, Tosoh Bioscience Ltd	Venous	Venous	Laboratory	Laboratory	40	–
Lenters-Westra and Slingerland [54]	2010b	Innovastar Clover	Ion exchange	Tosoh G8, Tosoh Bioscience Ltd	Venous	Venous	Laboratory	Laboratory	40	–
Lenters-Westra and Slingerland [7]	2014	Nycocard Quo-Test DCA Vantage Afinion	Ion exchange	Tosoh G8, Tosoh Bioscience Ltd	Venous	Venous	Laboratory	Laboratory	40	–
Malkani et al. [55]	2013	Innovastar Quo-Test	Ion exchange	Tosoh G8, Tosoh Bioscience Ltd	Venous	Venous	Clinical Laboratory	POC	45	–
Manley et al. [56]	2014	Quo-Lab B-Analyst	Ion exchange	Menarini HA8160, A Menarini diagnostics Ltd	Venous	Venous	Laboratory	Laboratory	128	–
Marley et al. [57]	2015	Cobas B101 DCA Vantage	Immunoassay	Cobas Integra 800, Roche Diagnostics Ltd	Capillary	Venous	Clinical	Clinical	241	–
Martin et al. [58]	2005	DCA 2000	Ion exchange	Biorad Variant II, Bio-Rad Laboratories Ltd	Capillary	Venous	–	POC	88	7.1
Mattewal et al. [59]	2007	A1cNow	Ion exchange	Tosoh 2.2, Tosoh Bioscience Ltd	Venous	Venous	Laboratory	Laboratory	34	7.1
Menendez-Valladares et al. [60]	2015	B-analyst	Ion exchange	Menarini HA8180, A Menarini diagnostics Ltd	Venous	Venous	Laboratory	Laboratory	120	–
Nam et al. [61]	2011	A1cNow	Ns	a	Capillary	Venous	Laboratory	POC	83	9.1
Petersen et al. [8]	2010	DCA Vantage Afinion	Ion exchange	Biorad Variant II, Bio-Rad Laboratories Ltd	Venous	Venous	Laboratory	Laboratory	86/88	7.9
Phillips et al. [62]	2014	Quo-Test, HemoCue, Afinion, B-analyst	Ion exchange	Menarini, A Menarini diagnostics Ltd	Venous	Venous	–	–	39–80	–
Pope et al. [28]	1993	DCA 2000	Ion exchange	Diamat, Bio-Rad Laboratories Ltd ^a	Capillary	Venous	Clinical	POC	15–24	–
Rossi et al. [63]	2015	A1cNow	Ion exchange	Tosoh G8, Tosoh Bioscience Ltd	Capillary	Venous	Laboratory	Laboratory	81	–
Sanchez-Mora et al. [64]	2011	DCA Vantage Afinion	Ion exchange	Menarini HA 8160, A Menarini diagnostics Ltd	Venous	Venous	Laboratory	Laboratory	53	–
Schweitzer et al. [65]	2012	A1cNow	Ion exchange	Biorad Variant II, Bio-Rad Laboratories Ltd	–	–	–	–	453	8.6
Shemesh et al. [66]	2003	DCA 2000	Ion exchange	Pharmacia MonoS, GE Healthcare Life Sciences ^a	Venous	Venous	–	POC	39	6.5
Shemesh et al. [67]	2006	DCA 2000	Ion exchange	Pharmacia MonoS, GE Healthcare Life Sciences ^a	Venous	Venous	–	POC	117	5.5
Shephard and Whiting [68]	2006	DCA 2000	Ion exchange	Pharmacia MonoS, GE Healthcare Life Sciences ^a	Capillary	Venous	–	–	100	7.1
Shimoda et al. [69]	2013	A1cNow	Ion exchange	Tosoh G7, Tosoh Bioscience Ltd	Venous	Venous	Clinical	POC	144	7.3
St John et al. [9]	2006	DCA 2000 A1cNow	Affinity	Primus CLC330, Trinity Biotech Plc ^a	Venous	Venous	Clinical	POC	112	–
		Nycocard								

Table 1 (continued)

Author	Year	POC device(s)	Comparator method	Comparator instrument and manufacturer	POC sample	Comparator sample	Operator	Setting	Number of patients	Mean HbA _{1c} %
Szymezak et al. [70]	2008	DCA Vantage	Ion exchange	Biorad Variant II, Bio-Rad Laboratories Ltd	Venous	Venous	Laboratory	Laboratory	100	
Tamborlane et al. [29]	2005	DCA 2000	Ion exchange	Tosoh A1c 2.2 Plus	Capillary	Capillary	—	POC		
Thueringer et al. [71]	2012	A1cNow	Ion exchange	Tosoh Bioscience Ltd	Capillary	Venous	Patients	POC	50	9.9
Twomey et al. [72]	2008	DCA 2000	Ion exchange	Biorad Variant II, Bio-Rad Laboratories Ltd	Venous	Venous	Laboratory	Laboratory		
Villar-del-Campo et al. [24]	2014	DCA	Ion exchange	Biorad Variant II, Bio-Rad Laboratories Ltd	Capillary	Venous	Clinical	Clinical	102	6.5
Wan Mohd Zin et al. [73]	2013	Afinion	Ion exchange	Menarini, A Menarini diagnostics Ltd	Capillary	Venous	—	—	139	7.1
		Quo-Test		Adams HA8160, Atkray Inc.	Venous	Venous				
Wehmeier [74]	2012	Innovastar	Ns	^a	Venous	Venous	Laboratory	Laboratory	30	
Williams et al. [75]	2014	Quo-Lab	Ion exchange	Biorad Variant II, Bio-Rad Laboratories Ltd	Venous	Venous	—	—	41	—
Wood et al. [30]	2012	DCA Vantage Afinion	Ion exchange	Tosoh, Tosoh Bioscience Ltd ^a	Capillary	Capillary	Clinical	POC	40	
Zhou et al. [76]	2014	DCA	Ion exchange	Tosoh G8, Tosoh Bioscience Ltd	Venous	Venous	—	—	50	—

Ns, not specified; HPLC, high performance liquid chromatography based method; Ion exchange, ion exchange chromatography; Affinity, affinity chromatography. Numbers in brackets after author name are references given in the list of references [1–84]. ^aUnclear whether NGSP certified comparator methods were used. A1c gear – Sakae, Japan; A1cNow – PTS Diagnostics, USA; Afinion – Alere Inc., USA; B-analyst – Menarini, Italy; Clover – Neon Diagnostics, UK; Cobas b101 – Roche Diagnostics, Switzerland; DCA – Siemens, Germany; HemoCue – HemoCue AB, Sweden; InnovaStar – Diasys, Germany; NycoCard – Alere Inc., USA; Quo-Laboratory – EKF Diagnostics, UK; Quo-Test – EKF Diagnostics, UK; SDA1cCare – SD Biosensor, Korea.

was small across all studies, bias within a single device could vary substantially between studies. For example, the mean bias across studies was -0.05% HbA_{1c} for the A1cNow device, but it varied from -0.70% (95% confidence interval, CI -0.82 to 0.58%) [56] to $+0.67\%$ (95% CI 0.52 – 0.82%) [21]. Similarly for the DCA device, mean bias varied from -0.96% to $+0.28\%$. Even when the mean bias was small, individual differences in measurements within a single study could be larger, as shown by the two SD bars (Figure 1), which typically extended over a range of 1% – 1.5% HbA_{1c} (10.9 – 16.4 mmol/mol). Moreover, the width of the prediction intervals in Figure 2 indicate that for most devices the mean bias of a future study could vary widely. Summary estimates of the SD of the differences (Figure 3) ranged from 0.21% to 0.53% but again individual SDs could be larger and varied between studies. Meta-regression did not find that year of publication significantly affected the mean bias for any of the devices but there was a trend towards a less negative bias over time in the DCA device (coefficient 0.014% HbA_{1c} per year, $p=0.081$). A post hoc meta-regression found that the standard deviation of the bias in the DCA device significantly decreased over time (-0.01% HbA_{1c} per year, 95% CI -0.015 to -0.006 , $p<0.0001$), suggesting that the variability in measurements contributing to the mean bias in the DCA device is decreasing over time. A subgroup analysis to compare the size of the SD in studies published prior to 2006 with those published between 2006 and 2015 found that those published in the last 10 years had a mean SD of 0.33% (95% CI 0.30 – 0.37) HbA_{1c} compared with a SD of 0.45% (95% CI 0.38 – 0.52) HbA_{1c} before 2006.

Sensitivity analyses excluding studies in which approximations were necessary during data extraction and those at highest risk of bias (Supplementary Data, Figures 2 and 3) gave broadly similar results. Nine studies which used NGSP or IFCC reference laboratories for the comparator method no longer had data for three devices (A1cGear, HemoCue and SDA1cCare). There was no longer a significant mean bias for the DCA device and variability was lower (mean bias 0.00% , 95% CI -0.16 to 0.15), whereas the Quo-Test had a significantly negative bias (mean bias -0.27% , 95% CI -0.50 to -0.03) (Supplementary Data, Figure 4). Studies published in 2014–2015 also had a non-significant mean bias for the DCA device (-0.08% , 95% CI -0.21 to 0.06).

Subgroup analyses for the DCA device comparing different settings found no difference in mean bias in studies carried out in a point of care setting -0.27% (95% CI -0.38 to -0.16) compared with a laboratory setting -0.30% (95% CI -0.47 to -0.14). Similarly, mean bias in studies carried out by a laboratory operator was not




Figure 1: Forest plot showing mean bias (in % HbA_{1c}) ± 2SD of HbA_{1c} measured using point-of-care devices compared with laboratory tests, by POC device, ordered by mean bias.

significantly different from that in those carried out by a clinical operator (−0.30%, 95% CI −0.47 to −0.14 vs. −0.30%, 95% CI −0.42 to −0.18, $p=0.956$) (Supplementary

Data, Figure 5). Sensitivity analysis gave similar results (not shown). Subgroup analyses carried out for the A1cNow device also found no differences in mean bias



–0.16 in adults compared with –0.30%, –0.50 to –0.10 in children). There was no difference in mean bias between venous or capillary blood in the A1cNow or DCA (Supplementary Data, Figures 6 and 7).




Figure 3: Forest plot showing standard deviation of differences, 95% confidence intervals of SD and prediction intervals for included studies, by POC device, ordered by SD.

There were sufficient data to carry out meta-analysis on diagnostic accuracy for five devices (Afinion, DCA, A1cNow, Quo-Test and Nycocard). Sensitivity across all

the devices was similar, although specificity varied more, with the Afinion and DCA having the highest specificity at a cutoff of 6.5% HbA_{1c} (48 mmol/mol) (Supplementary

Data, Table 2 and Figure 8). Repeating this analysis only in studies published from 2006 to 2015 did not change the results.

Imprecision measured using replicate analyses of a single sample for each analyzer at low, medium and high HbA_{1c} are presented as mean CV in the Supplementary Data (Table 3). All devices bar one which only had a single evaluation had mean CVs above 2% at low HbA_{1c} values (lower than 6% HbA_{1c} , 42 mmol/mol). At high HbA_{1c} (above 8% HbA_{1c} , 64 mmol/mol), mean CV was <2% for 4 devices, all of which only included data from one study. Five devices had a total CV <2%. Only two devices had all measured CVs below 2% (Cobas b101 and A1cgear) but both of these were only based on a single evaluation. Imprecision of comparator methods is shown in Table 4 in Supplementary Data. Each CV was from a single study, in some cases using multiple data points from repeat analysis of different samples. CVs ranged from 0.8% for Tosoh (model not specified) to 2.3% for the Diamat instrument. Between-laboratory precision reported in the 2016 CAP survey [77] is shown for the comparator methods and Afinion, DCA 2000 and DCA Vantage devices in Table 4 in Supplementary Data which range from 1.6% for the Tosoh G8 to 3.1% for Afinion.

Study quality/risk of bias

Sixteen studies reported receiving funding from the POC device manufacturers [7, 9, 27, 30, 35, 36, 39, 41, 47, 50, 51, 56, 58, 59, 65, 70]. In 18 of the included studies, the instrument was provided free of charge by the manufacturers [3, 7, 9, 24, 27, 31, 36, 38, 39, 44, 50, 53, 54, 56, 59, 65, 69, 70]. In eight studies, it was unclear what laboratory methodology was used in the comparator method [22, 33, 35, 41, 43, 51, 61, 74]. Another seven studies did not use a wide range of HbA_{1c} concentrations to assess mean bias across all HbA_{1c} concentrations [24, 25, 41, 50, 57, 71, 73]. Nine studies used an NGSP- or IFCC-certified laboratory for the comparator method [3, 7, 29, 41, 50, 54, 56, 59, 60].

Discussion

This analysis has shown that the majority of POC HbA_{1c} devices have a mean negative bias compared to laboratory methods. Devices with the most comparisons were found to have higher specificity, rather than consistently high sensitivity, for detecting HbA_{1c} above the diagnostic threshold. Unexplained variation in the mean bias was also observed

for most devices, as indicated by large variations in SDs of the bias as well as high imprecision from the repeated analysis of the same sample. This is the first meta-analysis to compare performance between HbA_{1c} measured using POC devices and laboratory methods drawn from published data. We have used a novel approach to pool data on mean bias, variability, device precision and sensitivity and specificity. Our analysis has demonstrated that differences between POC HbA_{1c} and comparator methods can vary considerably within a single device across all the included studies with POC values ranging from as much as 1.5% HbA_{1c} below to 1.5% HbA_{1c} above the comparator method HbA_{1c} across all devices.

Many evaluations included in our review compared capillary blood on the POC device with venous blood on the comparator method. Although there may be slight differences in HbA_{1c} in different sample types [78, 79] we have been able to eliminate this as a source of bias in two devices (A1cNow and DCA). We did not find evidence that changes in calibration over time resulted in significantly lower levels of bias in device; however, these data suggests that variability in the DCA device may be improving over time, suggesting that its precision may also have improved over time. The diagnostic accuracy of devices was not affected when only the most recent studies were included.

Acceptable limits for bias and imprecision vary between organizations; the College of American Pathologists (CAP) criteria state that acceptable limits of bias are $\pm 6\%$ [80], whereas a within-laboratory CV of <2% (in % units) is widely recommended [4, 80]. More recently, the IFCC Task Force on HbA_{1c} Standardization have recommended using a sigma metrics approach to evaluating analytical error. This takes into account both bias and imprecision in one model based on total allowable error, where bias and imprecision are not treated as separate performance measures [81]. The CAP survey [80] includes data from HbA_{1c} POC devices and shows comparable performance to a number of routine laboratory analysers. However, the survey is aimed at clinical laboratories and participation is not mandatory for instruments used in non-laboratory settings. Those that enter the survey may therefore be a self-selecting group and under-represent the true extent of the use of POC devices in the USA.

Limitations

Sensitivity and specificity at the diagnostic threshold (6.5% HbA_{1c} , 48 mmol/mol) were not reported in the included papers. However, we have been able to combine

data from plots and unpublished data from authors to compare sensitivity and specificity of five devices at the diagnostic threshold.

Some laboratory instruments have been reported to have systematic biases against reference laboratories [46] which may have led to under- or over-estimation of biases in some studies in our review. Only nine studies reported to use an NGSP or IFCC reference laboratory for the comparator method which generally resulted in lower variability. The actual number of studies using reference laboratories may have been higher as many of the papers included in this review were conference abstracts which were limited by word count.

Even though the intended end-users of POC devices are clinicians or healthcare workers, the majority of the accuracy studies were carried out in a laboratory setting by laboratory staff. The translatability of the evaluations included in the meta-analysis to the clinical end-user is therefore a major limitation of this study. We have been able to examine the impact of operator and setting on bias for two devices, DCA and A1cNow, both of which showed no significant difference in performance between a clinical and laboratory operator. It is important to note that practices and regulations in POC testing vary between countries and therefore POC testing may impact differently on patient management between geographic locations.

Mean values from duplicate POC tests were used by some researchers [64], and this may lead to the reported variability being lower than actual variability. In some studies, it was not clear whether results were reported as relative mean bias or absolute mean bias; however, we have excluded studies with unclear reporting in a sensitivity analysis and found similar results.

Clinical implications

Because the majority of devices had an overall negative bias as well as large standard deviations, their use in a clinical setting may result in differences in treatment decisions compared with results from laboratory comparator methods. POC HbA_{1c} results which are lower than laboratory methods may result in undertreatment in some cases. The implications of this on overall patient care may result in fewer short-term hypoglycemic episodes, however, the long-term use may leave some patients exposed to higher than optimal levels of glycemia. The long-term effects of this are not known, but studies have shown that even small increases in HbA_{1c} increase the risk of micro- and macrovascular complications [1]. This may be outweighed by the other benefits of POC testing including improved

ability of the clinicians to make appropriate and timely treatment decisions [82, 83]. There was insufficient published data to enable comparison of performance between settings and operators for most devices; further evaluations in a clinical setting are now needed.

As well as bias, good reproducibility or precision is crucial in device performance; however, imprecision in nearly all devices was poor with the majority of devices having mean CVs >2% at low HbA_{1c} levels. In contrast, only one of the comparator instruments had a mean CV over 2%. These imprecision data are based on repeated analysis of a single sample; our review has shown that there is also considerable variability in HbA_{1c} results in most devices in evaluations over a range of HbA_{1c} concentrations which may better represent clinical situations. These high levels of uncertainty in results may mean that some real changes in HbA_{1c} may not be picked up and could result in different treatment decisions. The combination of instrument performance and biological variability in HbA_{1c} [84] could substantially affect the numbers of patients diagnosed with diabetes and impact on clinical decision making for those with established diabetes. These high levels of imprecision and the large variability is a key finding of this review and may be a crucial factor influencing the adoption of these technologies.

This analysis contributes to the information available to clinicians who are considering whether to use POC HbA_{1c} testing. The decision to select an analyzer must be carefully considered based on clinical needs and what is deemed to be acceptable bias and precision of the analyser. The NGSP recommends that for method certification 37 out of 40 HbA_{1c} tests should be within $\pm 6\%$ relative to the standard reference laboratory measurement [80]. Our analysis shows that the mean measurement bias of two devices was >6% relative to the comparator method, however, because of the large variability observed within devices it is probable that many of the devices will give some measurements greater or <6% of the patient's true HbA_{1c}. It is therefore likely that some of the devices would fail to meet the requirements of 37 of 40 test results within 6% of reference laboratory measurements.

The impact of using POC HbA_{1c} testing on medication use, clinical decision making and patient outcomes now needs to be evaluated in a randomized trial with full economic evaluation.

Acknowledgments: This report is independent research supported by the National Institute for Health Research (NIHR) (Doctoral Research Fellowship, Ms Jennifer Hirst, DRF-2013-06-086). The views expressed in this publication are those of the author(s) and not necessarily those of

the NHS, NIHR or the Department of Health. We are grateful to Dr. Erna Lenters-Westra, Samir Malkani, Bernard Vialettes and Katrina Ruedy on behalf of Jamie Wood for providing us with unpublished data. We also thank Ms. Nia Roberts for her assistance in optimizing the search strategy.

Author contributions: All the authors have accepted responsibility for the entire content of this submitted manuscript and approved submission. J.H. designed the study, researched data, conducted the analysis and wrote manuscript, J.M. researched data and reviewed/edited manuscript. C.P. reviewed/edited manuscript and contributed to discussion. E.E. reviewed/edited manuscript and contributed to discussion. B.F. researched data, contributed to the analysis and reviewed/edited manuscript. R.S. contributed to the analysis, reviewed/edited manuscript and contributed to discussion. A.F. reviewed/edited manuscript and contributed to discussion.

Research funding: None declared.

Employment or leadership: None declared.

Honorarium: None declared.

Competing interests: C.P. is a member of the NIHR Diagnostic Evidence Co-operative Oxford. The views expressed are those of the authors and not necessarily those of the NHS, the NIHR or the Department of Health. The funding organization(s) played no role in the study design; in the collection, analysis, and interpretation of data; in the writing of the report; or in the decision to submit the report for publication. The researchers retained complete independence in the conduct of this study. A.F. is an NIHR Senior Investigator.

References

- Stratton IM, Adler AI, Neil HA, Matthews DR, Manley SE, Cull CA, et al. Association of glycaemia with macrovascular and microvascular complications of type 2 diabetes (UKPDS 35): prospective observational study. *BMJ* 2000;321:405–12.
- WHO. Use of glycated haemoglobin (HbA1c) in the diagnosis of diabetes mellitus, 4 December 2014. Available at: http://www.who.int/diabetes/publications/report-hba1c_2011.pdf.
- Lenters-Westra E, Slingerland RJ. Six of eight hemoglobin A1c point-of-care instruments do not meet the general accepted analytical performance criteria. *Clin Chem* 2010;56:44–52.
- Goodall I, Colman PG, Schneider HG, McLean M, Barker G. Desirable performance standards for HbA1c analysis – precision, accuracy and standardisation: consensus statement of the Australasian Association of Clinical Biochemists (AACB), the Australasian Diabetes Society (ADS), the Royal College of Pathologists of Australasia (RCPA), Endocrine Society of Australia (ESA), and the Australian Diabetes Educators Association (ADEA). *Clin Chem Lab Med* 2007;45:1083–97.
- Weykamp CW, Mosca A, Gillery P, Panteghini M. The analytical goals for hemoglobin A(1c) measurement in IFCC units and National Glycohemoglobin Standardization Program Units are different. *Clin Chem* 2011;57:1204–6.
- Sacks DB, Arnold M, Bakris GL, Bruns DE, Horvath AR, Kirkman MS, et al. Guidelines and recommendations for laboratory analysis in the diagnosis and management of diabetes mellitus. *Clin Chem* 2011;57:e1–47.
- Lenters-Westra E, Slingerland RJ. Three of 7 hemoglobin A1c point-of-care instruments do not meet generally accepted analytical performance criteria. *Clin Chem* 2014;60:1062–72.
- Petersen JR, Omoruyi FO, Mohammad AA, Shea TJ, Okorodudu AO, Ju H. Hemoglobin A1c: assessment of three POC analyzers relative to a central laboratory method. *Clin Chim Acta* 2010;411:2062–6.
- St John A, Davis TM, Goodall I, Townsend MA, Price CP. Nurse-based evaluation of point-of-care assays for glycated haemoglobin. *Clin Chim Acta* 2006;365:257–63.
- Wood JA, Kaminski BM, Kollman C, Diabetes P. The accuracy and precision of the axis-shield Afinion HbA1c measurement device. *Diabetes* 2011; Conference: 71st Scientific Sessions of the American Diabetes Association San Diego, CA United States. Conference Start: 20110624 Conference End: 20110628. Conference Publication: (var.pagings). 60:A237.
- Health Quality Ontario. Point-of-care hemoglobin A1c testing: an evidence-based analysis. *Ont Health Technol Assess Ser* 2014;14:1–30.
- Wan Y, Heneghan C, Stevens R, McManus RJ, Ward A, Perera R, et al. Determining which automatic digital blood pressure device performs adequately: a systematic review. *J Hum Hypertens* 2010;24:431–8.
- Bland JM, Altman DG. Statistical methods for assessing agreement between two methods of clinical measurement. *Lancet* 1986;1:307–10.
- Whiting PF, Rutjes AW, Westwood ME, Mallett S, Deeks JJ, Reitsma JB, et al. QUADAS-2: a revised tool for the quality assessment of diagnostic accuracy studies. *Ann Intern Med* 2011;155:529–36.
- DerSimonian R, Laird N. Meta-analysis in clinical trials. *Control Clin Trials* 1986;7:177–88.
- Williamson PR, Lancaster GA, Craig JV, Smyth RL. Meta-analysis of method comparison studies. *Stat Med* 2002;21:2013–25.
- Higgins JPT, Green S, editors. *Cochrane Handbook for Systematic Reviews of Interventions Version 5.1.0 [updated March 2011]*. The Cochrane Collaboration, 2011. Available from www.handbook.cochrane.org.
- Ahn S, Fessler JA. Standard errors of mean, variance, and standard deviation estimators. *EECS Department, The University of Michigan*, 2003:1–2.
- Riley RD, Higgins JP, Deeks JJ. Interpretation of random effects meta-analyses. *BMJ* 2011;342:d549.
- Harbord RM, Whiting P. Metandi: meta-analysis of diagnostic accuracy using hierarchical logistic regression. *Stata J* 2009;9:211–29.
- Affret A, Griz LH, Cesse EA, Specht YD, Carvalho EM, Fontbonne A. Assessment of a glycated hemoglobin point-of-care analyzer (A1CNow+) in comparison with an immunoturbidimetric method: a diagnostic accuracy study. *Sao Paulo Med J* 2015;133:460–4.
- Boz M, Gerard P, Scheen AJ, Lefebvre PJ, Castillo MJ. Rapid determination by immunoassay of glycosylated hemoglobin in capillary blood compared to an affinity method for boronate and

- ion capturing on venous blood. [French]. *Ann Biol Clin (Paris)* 1997;55:139–44.
23. Carrera Font MT, Sole Brichs MC, Sala Alvarez MC, Navarro Olivella JM, Servent Turo J, Felipe Fernandez MP. Capillary HbA1c determination on type 2 diabetes patients in a primary health centre. [Spanish]. *Aten Primaria* 2011;43:536–43.
 24. Villar-del-Campo MC, Rodriguez-Caravaca G, Gil-Yonte P, Cidoncha-Calderon E, Garcia-Cruces Mendez J, Donnay-Perez S. [Diagnostic agreement between two glycosylated a1b hemoglobin methods in Primary Care]. *Semergen* 2014;40:431–5.
 25. Chan CL, McFann K, Newnes L, Nadeau KJ, Zeitler PS, Kelsey M. Hemoglobin A1c assay variations and implications for diabetes screening in obese youth. *Pediatr Diabetes* 2014;15:557–63.
 26. El Arabi H, Willems D, Melot C, Dorchy H. [Evaluation of DCA vantage for rapid in-clinic measurement of HbA1c on capillary blood in young type 1 diabetic patients]. *Rev Med Brux* 2013;34:87–9.
 27. Greaves RF, Northfield J-A, Cameron FJ. Haemoglobin A1c: evaluation of three point of care analysers for use in a paediatric diabetes clinic. *Ann Clin Biochem* 2005;42(Pt 2):124–9.
 28. Pope RM, Apps JM, Page MD, Allen K, Bodansky HJ. A novel device for the rapid in-clinic measurement of haemoglobin A1c. *Diabet Med* 1993;10:260–3.
 29. Tamborlane WV, Kollman C, Steffes MW, Ruedy KJ, Dongyuan X, Beck RW, et al. Comparison of fingerstick hemoglobin A1c levels assayed by DCA 2000 with the DCCT/EDIC central laboratory assay: results of a Diabetes Research in Children Network (DirecNet) Study. *Pediatr Diabetes* 2005;6:13–6.
 30. Wood JR, Kaminski BM, Kollman C, Beck RW, Hall CA, Yun JP, et al. Accuracy and precision of the Axis-Shield Afinion hemoglobin A1c measurement device. *J Diabetes Sci Technol* 2012;6:380–6.
 31. Dupuy AM, Elong-Bertard C, Badiou S, Barrot A, Bargnoux AS, Cristol JP. Analytical performance and clinical use of a hemoglobin A1c point-of-care analyzer in a pediatric unit. *J Diabetes Sci Technol* 2013;7:1408–9.
 32. Arsie MP, Marchioro L, Lapolla A, Giacchetto GF, Bordin MR, Rizzotti P, et al. Evaluation of diagnostic reliability of DCA 2000 for rapid and simple monitoring of HbA1c. *Acta Diabetologica* 2000;37:1–7.
 33. Azevedo T, Ferreira M, Giestas A, Carvalho AC, Pinto S, Palma I. HbA1c-evaluation of the DCA2000 and quo-test systems. *Endocrine Reviews Conference: 93rd Annual Meeting and Expo of the Endocrine Society, ENDO 2011;32(3 Meeting Abstracts)*.
 34. Berry H, Hall G, Donovan J. Comparison of point of care haemoglobin A1c meters. *Clin Chem Lab Med* 2014;52:eA326–7.
 35. Cagliero E, Levina EV, Nathan DM. Immediate feedback of HbA1c levels improves glycemic control in type 1 and insulin-treated type 2 diabetic patients. *Diabetes Care* 1999;22:1785–9.
 36. Clark AL, Irvin BR, Isbell AW, Ammirati EA. Multi-site correlation of A1cNow %A1C assay to a standardized laboratory. *Clin Chem* 2005;51:A259–A.
 37. Criel M, Jonckheere S, Langlois M. Evaluation of three hemoglobin A1c point-of-care instruments. *Acta Clinica Belgica: International Journal of Clinical and Laboratory Medicine* 2015;70:S3.
 38. Dupuy AM, Badiou S, Elong-Bertard C, Bargnoux AS, Cristol JP. Analytical performance of the Axis-Shield Afinion for hemoglobin A1c measurement: impact of lot number. *Clin Lab* 2014;60:369–76.
 39. Ejilemele A, Unabia J, Ju H, Petersen JR. A1c Gear: laboratory quality HbA1c measurement at the point of care. *Clin Chim Acta* 2015;445:139–42.
 40. Estevez Fidalgo P, Pineiro Caamano A, Otero Santiago M, Garcia Aschauer JM, Lorenzo Freire MT, Espana Barrada R, et al. Validation of HbA1c results obtained with a point of care instrument. *Clin Chem Lab Med* 2014;52:S1523.
 41. Ginde AA, Cagliero E, Nathan DM, Camargo CA, Jr. Point-of-care glucose and hemoglobin A1c in emergency department patients without known diabetes: implications for opportunistic screening. *Acad Emerg Med* 2008;15:1241–7.
 42. Guerci B, Durain D, Leblanc H, Rouland JC, Passa P, Godeau T, et al. Multicentre evaluation of the DCA 2000 system for measuring glycated haemoglobin. DCA 2000 Study Group. *Diabetes Metab* 1997;23:195–201.
 43. Harris S, Tokmakejian S, Edmonds M, Webster-Bogaert S, Mahon J, Porter S, et al. Validation of the bayer DCA2000 (TM) as a point-of-care measure of glycosylated haemoglobin. *Diabetes* 2000;49:A110–A.
 44. Hawkins RC. Comparison of four point-of-care HbA1c analytical systems against central laboratory analysis. *Singapore Med J* 2003;44:8–11.
 45. Heng PY, Fun S, Chui KL, Taufik MA, Lew MY, Wong MS. An evaluation of the performance of Roche Cobas b101 point-of-care system in the measurement of haemoglobin A1C. *Clin Chem Lab Med* 2015;53:S623.
 46. Holmes EW, Ersahin C, Augustine GJ, Charnogursky GA, Gryzbac M, Murrell JV, et al. Analytic bias among certified methods for the measurement of hemoglobin A1c: a cause for concern? *Am J Clin Pathol* 2008;129:540–7.
 47. Jiang F, Hou X, Lu J, Zhou J, Lu F, Kan K, et al. Assessment of the performance of A1cNow(+) and development of an error grid analysis graph for comparative hemoglobin A1c measurements. *Diabetes Technol Ther* 2014;16:363–9.
 48. Jou JM, Bedini JL, Seuma J, Luna N, Kinder M. Evaluati on of the point of care a1cnow+ analyzer for HBA1C. *International Journal of Laboratory Hematology* 2013; Conference: 26th International Symposium on Technological Innovations in Laboratory Hematology, ISLH 2013 Toronto, ON Canada. Conference Start: 20130510 Conference End: 20130512. Conference Publication: (var.pagings). 35:50.
 49. Karami A, Baradaran A. Comparative evaluation of three different methods for HbA1c measurement with High-performance liquid chromatography in diabetic patients. *Adv Biomed Res* 2014;3:94.
 50. Knaebel J, Irvin BR, Xie CZ. Accuracy and clinical utility of a point-of-care HbA1c testing device. *Postgrad Med* 2013;125:91–8.
 51. Leal S, Soto-Rowen M. Usefulness of point-of-care testing in the treatment of diabetes in an underserved population. *J Diabetes Sci Technol* 2009;3:672–6.
 52. Leca V, Ibrahim Z, Lombard-Pontou E, Maraninchi M, Guieu R, Portugal H, et al. Point-of-care measurements of HbA1c: Simplicity does not mean laxity with controls. *Diabetes Care* 2012;35:e85.
 53. Lee K, Jun SH, Han M, Song SH, Park KU, Song WH, et al. Performance evaluation of SD A1cCare as a HbA1c analyzer for point-of-care testing. *Clin Biochem* 2015;48:625–7.
 54. Lenters-Westra E, Slingerland RJ. Evaluation of the Quo-Test hemoglobin A1c point-of-care instrument: second chance. *Clin Chem* 2010;56:1191–3.

55. Malkani S, Korpi-Steiner N, Rao LV. Reducing analytical variation between point-of-care and laboratory HbA1c testing. *J Diabetes* 2013;5:192–6.
56. Manley SE, Hikin LJ, Round RA, Manning PW, Luzio SD, Dunseath GJ, et al. Comparison of IFCC-calibrated HbA1c from laboratory and point of care testing systems. *Diabetes Res Clin Pract* 2014;105:364–72.
57. Marley JV, Oh MS, Hadgraft N, Singleton S, Isaacs K, Atkinson D. Cross-sectional comparison of point-of-care with laboratory HbA_{1c} in detecting diabetes in real-world remote Aboriginal settings. *BMJ Open* 2015;5:e006277.
58. Martin DD, Shephard MD, Freeman H, Bulsara MK, Jones TW, Davis EA, et al. Point-of-care testing of HbA1c and blood glucose in a remote Aboriginal Australian community.[Erratum appears in *Med J Aust* 2005 Jun 20;182:656 Note: Bulsara, Max K [added]]. *Med J Aust* 2005;182:524–7.
59. Matthewal A, Aldasouqi S, Solomon D, Gossain V, Koller A. A1cNow InView: a new simple method for office-based glycohemoglobin measurement. *J Diabetes Sci Technol* 2007;1:879–84.
60. Menendez-Valladares P, Fernandez-Riejos P, Sanchez-Mora C, Perez-Perez A, Sanchez-Margalet V, Gonzalez-Rodriguez C. Evaluation of a HbA1c point-of-care analyzer. *Clin Biochem* 2015;48:686–9.
61. Nam S, Han H-R, Song H-J, Song Y, Kim KB, Kim MT. Utility of a point-of-care device in recruiting ethnic minorities for diabetes research with community partners. *J Health Care Poor Under-served* 2011;22:1253–63.
62. Phillips SG, Nwagbo Y, Ashton K. Analytical evaluation of POCT HbA1c instruments. *Clin Chem Lab Med* 2014;52:eA324–5.
63. Rossi L, Della Bartola L, Giampietro O, Consani C, Pellegrini G, Matteucci E. Point-of-care HbA1c testing in a clinical setting: performance analysis. *Clin Chem Lab Med* 2015;53:S543.
64. Sanchez-Mora C, S Rodriguez-Oliva M, Fernandez-Riejos P, Mateo J, Polo-Padillo J, Goberna R, et al. Evaluation of two HbA1c point-of-care analyzers. *Clin Chem Lab Med* 2011;49:653–7.
65. Schweitzer M, Cavan DA, Ziegler R, Cranston I, Parkin C, Wagner RS. Is HbA1c a reliable measure for assessing glycaemic control? *Diabetologia* 2012; Conference: 48th Annual Meeting of the European Association for the Study of Diabetes, EASD 2012 Berlin Germany. Conference Start: 20121001 Conference End: 20121005. Conference Publication: (var.pagings). 55:S404.
66. Shemesh T, Piers LS, O'Dea K. Use of the Bayer DCA 2000+ for the measurement of glycated haemoglobin in a remote Australian Aboriginal community. *Ann Clin Biochem* 2003; 40(Pt 5):566–8.
67. Shemesh T, Rowley KG, Shephard M, Piers LS, O'Dea K. Agreement between laboratory results and on-site pathology testing using Bayer DCA2000+ and Cholestech LDX point-of-care methods in remote Australian Aboriginal communities. *Clin Chim Acta* 2006;367:69–76.
68. Shephard M, Whiting M. Assessment of the practicability and analytical performance of a point-of-care affinity chromatography haemoglobin A1c analyser for use in the non-laboratory setting. *Ann Clin Biochem* 2006;43(Pt 6):513–5.
69. Shimoda S, Maeda T, Furukawa N, Ichinose K, Taketa K, Igata M, et al. Evaluation of a new device for measurement of hemoglobin A1c for Japanese subjects. *Diabetology International* 2013;4:112–6.
70. Szymezak J, Leroy N, Lavalard E, Gillery P. Evaluation of the DCA Vantage analyzer for HbA1c assay. *Clin Chem Lab Med* 2008;46:1195–8.
71. Thueringer JT, Ishani A, Rector T, Greer N, Ercan-Fang N. Comparison of a1c measurements from a home test device to a hospital laboratory. *Diabetes* 2012; Conference: 72nd Scientific Sessions of the American Diabetes Association Philadelphia, PA, United States. Conference Start: 20120608 Conference End: 20120612. Conference Publication: (var.pagings). 61:A218.
72. Twomey PJ, Rayman G, Pledger DR. Implications of different DCCT-aligned HbA(1c) methods on GMS clinical indicators. *Diabetic Med* 2008;25:97–100.
73. Wan Mohd Zin RM, Ahmad Kamil ZI, Tuan Soh TR, Embong M, Wan Mohamad WN. Haemoglobin A1c: comparing performance of two point of care devices with laboratory analyser. *BMC Res Notes* 2013;6:540.
74. Wehmeier M. Evaluation of POCT HbA1C analyzer innovastar for use in diabetes clinics. *Clin Chem Lab Med* 2012; Conference: Congress of Clinical Chemistry and Laboratory Medicine 9th Annual Meeting of the German Society for Clinical Chemistry and Laboratory Medicine, DGKL 2012 Mannheim Germany. Conference Start: 20120926 Conference End: 20120929. Conference Publication: (var.pagings). 50:A266–7.
75. Williams PF, Tan K, Sherfan J, Woolnough C, Lin E, Brown M, et al. Comparison of three point of care (POC) HbA1c devices with cation exchange and capillary electrophoresis. *Diabetes* 2014;63:A220.
76. Zhou R, Tong Q, Zuo C, Song ZX, Ma HA, Yue YH, et al. Point-of-care testing of HbA1C is traceable to IFCC reference method by external calibration. *Clin Chim Acta* 2014;433:249–53.
77. CAP. College of American Pathologists (CAP) GH5 Survey Data 2016 [09/06/2016]. Available at: <http://www.ngsp.org/CAP/CAP16a.pdf>.
78. Ferreras Amez JM, Aldea Molina E, Ortega Jimenez N, Guardia Sancho L, Gamez Gomez MA, Blasco Valle M. Assessment of glycosylated hemoglobin in capillary blood compared to venous blood. [Spanish]. *Atencion Primaria* 2010;42:60–1.
79. Keramati T, Razi F, Tootee A, Larijani B. Comparability of hemoglobin A1c level measured in capillary versus venous blood sample applying two point-of-care instruments. *J Diabetes Metab Disord* 2014;13:94.
80. NGSP. Harmonising Hemoglobin A1c Testing 2010 [24/8/2015]. Available at: <http://www.ngsp.org/>.
81. Weykamp C, John G, Gillery P, English E, Ji L, Lenters-Westra E, et al. Investigation of 2 models to set and evaluate quality targets for hb a1c: biological variation and sigma-metrics. *Clin Chem* 2015;61:752–9.
82. Ferenczi A, Reddy K, Lorber DL. Effect of immediate hemoglobin A1c results on treatment decisions in office practice. *Endocr Pract* 2001;7:85–8.
83. Miller CD, Barnes CS, Phillips LS, Ziemer DC, Gallina DL, Cook CB, et al. Rapid A1c availability improves clinical decision-making in an urban primary care clinic. *Diabetes Care* 2003;26:1158–63.
84. Lenters-Westra E, Roraas T, Schindhelm RK, Slingerland RJ, Sandberg S. Biological variation of hemoglobin A1c: consequences for diagnosing diabetes mellitus. *Clin Chem* 2014;60:1570–2.

Supplemental Material: The online version of this article (DOI: 10.1515/cclm-2016-0303) offers supplementary material, available to authorized users.