

Test–retest reliability of scotopic and mesopic fundus-controlled perimetry using a modified MAIA in normal eyes

Maximilian Pfau¹, Moritz Lindner¹, Monika Fleckenstein¹, Robert P. Finger¹, Gary S. Rubin^{2,3}, Wolf M. Harmening¹, Marco U. Morales⁴, Frank G. Holz¹, Steffen Schmitz-Valckenberg¹

1. Department of Ophthalmology, University of Bonn, Ernst-Abbe-Str. 2, Bonn, Germany
2. UCL Institute of Ophthalmology, London, United Kingdom
3. NIHR Biomedical Research Centre for Ophthalmology, London, United Kingdom.
4. Division of Clinical Neurosciences, University of Nottingham, U.K.

Running head: Repeatability of scotopic fundus-controlled perimetry
Key words: microperimetry, rod function, cone function, dark adaptation, Bland-Altman, healthy eyes
Number of words: 4462 (excluding title page, abstract, legends, and references)
Number of figures: 4
Number of tables: 5
Number of supplements: 2

Correspondence:

Prof. Dr. Steffen Schmitz-Valckenberg, FEBO
Department of Ophthalmology
University of Bonn
Ernst-Abbe-Str. 2
53127 Bonn
Germany
Tel.: +49 228 287 16826
Fax: +49 228 287 11470
e-mail: steffen.schmitz-valckenberg@ukb.uni-bonn.de

Financial Support: CenterVue SpA, Padova, Italy has provided research material (S-MAIA) for the conduct of this study. CenterVue had no role in the design or conduct of the experiments.

Author Disclosure Information: **M. Pfau:** Heidelberg Engineering, Optos, Zeiss, F; Heidelberg Engineering, R; **M. Lindner:** Heidelberg Engineering, Optos, Zeiss, Genentech F; **M. Fleckenstein:** Heidelberg Engineering, Optos, Zeiss, F; Heidelberg Engineering, R; **R.P. Finger:** Heidelberg Engineering, Optos, Zeiss, F; **G. Rubin;** **W. Harmening:** Heidelberg Engineering, Optos, Zeiss, F; **M.U. Morales:** Centervue I; **F.G. Holz:** Heidelberg Engineering, Optos, Zeiss, F; Heidelberg Engineering, Zeiss, C; Heidelberg Engineering, R; **S. Schmitz-Valckenberg:** Heidelberg Engineering, Optos, Zeiss, F; Heidelberg Engineering, R

Abstract:

Purpose: To assess the intra-session test-retest reliability of scotopic-cyan and scotopic-red fundus-controlled perimetry (FCP) in normal subjects using a modified MAIA “micro-perimeter” (Macular-Integrity-Assessment).

Methods: Forty-seven normal eyes of 30 subjects (33.8 years) underwent duplicate mesopic (achromatic stimuli, 400-800nm), scotopic-cyan (505nm) and scotopic-red (627nm) FCP, using a grid of 49 stimuli over 14° of the central retina. Test-retest reliability for point-wise-sensitivity (PWS), stability of fixation, reaction-time and test-duration were analyzed using mixed-effects models.

Results: PWS test-retest reliability was good among all three types of retinal sensitivity assessments (Coefficient- of-repeatability (CoR) of 4.75dB for mesopic, 5.26dB for scotopic-cyan and 4.06dB for scotopic-red testing). While the mean sensitivity decreased with eccentricity for both mesopic and scotopic-red testing, it was highest at 7° eccentricity for scotopic-cyan assessment ($p < 0.001$).

Conclusions: The modified MAIA device allows for reliable scotopic FCP in normal subjects. Our findings suggest that testing of scotopic-cyan sensitivity largely reflects rod function.

Introduction

Tremendous progress has been achieved since the first description of fundus-controlled perimetry (FCP) in 1851 by Hermann von Helmholtz.[1] Early approaches based on direct ophthalmoscopy or direct slit-lamp ophthalmoscopy were highly examiner dependent and allowed only photopic testing due to high levels of light necessary for retinal illumination.[1–5] Today, using infrared light (IR) and scanning laser ophthalmoscopy (SLO), it has become possible to reduce the background illumination for FCP to a degree which has opened the door for a topographic point-by-point assessment of scotopic function.[6–8]

Multiple retinal diseases are associated with rod photoreceptor functional impairment including pseudoxanthoma elasticum,[9] macular telangiectasia type 2,[10] retinitis pigmentosa,[11–13] and rod-cone dystrophies.[14] Moreover, histological evidence suggests that rod loss precedes cone loss in both exudative and non-exudative age-related macular degeneration (AMD)[15,16] – the leading cause of legal blindness in developed countries.[17] In addition to rod loss, cone degeneration has also been observed histologically in AMD.[16] Psychophysical data in patients with early and late AMD have further underlined that rod dysfunction exceeds cone dysfunction, whereby the underlying mechanisms for the greater vulnerability of rods vs. cones is largely unknown.[18–24] Cone specific dysfunction is also detectable in AMD both psychophysically and with objective methods such as multifocal pupillographic perimetry or multifocal electroretinography.[25–29]

The first commercially available device to perform automated dark adapted perimetry was a modified first-generation Humphrey Field Analyser (HFA, Carl Zeiss Meditec Inc, USA).[30] This device did not correct for fixation (i.e. not fundus-controlled), which was likely the reason for the inconclusive results for structure-function correlation analyses in patients with AMD.[18] Later, it was demonstrated that scotopic sensitivity loss exceeded photopic sensitivity impairment in areas with increased levels of fundus autofluorescence.[31] Secondly, the no-longer commercially available scanning laser ophthalmoscope (SLO 101, Rodenstock, Ottobrunn, Germany) has been used for scotopic FCP – however, only a single congress abstract has been available so far to the best of our knowledge (Van de Velde FJ,

et al. IOVS 1993; 34:ARVO Abstract 3542). Thirdly, a modified version (MP-1S) of the well-established MP-1 microperimeter (Nidek Technologies, Padova, Italy) has become available in the meantime.

While many reports on mesopic function with the MP-1 device are available, only a few reports have been published on scotopic FCP using the MP-1S.[20,21,24,32–35] For example, we have recently shown that the presence of reticular drusen is spatially confined to outer retinal thinning and impairment of scotopic function.[21] Moreover, a correlation of the extent of outer retinal thinning and the extent of scotopic sensitivity loss was shown. Another modified version of the MP-1 has also been used for fundus-controlled two-color perimetry and dark adaptometry.[36,37] However, the dynamic range of the stimulus-presenting LCD display of the MP-1 (20 dB) is considered to be low and prone to ceiling and floor effects in both mesopic and scotopic examinations.[20,32,33] Moreover, the quality of the reference image (IR fundus camera) has been reported to be low, making reliable detection of retinal pathologies and test pattern placement challenging.[38]

Hence, there is an unmet need for a reliable scotopic SLO-based FCP system in both clinical and research settings. A modified version of the well-established Macular Integrity Assessment (MAIA, CenterVue, Padova, Italy) device with two additional projection LEDs and the possibility to reduce the SLO laser power for scotopic testing has become recently available. It allows for mesopic FCP testing with achromatic stimuli as well as scotopic testing with both cyan (505 nm) and/or red (627 nm) stimuli. While the scotopic cyan testing is intended to be largely derived from rod photoreceptor-mediated function, also excluding S-cone activity, the scotopic red would be more influenced by cone-mediated function, rather reflecting a mixture of both rod and cone mediated responses.[39–41]

The aim of the study was to assess the test-retest reliability of scotopic cyan and scotopic red FCP in normal subjects using a modified MAIA. Specifically, we evaluated point-wise sensitivity, fixation stability and reaction time with regard to reliability. The results were also compared to mesopic testing performed during the same session using the same device.

111 Furthermore, the differences in retinal sensitivity and reaction time among the types of
112 testing were analyzed with regard to photoreceptor-mediation (cone vs. rod). Finally, we
113 discuss the results of test-retest reliability in comparison to previous reports of mesopic
114 testing using the MAIA device.

Methods

Subjects

This study was designed to enroll 30 subjects with at least one eye with no known retinal disease at the Department of Ophthalmology, University of Bonn, Germany. Subjects were recruited from inpatients, family members of inpatients, medical students and colleagues. The inclusion criteria for study eyes were best-corrected visual acuity of 6/6as assessed by autorefraction (ARK-560A, Nidek, Gamagori, Japan). Exclusion criteria were refractive errors > 5.00 diopters of spherical equivalent and > 1.50 diopters of astigmatism as well as history of glaucoma, retinal diseases, uveitis, diabetes, history of prior vitreoretinal surgery or retinal laser treatment. If both eyes met the inclusion criteria, both eyes underwent the full imaging and functional testing protocol in random order. Apart from taking the medical history, spectral domain optical coherence tomography (SD-OCT) raster scanning was performed using a 30° x 25° scan field (61 B-scans, automated real time (ART) mode 20 frames, centered on the fovea, Spectralis OCT2, Heidelberg Engineering, Heidelberg, Germany).

The study was approved by the Institutional Review Board of the University of Bonn (ethics approval ID: 191/16). After explanation of the nature and possible consequences of the study, informed written consent was obtained from all subjects. The protocol followed the tenets of the Declaration of Helsinki.

Fundus-controlled perimetry (FCP)

All study eyes underwent two mesopic as well as two scotopic cyan and two scotopic red FCP tests of the central retina using a modified MAIA (software version 2.0.4). The basic design of the MAIA platform has been described previously.[42] Briefly, it employs confocal scanning laser ophthalmoscopy (cSLO) with a central wavelength of 850 nm (36.5° x 36.5°, 25 fps) for fundus tracking. In the current study, a testing grid with 49 stimuli was used. The stimuli grid distribution contained 4 concentric rings with 12 evenly distributed stimuli (angular

distance of 30° to each other), separated at 1°, 3°, 5° and 7° from a central stimulus (Fig. 1).

Prior to testing, pupil dilatation was performed using 1.0% tropicamide.

A 645 nm red ring with 1° diameter size was used as target of fixation (with the preset luminance for mesopic or the preset luminance for scotopic testing). Goldmann III sized stimuli were applied, directly projected on the retina by means of LED projectors and each presented for 200 ms. For mesopic testing, achromatic stimuli (400 to 800 nm) were presented using a 4-2 staircase threshold strategy whilst patients observe the fixation target against a background of 1.27 cd/m² (4 asb).[43,44] The minimum and maximum luminance of the stimuli were 0.08 cd/m² (0.25 asb) and 318 cd/m² (1000 asb), respectively for a dynamic range of 36 dB. For scotopic testing, two different LED types were used - cyan (505 nm) and red (627 nm) - and projected with a 2-1 staircase threshold strategy whilst patients observe the fixation target against a background of <0.0001 cd/m². [43,44] The minimum and maximum luminance of the stimuli were 0.0025 cd/m² (0.008 asb) and 0.25 cd/m² (0.800 asb), representing a dynamic range of 20 dB. The emission spectrum produced by the three types of LEDs is shown in supplement fig. 1.

Irrespective of the type of testing (mesopic, scotopic cyan, scotopic red), the MAIA started examinations with one paracentral test stimulus in each quadrant, (eccentricity of 3°, angular position of 0°, 90°, 180°, 270° respectively). These four threshold values were then used to adjust the initial brightness levels for measuring the remaining test loci in each of the corresponding quadrants. For mesopic function, the testing sensitivity started with a level of 2 dB brighter of the respective initial threshold value using a 4-2 dB full-threshold strategy. For scotopic testing, the testing sensitivity started with a level of 2 dB dimmer than the respective initial threshold, particularly to avoid bleaching, and using a 2-1 dB full-threshold strategy. These differences between mesopic and scotopic FCP were not modifiable as they were included in the presetting of the device. After the first test and for each of the three types of functional testing, subjects were asked to sit back and rest until they were ready to proceed to the second test. The second test was performed using the follow-up mode. For mesopic

function, the starting brightness for each stimulus was automatically set to 2 dB brighter than the previously determined threshold, while the threshold sensitivity for scotopic testing was preset as the initially determined brightness.

FCP was performed by a single experienced examiner in a darkened room. At room light, a brief instruction regarding the testing procedure was given and the correct operation of the subject response trigger was practiced. This was followed immediately by the first mesopic test. The sequence of testing is illustrated in supplement fig. 2. For mesopic testing, the room light was switched off during the examination and briefly turned on between the different mesopic tests (in case of bilateral inclusion, the two eyes firstly underwent one after the other the two mesopic tests). Following mesopic and prior to scotopic testing, eyes underwent 30 mins of dark adaptation (while waiting in the examination room, lights-off; light level < 0.1 lux). The four scotopic tests were then consecutively performed in one eye and finally – if both eyes were included - in the other eye. The fellow eye was always covered with an opaque eye patch. Test reliability was assessed by measuring the frequency of false-positive responses (measured by presentation of suprathreshold stimuli to the optic nerve head). Participants with a false-positive rate greater than 33.3% in any of the tests were excluded from the analysis.

Outcome measures and statistical analyses

Statistical analyses were performed using software environment R.[45] Graphs were created using the ggplot2 library.[46]

The primary outcome measure was the point-wise sensitivity (PWS) intra-session test-retest reliability for the scotopic cyan and scotopic red retinal sensitivity testing in comparison to mesopic FCP in normal eyes, assessed by 95% coefficient of repeatability (CoR) as recommended by Bland and Altman and previously applied by Chen and co-workers for assessment of test-retest variability of mesopic FCP using the Nidek MP1.[47]

Prior to the evaluation of the primary outcome measure, the data was screened for possible systematic learning curve and bleaching effects. The frequency of stimuli with a dense scotoma (failure to respond to the brightest stimulus, i.e. <0 dB) was computed and compared with regard to the three types of testing, the stimulus eccentricity and the test number (2nd vs. 1st test) using mixed-effects logistic regression considering

For all of the following analyses, sensitivity measurements " <0 dB" were incorporated as "0 dB" in analogy to calculation of the different visual field indices.[48] The data was interpreted as equidistant (interval scaled) data according to Weber–Fechner's law.[49–51] The effect of stimulus eccentricity, angular position, and test number (2nd vs. 1st test) on the sensitivity in all three types of testing was assessed using linear mixed-effects models. Hereby, the eccentricity, angular position and test number (2nd vs. 1st test) were considered as fixed effects while considering stimulus position nested within patient as a random effect.

The primary outcome measure (95% coefficient of repeatability, CoR) was determined by multiplying the within-subject standard deviation by 2.77 (1.96 times $\sqrt{2}$). A mixed-design analysis of variance was used to determine the within-subject standard deviation.

Additionally, the frequencies of change in five different categories (no change, $< \pm 2$ dB, $< \pm 4$ dB, $< \pm 6$ dB, $> \pm 6$ dB) were computed as absolute count and cumulative percentage.[52] For visualization of the test-retest reliability, Bland-Altman plots with the mean difference and the 95% limits of agreement (LoA) are provided.[47]

The effective dynamic range and the number of discriminable steps was analyzed as proposed by Wall and associates.[53] The ceiling was defined as value above which less than 0.5% of the values fall in normal eyes. The floor was defined as by the frequency of frequency of 0 dB on the retest (first definition: 5% of retest values equal to 0 dB, second definition: 0 dB most frequent retest value). The definition of the number of discriminable steps from normal to the floor is described in the above mentioned publication in detail.[53]

Furthermore, the stability of fixation was assessed by the area of an ellipse (in degree²) that covered 63% and 95% of fixation points, respectively. After log transformation, the test-retest

222 reliability of the fixation assessment (normal distributed according to a Shapiro–Wilk test)
223 was evaluated using intraclass correlation ICC(3,1). The test-retest reliability of the average
224 reaction time (normal distributed according to a Shapiro–Wilk test) was assessed using the
225 CoR and differences in average reaction times among the types of testing were compared
226 using a one-way analysis of variance followed by a post-hoc Tukey's test.

Results

A total of 32 subjects were recruited between June 2016 and July 2016. FCP testing was not possible in two myopic subjects because ocular media opacities (vitreous floaters) interfered with the automated fundus tracking at the beginning of scotopic testing. These two subjects were excluded from the analysis. The remaining 30 individuals (median age 33.8 years, range 12.8 – 80.1, 14 females) underwent the complete standardized protocol, including duplicate mesopic, scotopic cyan and scotopic red FCP, and were included in the following analysis. In a subset of 17 subjects (median age 25.1 years. range 12.8 - 70.2., 8 females), both eyes were tested.

To avoid invalid inference due to inter-eye (intra-individual) correlation, only the first examined eye of each subject is included in the analysis shown below ($n_{\text{subjects}}=30$, $n_{\text{eyes}}=30$). The stratified analysis (for the second eyes) for subjects with bilateral enrollment is shown in supplement table 1 ($n_{\text{subjects}}=17$, $n_{\text{eyes}}=17$).

Frequency of stimuli indicating a dense scotomata (i.e. <0 dB)

For mesopic testing no stimulus indicating a dense scotoma (i.e. <0 dB) was observed, while for scotopic red testing there was a single stimulus with a dense scotoma (table1). The number of stimuli with a dense scotoma for scotopic cyan testing was 72 (4.89%) for the first and 80 (5.44%) for the second test, respectively. Further analysis revealed that the majority of these dense scotomata was spatially confined to an eccentricity of 0°-1° (first test: 58 of 72 (80.6%), second test 78 of 90 (86.7%)). A mixed-effects logistic regression analysis revealed, that in scotopic cyan testing the likelihood of non-seen stimuli was significantly higher towards the fovea ($p<0.001$). The test number (2nd vs. 1st test) exhibited no statistically significant effect in this regard.

Predictors of point-wise sensitivity

Linear mixed-effects models were used to examine the influence of stimulus location (eccentricity and angular position), test number (2nd vs. 1st test) and possible interactions for all three types of testing thoroughly (table 2). For mesopic testing, the sensitivity was highest at 1° and decreased towards the periphery ($p < 0.001$). Sensitivity along the horizontal meridian was slightly higher than along vertical meridian (nasal ± 0.00 dB, temporal +0.24 dB, superior -0.55 dB, inferior -1.19 dB, $p < 0.001$, supplement fig. 3). A slight improvement between the first and second test was observable (+0.39 dB, $p < 0.001$). Of note, this test-retest improvement was not observable in the second eyes (supplement table 1).

In scotopic cyan testing, the sensitivity increased significantly towards the periphery (1° +0.47 dB, 3° +5.76 dB, 5° +6.25 dB, 7° 6.65 dB, $p < 0.001$). Overall, the superior hemifield exhibited a significantly higher sensitivity than the inferior hemifield (fig. 2). The observation was most pronounced in the vertical meridian (superior +1.85 dB, inferior -0.75 dB). There was a significant decrease in sensitivity between the first and second test (-3.1 dB, $p < 0.001$) that exhibited a significant interaction with the eccentricity – i.e. the decrease was most pronounced at 0° and 1° and minor at 3° and 7° ($p < 0.001$). At 5° no relevant effect was observable.

In scotopic red testing, the measured sensitivity was highest at 1° and decreased towards the periphery. The horizontal meridian exhibited a slightly higher sensitivity than the vertical meridian (nasal ± 0.00 dB, temporal -0.18 dB, superior -0.74 dB, inferior -1.35 dB, $p < 0.001$, fig. 2). The sensitivity was slightly decreased in the second test (-0.4 dB, $p < 0.001$).

Coefficient of Repeatability (CoR) and frequency of change in PWS

The Coefficient of Repeatability (CoR), i. e. the value below which the absolute differences between two measurements would lie with 0.95 probability, was 4.75 dB for mesopic, 5.26 dB for scotopic cyan and 4.06 dB for scotopic red testing (table 3). For mesopic and scotopic

red testing, the CoR were in the same range across all eccentricities. The CoRs for scotopic cyan testing were lower at 3°, 5° and 7° than at 0°-1° (4.27 dB, 4.24 dB, 4.26 dB vs. 7.36 dB). Table 4 lists the frequency of change in PWS. For all three types of testing, the cumulative percentage of stimulus points within ± 2 dB repeatability was > 75%.

Bland-Altman plots and 95% limits of agreement (LoA)

The 95% limits of agreement (LoAs) according to Bland-Altman statistics ranged from -5.08 dB (95% CI, -5.29 – -4.86) to 4.29 dB (95% CI, 4.08 – 4.50) for mesopic testing. Graphical analysis showed no ceiling or floor effects (fig. 3A). For scotopic red testing, similar LoAs, ranging from -3.59 dB (95% CI, -3.77 – -3.41) to 4.38 dB (95% CI, 4.20 – 4.56), were calculated. As opposed to mesopic testing, a marked ceiling effect was detectable for scotopic red FCP (fig. 3C). Finally, the widest limits of agreement as well as signs for both floor and ceiling effects were found for scotopic cyan testing with LoAs ranging from -4.72 dB (95% CI, -4.96 – -4.49) to 5.65 dB (95% CI, 5.41 – 5.88) and also showing several outliers (fig. 3B).

Effective dynamic range

Last, the effective dynamic range and the number of discriminable steps were analyzed as proposed by Wall and associates.[53] The ceiling (defined as value above less than 0.5% of measurements fall in normal eyes) was 32 dB for mesopic, 20 dB for scotopic cyan and 20 dB for scotopic red testing (supplement table 4). The floor was not determinable using normal eyes for mesopic and scotopic red testing due to the absence of scotomata. In scotopic cyan testing, the fovea and parafovea served as physiologic scotoma allowing the assessment of floor effects. The floor for scotopic cyan testing was depending on the definition at 9 dB (5% of retest values 0 dB) or at 3 dB (0 dB most frequent retest value). The number of discriminable steps was 4 (supplement fig. 2).

Stability of fixation

The analysis of the test-retest reliability for stability of fixation (after log transformation to reduce skew) disclosed moderate to high agreement between the two tests for all three types of retinal sensitivity testing with ICC(3,1) values of 0.891 for mesopic, 0.671 for scotopic cyan and 0.879 for scotopic red testing, respectively (supplement table 3). The difference in fixation stability between the mesopic tests and the scotopic cyan ($p = 0.029$) and scotopic red ($p = 0.077$) tests was found to be statistically significant and almost statistically significant respectively. No overall difference in fixation stability was observed between the scotopic cyan and red tests ($p = 0.920$) (fig. 4A; supplement table 4).

Reaction time

The test-retest reliability for the average reaction times was high with CoR values of 126 ms for mesopic, 146 ms for scotopic cyan and 137 ms for scotopic red testing (supplement table 5). Mesopic testing showed the shortest reaction time (mean \pm SD: 486 ± 46 ms, pooled data of first and second test), followed by scotopic red (619 ± 50 ms) and then scotopic cyan (712 ± 52 ms) retinal sensitivity assessment ($p < 0.001$), (fig. 4B; supplement table 6).

Testing duration

The mean duration for the first test was 6 minutes and 24 seconds for mesopic, 8 minutes and 8 seconds for scotopic cyan and 7 minutes and 45 seconds for scotopic red testing (Table 5). For both scotopic cyan and scotopic red testing, the test duration was significantly

328 shorter for the second as compared to the first test (table 5). No reduction in test duration
329 was observed between the two mesopic tests. As outlined above, the strategy with regards
330 to the initial brightness levels for test loci in the second test and using the follow-up mode
331 was different between mesopic (2 dB above the value determined in the first test) and
332 scotopic assessments (initial brightness equal to value determined in the first test).

Discussion

The current study demonstrates that the recently introduced modified MAIA device allows for a practical, reliable and meaningful assessment of scotopic FCP in normal subjects.

Compared to mesopic testing that has been available for several years and which was also evaluated in this study, assessment of scotopic function is obviously more complex and more time-consuming since it requires dark adaptation and examination in a complete darkened room and, thus, is more prone to confounding factors. The inability to ensure a meaningful tracking of the fundus under dark-adapted conditions (due to the reduced SLO laser power) in two subjects with vitreous floaters (who had to be excluded) indicates that the assessment of scotopic function may not be feasible in some normal subjects. On the other hand, the number of false-positive answers was below 33% in all subjects. Furthermore, the average test duration for each of the two scotopic tests was below 9 min - without the time for dark adaptation – and, therefore on average only approximately 30% longer than mesopic testing.

The more elaborated protocol with two different types of scotopic testing (cyan and red) has been employed in the device for a more accurate discrimination of specific photoreceptor-mediated function.[39–41] Indeed, several results of this study are in accordance with the view that the scotopic cyan test condition largely reflects rod function, while the scotopic red condition derives from both rod and cone photoreceptors. First, the reported anatomical distribution of rod cells corresponds to the topographic differences of median sensitivity values for scotopic cyan testing (Figure 2) that showed the highest median sensitivity at 7°, a centripetal decrease, including a sharp decline between 3° and 1° and slight but consistent higher values towards the superior retina.[54,55] By contrast, this topographic pattern was not detectable by scotopic red testing. Secondly, the decrease in sensitivity in the second as compared to the first scotopic cyan test at 0° and 1° was most likely related to mild bleaching effects on the sensitive rods due to the fixation target. The third line of evidence comes from differences in reaction time among tests. Although no direct instruction was given to the subjects to respond as quickly as possible, reaction time variability was generally small (Fig

4B). Longer reaction times for both scotopic as compared to mesopic test conditions are likely to reflect the considerable lower stimulus intensities employed and the increasing recruitment of rod photoreceptors in those tests.[56] For scotopic red, subjects commonly reported disparate color percepts depending on the degree of eccentricity. In line with the retinal distribution of L-cone and rod distribution, the stimuli appeared to be “red” at 0°-5° and “white” at 7°.[54,55] It would be conceivable that this difference in perception caused a “degree of uncertainty” and thus longer reaction times. For scotopic cyan, the even longer reaction times at similar levels of luminance most likely additionally reflects the so-called slow rod pathway.[57] It is well-known that rod latency is depending on the degree of dark adaptation. While under conditions of equal cone and rod adaption, the cone-rod latency has been shown to be 8-20 ms in humans, rod signaling is delayed at higher levels of luminance.[57–61] This so-called slow rod pathway is presumably mediated via rod bipolar and All-amacrine cells.[62] The published cone-rod latency of 60-80 ms under these conditions is of a similar magnitude as the difference of 96 ms of average reaction time between scotopic cyan and scotopic red testing.[57]

Several results of the statistical analysis of the intra-session duplicate testing suggest that a reliable assessment of both types of scotopic function is feasible using this instrument. These include the strong agreement for the PWS retest reliability and that > 75% of threshold values were within ± 2 dB of repeatability. No major learning effect was found in the current protocol, involving a brief instruction of the procedure and practice of the response trigger function. A minor inter-session learning effect might have been the reason for the higher count of 10 versus 2 dense scotomata (out of 564 stimuli) in the first as compared to the second scotopic cyan testing at an eccentricity of 3°.

The slightly higher CoR for mesopic testing in this study (4.75 dB) as compared to the previous reports by Wu et al. (3.74 dB, using the MAIA) might be caused by differences in the testing protocol.[63] While the former included training with presenting of test stimuli prior to actual testing and used a grid of 37 loci, the current study - in addition to no dedicated

training (see above) - used a grid with 49 stimuli (longer duration). In fact, the twelve most eccentric test points (in areas that were not included by Wu et al.) showed the lowest overall CoR for mesopic testing.

The measures for reliability for both scotopic testing were mostly within the same range as compared to the duplicate assessment of mesopic testing in this study. One exception was the assessment of scotopic cyan testing at 0°-1° degree. In this area, very high threshold values were consistently detected which would indicate a floor effect. A likely reason for this phenomenon is the absence or relatively small amount of rods in this area and/or bleaching effects of the fixation target interfering with the detection of low luminance levels.[64] The latter hypothesis is supported by the reduction in sensitivity between the first and second scotopic cyan test, which was confined to 0° and 1° (i.e. the location of the fixation target). Moreover, the CoR for scotopic cyan testing at 0°-1° has to be interpreted with caution, since it is most likely an overestimation of the repeatability due to the floor truncation and the inclusion of non-seen stimuli "<0 dB" as "0 dB". Therefore, assessment and interpretation of rod function based on scotopic cyan sensitivity would be challenging in the central retina. In fact, the central area is excluded in the preset grid of the device and was only included in the customized grid of this study to specifically assess scotopic testing in this area also. Compared to mesopic testing, the LoAs and their graphical analysis suggest that scotopic red is more prone to ceiling effects, while both floor and ceiling effects must be particularly considered for scotopic cyan testing.

Indeed, the physical dynamic range of a perimetric test does not necessarily represent the effective dynamic range (EDR), which can be truncated due to the limits of perception (floor) and due to the test-retest reliability (floor).[53] The analyses in analogy to Wall and associates disclosed that the effective ceiling for mesopic testing was 32 dB instead of 36 dB. The floor was not quantifiable in normal eyes for mesopic and scotopic red testing due to the lack of scotomata. In scotopic cyan testing the floor was estimable – however precaution must be taken due to the observed bleaching effects at 0° and 1°. With re-dark adaption

between the test-retest, the floor of the EDR for scotopic cyan testing would have been most likely lower. The number of discriminable steps for scotopic cyan testing was equivalent in comparison to previously published data for standard automated perimetry using the Humphrey Field Analyzer (24-2, Goldmann III) and frequency-doubling technology perimetry.[53]

The high ICC for the BCEA indicates high test-retest repeatability for the assessment of fixation stability. The lower stability of fixation compared to mesopic testing is in accordance with several previous publications that have consistently reported about the difficulties of ensuring a stable fixation under scotopic condition.[20,32] Some subjects states that the fixation target disappeared during scotopic testing and re-appeared after blinking or short fixation movements. This observation could be explained by the Troxler-effect (especially for low-contrast targets).[65]

The current study is limited by inclusion of only normal subjects and the limited number of subjects. In addition, longer test-retest intervals (e.g. days or weeks) were not assessed. Furthermore, the test-retest reliability would be most likely higher with repeated dark adaption in between the tests.

In summary, we performed a systematic analysis of scotopic retinal sensitivity assessment in normal subjects using the recently introduced modified MAIA device. We demonstrated a strong agreement for intra-session PWS reliability for both scotopic cyan and scotopic red testing. Results for assessment of reliability were mostly comparable and in the same range as mesopic testing. In accordance with anatomical and functional differences between rods and cones, including the topographic distribution, the unequal signaling pathway and the different range of luminance sensitivity, the results indicate that particular attention to various phenomena must be considered when interpreting scotopic test results. This includes the increased susceptibility for floor and ceiling effects and the occurrence of severe outliers. Although scotopic testing was more demanding, it was feasible in almost all subjects in the study. Finally, several observations strongly suggest that the testing of scotopic cyan

sensitivity largely reflects rod function. Further hardware and software updates of the device might become available and may improve the performance, reliability and the ability to interpret testing results. This may offer the opportunity for more refined structural-functional analyses in a broad spectrum of retinal diseases and the potential use as a functional outcome measure in future interventional clinical trials.

448 **Acknowledgements**

449 We are grateful for the technical support of Carlo Pellizzari from Centervue SpA.

References

1. Helmholtz H von: Beschreibung des Augenspiegels; in : Beschreibung Eines Augen-Spiegels. Springer, 1851, pp 28–34.
2. Trantas NG: Applications et resultats d'un moyen simple d'examen de la photosensibilite de la retine. Bull Soc Ophthalmol Fr 1955;55:499–513.
3. Meyers MP: The use of the Visuscope. Am J Ophthalmol 1959;47:677–681.
4. Awaya S: Spot scotometry; A new method to examine scotomas under direct ophthalmoscopy by using visuscope (euthyscope). Jpn J Ophthalmol 1972;16:144–157.
5. Watzke RC, Allen L: Subjective slitbeam sign for macular disease. Am J Ophthalmol 1969;68:449–453.
6. Kani K, Ogita Y: Fundus controlled perimetry; in : Third International Visual Field Symposium Tokyo, May 3–6, 1978. Springer, 1979, pp 341–350.
7. Webb RH, Hughes GW: Scanning laser ophthalmoscope. IEEE Trans Biomed Eng 1981;488–492.
8. Timberlake GT, Mainster MA, Webb RH, Hughes GW, Trempe CL: Retinal localization of scotomata by scanning laser ophthalmoscopy. [Internet]. Invest Ophthalmol Vis Sci 1982 Jan 1;22:91–97.
9. Holz FG, Jubb C, Fitzke FW, Bird AC, Pope FM: Dark adaptation and scotopic perimetry over “peau d’orange” in pseudoxanthoma elasticum. Br J Ophthalmol 1994 Jan;78:79–80.
10. Schmitz-Valckenberg S, Ong EL, Rubin GS, Peto T, Tufail A, Egan CA, et al.: Structural and functional changes over time in MacTel patients. Retina 2009;29:1314–1320.
11. Berson EL, Gouras P, Gunkel RD: Rod responses in retinitis pigmentosa, dominantly inherited. Arch Ophthalmol 1968;80:58–67.
12. Berson EL, Gouras P, Gunkel RD, Myrianthopoulos NC: Rod and cone responses in sex-linked retinitis pigmentosa. Arch Ophthalmol 1969;81:215–225.
13. Arden GB, Carter RM, Hogg CR, Powell DJ, Ernst WJ, Clover GM, et al.: Rod and cone activity in patients with dominantly inherited retinitis pigmentosa: comparisons between psychophysical and electroretinographic measurements. Br J Ophthalmol 1983;67:405–418.
14. Lorenz B, Gyürüs P, Preising M, Bremser D, Gu S, Andrassi M, et al.: Early-onset severe rod–cone dystrophy in young children with RPE65 mutations. Invest Ophthalmol Vis Sci 2000;41:2735–2742.
15. Curcio CA, Medeiros NE, Millican CL: Photoreceptor loss in age-related macular degeneration. Invest Ophthalmol Vis Sci 1996;37:1236–1249.
16. Shelley EJ, Madigan MC, Natoli R, Penfold PL, Provis JM: Cone degeneration in aging and age-related macular degeneration. Arch Ophthalmol 2009;127:483–492.
17. Lim LS, Mitchell P, Seddon JM, Holz FG, Wong TY: Age-related macular degeneration. Lancet 2012;379:1728–1738.
18. Owsley C, Jackson GR, Cideciyan A V, Huang Y, Fine SL, Ho AC, et al.: Psychophysical evidence for rod vulnerability in age-related macular degeneration. Invest Ophthalmol Vis Sci 2000;41:267–273.
19. Jackson GR, Owsley C, Curcio CA: Photoreceptor degeneration and dysfunction in aging and age-related maculopathy. Ageing Res Rev 2002;1:381–396.
20. Steinberg JS, Fitzke FW, Fimmers R, Fleckenstein M, Holz FG, Schmitz-Valckenberg S: Scotopic and photopic microperimetry in patients with reticular drusen and age-related macular degeneration. JAMA Ophthalmol 2015;133:690–697.
21. Steinberg JS, Saßmannshausen M, Fleckenstein M, Fimmers R, Oishi A, Holz FG, et al.: Correlation of Partial Outer Retinal Thickness With Scotopic and Mesopic Fundus-

- Controlled Perimetry in Patients With Reticular Drusen. *Am J Ophthalmol* 2016;168:52–61.
22. Jackson GR, Edwards JG: A short-duration dark adaptation protocol for assessment of age-related maculopathy. *J Ocul Biol Dis Infor* 2008;1:7–11.
23. Jackson GR, Scott IU, Kim IK, Quillen DA, Iannaccone A, Edwards JG: Diagnostic Sensitivity and Specificity of Dark Adaptometry for Detection of Age-Related Macular Degeneration Sensitivity and Specificity of AdaptDx. *Invest Ophthalmol Vis Sci* 2014;55:1427–1431.
24. Nebbioso M, Barbato A, Pescosolido N: Scotopic microperimetry in the early diagnosis of age-related macular degeneration: preliminary study. *Biomed Res Int* 2014;2014.
25. Rosli Y, Bedford SM, James AC, Maddess T: Photopic and scotopic multifocal pupillographic responses in age-related macular degeneration. *Vision Res* 2012;69:42–48.
26. Smith VC, Pokorny J, Diddie KR: Color matching and the Stiles–Crawford effect in observers with early age-related macular changes. *JOSA A* 1988;5:2113–2121.
27. Holz FG, Gross-Jendroska M, Eckstein A, Hogg CR, Arden GB, Bird AC: Colour contrast sensitivity in patients with age-related Bruch’s membrane changes. *Ger J Ophthalmol* 1995;4:336–341.
28. Gerth C, Hauser D, Delahunt PB, Morse LS, Werner JS: Assessment of multifocal electroretinogram abnormalities and their relation to morphologic characteristics in patients with large drusen. *Arch Ophthalmol* 2003;121:1404–1414.
29. Gerth C, Delahunt PB, Alam S, Morse LS, Werner JS: Cone-mediated multifocal electroretinogram in age-related macular degeneration: progression over a long-term follow-up. *Arch Ophthalmol* 2006;124:345–352.
30. Jacobson SG, Voigt WJ, Parel J-M, Apathy PP, Nghiem-Phu L, Myers SW, et al.: Automated light-and dark-adapted perimetry for evaluating retinitis pigmentosa. *Ophthalmology* 1986;93:1604–1611.
31. Scholl HPN, Bellmann C, Dandekar SS, Bird AC, Fitzke FW: Photopic and scotopic fine matrix mapping of retinal areas of increased fundus autofluorescence in patients with age-related maculopathy. *Invest Ophthalmol Vis Sci* 2004;45:574–583.
32. Crossland MD, Luong VA, Rubin GS, Fitzke FW: Retinal specific measurement of dark-adapted visual function: validation of a modified microperimeter. *BMC Ophthalmol* 2011;11:1.
33. Midena E, Vujosevic S, Convento E, Cavarzeran F, Pilotto E: Microperimetry and fundus autofluorescence in patients with early age-related macular degeneration. *Br J Ophthalmol* 2007;91:1499–1503.
34. Salvatore S, Fishman GA, McAnany JJ, Genead MA: Association of Dark-Adapted Visual Function with Retinal Structural Changes in Patients with Stargardt Disease. *Retina* 2014 May;34:989–995.
35. Wu Z, Jung CJ, Ayton LN, Luu CD, Guymer RH: Test–Retest Repeatability of Microperimetry at the Border of Deep Scotomas Test–Retest Repeatability of Microperimetry. *Invest Ophthalmol Vis Sci* 2015;56:2606–2611.
36. Bowl W, Stieger K, Lorenz B: Fundus-controlled two-color dark adaptometry with the Microperimeter MP1. *Graefes Arch Clin Exp Ophthalmol* 2015 Jun;253:965–972.
37. Bowl W, Lorenz B, Jager M, Friedburg C: Improving detection of mild loss of retinal light increment sensitivity at the posterior pole with the microperimeter MP1. *Invest Ophthalmol Vis Sci* 2013 Jul;54:4666–4674.
38. Rohrschneider K, Springer C, Bültmann S, Vö HE: Microperimetry—comparison between the micro perimeter 1 and scanning laser ophthalmoscope—fundus perimetry. *Am J Ophthalmol* 2005;139:125–134.

39. Wentworth HA: A quantitative study of achromatic and chromatic sensitivity from center to periphery of the visual field. *Psychol Monogr* 1930;40:i.
40. Sloan LL: Instruments and technics for the clinical testing of light sense: III. An apparatus for studying regional differences in light sense. *Arch Ophthalmol* 1939;22:233–251.
41. Sloan LL: The threshold gradients of the rods and the cones: in the dark-adapted and in the partially light-adapted eye. *Am J Ophthalmol* 1950;33:1077–1089.
42. Midena E: *Microperimetry and multimodal retinal imaging*. Springer, 2014.
43. Johnson CA, Chauhan BC, Shapiro LR: Properties of staircase procedures for estimating thresholds in automated perimetry. *Invest Ophthalmol Vis Sci* 1992;33:2966–2974.
44. Békésy G v: A new audiometer. *Acta Otolaryngol* 1947;35:411–422.
45. Team RC: R: A language and environment for statistical computing. R Foundation for Statistical Computing, Vienna, Austria. 2013 2014;
46. Wickham H: *ggplot2: elegant graphics for data analysis*. Springer Science & Business Media, 2009.
47. Bland JM, Altman D: Statistical methods for assessing agreement between two methods of clinical measurement. *Lancet* 1986;327:307–310.
48. Flammer J: The concept of visual field indices. *Graefe's Arch Clin Exp Ophthalmol* 1986;224:389–392.
49. Bouguer P: *Traité d'optique sur la gradation de la lumière*. 1760.
50. Weber EH: *De Pulsu, resorptione, auditu et tactu: Annotationes anatomicae et physiologicae...* CF Koehler, 1834.
51. Fechner GT: *Das psychische Maß. na*, 1858.
52. Chen FK, Patel PJ, Xing W, Bunce C, Egan C, Tufail AT, et al.: Test–retest variability of microperimetry using the Nidek MP1 in patients with macular disease. *Invest Ophthalmol Vis Sci* 2009;50:3464–3472.
53. Wall M, Woodward KR, Doyle CK, Zamba G: The effective dynamic ranges of standard automated perimetry sizes III and V and motion and matrix perimetry. *Arch Ophthalmol* 2010;128:570–576.
54. Curcio CA, Sloan KR, Kalina RE, Hendrickson AE: Human photoreceptor topography. *J Comp Neurol* 1990;292:497–523.
55. Curcio CA, Sloan KRJ, Packer O, Hendrickson AE, Kalina RE: Distribution of cones in human and monkey retina: individual variability and radial asymmetry. *Science* 1987 May;236:579–582.
56. Kohfeld DL: Simple reaction time as a function of stimulus intensity in decibels of light and sound. *J Exp Psychol* 1971;88:251.
57. Cao D, Zele AJ, Pokorny J: Linking impulse response functions to reaction time: Rod and cone reaction time data and a computational model. *Vision Res* 2007;47:1060–1074.
58. Sun H, Pokorny J, Smith VC: Rod-cone interactions assessed in inferred magnocellular and parvocellular postreceptoral pathways. *J Vis* 2001;1:5.
59. MacLeod DI: Rods cancel cones in flicker. *Nature* 1972;
60. Van den Berg T, Spekreijse H: Interaction between rod and cone signals studied with temporal sine wave stimulation. *JOSA* 1977;67:1210–1217.
61. Barbur JL: Reaction-time determination of the latency between visual signals generated by rods and cones. *Ophthalmic Physiol Opt* 1982;2:179–185.
62. Sharpe LT, Stockman A: Rod pathways: the importance of seeing nothing. *Trends Neurosci* 1999;22:497–504.
63. Wu Z, Ayton LN, Guymer RH, Luu CD: Intrasection Test–Retest Variability of Microperimetry in Age-Related Macular Degeneration Test–Retest Variability of

- 602 Microperimetry in AMD [Internet]. . Invest Ophthalmol Vis Sci 2013 Nov 8;54:7378–
603 7385.
- 604 64. Stockman A, Langendörfer M, Smithson HE, Sharpe LT: Human cone light adaptation:
605 From behavioral measurements to molecular mechanisms. J Vis 2006;6:5.
- 606 65. Troxler D: Über das Verschwinden gegebener Gegenstände innerhalb unseres
607 Gesichtskreises. Ophthalmol Bibl 1804;2:1–53.
608