









OPEN ACCESS

# Why is safety in intrapartum electronic fetal monitoring so hard? A qualitative study combining human factors/ergonomics and social science analysis

Guillaume Lamé <sup>1,2</sup> Elisa Giulia Liberati <sup>2</sup> Aneurin Canham,<sup>3</sup> Jenni Burt <sup>2</sup> Lisa Hinton <sup>4</sup> Tim Draycott,<sup>5</sup> Cathy Winter,<sup>5</sup> Francesca Helen Dakin,<sup>4</sup> Natalie Richards,<sup>2</sup> Lucy Miller,<sup>6</sup> Janet Willars <sup>7</sup> Mary Dixon-Woods <sup>2</sup>

► Additional supplemental material is published online only. To view, please visit the journal online (<http://dx.doi.org/10.1136/bmjqs-2023-016144>).

For numbered affiliations see end of article.

## Correspondence to

Professor Mary Dixon-Woods, THIS Institute, Department of Public Health and Primary Care, University of Cambridge, Cambridge CB1 8RN, UK; [director@thisinstitute.cam.ac.uk](mailto:director@thisinstitute.cam.ac.uk)

Received 21 March 2023  
Accepted 16 September 2023



© Author(s) (or their employer(s)) 2023. Re-use permitted under CC BY-NC. No commercial re-use. See rights and permissions. Published by BMJ.

**To cite:** Lamé G, Liberati EG, Canham A, *et al.* *BMJ Qual Saf* Epub ahead of print: [please include Day Month Year]. doi:10.1136/bmjqs-2023-016144

## ABSTRACT

**Background** Problems in intrapartum electronic fetal monitoring with cardiotocography (CTG) remain a major area of preventable harm. Poor understanding of the range of influences on safety may have hindered improvement. Taking an interdisciplinary perspective, we sought to characterise the everyday practice of CTG monitoring and the work systems within which it takes place, with the goal of identifying potential sources of risk.

**Methods** Human factors/ergonomics (HF/E) experts and social scientists conducted 325 hours of observations and 23 interviews in three maternity units in the UK, focusing on how CTG tasks were undertaken, the influences on this work and the cultural and organisational features of work settings. HF/E analysis was based on the Systems Engineering Initiative for Patient Safety 2.0 model. Social science analysis was based on the constant comparative method.

**Results** CTG monitoring can be understood as a complex sociotechnical activity, with tasks, people, tools and technology, and organisational and external factors all combining to affect safety. Fetal heart rate patterns need to be recorded and interpreted correctly. Systems are also required for seeking the opinions of others, determining whether the situation warrants concern, escalating concerns and mobilising response. These processes may be inadequately designed or function suboptimally, and may be further complicated by staffing issues, equipment and ergonomics issues, and competing and frequently changing clinical guidelines. Practice may also be affected by variable standards and workflows, variations in clinical competence, teamwork and situation awareness, and the ability to communicate concerns freely.

**Conclusions** CTG monitoring is an inherently collective and sociotechnical practice. Improving it will require accounting for complex system interdependencies, rather than focusing solely on discrete factors such as individual technical proficiency in interpreting traces.

## INTRODUCTION

Problems with intrapartum electronic fetal monitoring (EFM) during labour

## WHAT IS ALREADY KNOWN ON THIS TOPIC

- ⇒ Intrapartum electronic fetal monitoring has been repeatedly identified as a major area of preventable harm in maternity care
- ⇒ Improvement efforts have typically focused on training individual clinicians in cardiotocography trace interpretation, but have had limited impact in improving care overall

## WHAT THIS STUDY ADDS

- ⇒ Intrapartum electronic fetal monitoring should be considered as a practice fraught with sociotechnical complexity, where people, tasks, tools and technology, organisation and environment interact to affect outcomes.
- ⇒ Fetal monitoring work systems may be suboptimally designed or fail to work well in practice, with substantial gaps between work as imagined and work as done.

have repeatedly been identified as a major area of preventable harm.<sup>1-3</sup> However, it has been remarkably difficult to secure improvement.<sup>3 4</sup> One explanation for the limited progress lies in the possibility that improvement efforts may have been misdirected, typically focused primarily on individual behaviours in relation to cardiotocography (CTG) interpretation<sup>5</sup> rather than the wider range of influences on risks, including systems, processes, decisions and actions. Examination of the everyday work of EFM is essential

**HOW THIS STUDY MIGHT AFFECT RESEARCH, PRACTICE OR POLICY**

- ⇒ Updated guidelines or more training are unlikely, on their own, to improve electronic fetal monitoring. Improvement initiatives need to consider a broader set of factors and their interdependencies to create the conditions for better cardiotocography monitoring.
- ⇒ Some issues may be amenable to local action, but others will require large-scale coordination.

to understanding system factors and to identifying potential targets for intervention for improvement. To address this need, we undertook an interdisciplinary study drawing on both human factors/ergonomics (HF/E) and social science perspectives to examine the routine operation of local systems where EFM takes place.<sup>6</sup>

Increasingly widely promoted,<sup>7 8</sup> the discipline of HF/E is characterised by a focus on the components of work systems and their interactions.<sup>9</sup> One of its key commitments is to understanding how tasks are conducted in practice (work ‘as done’), rather than how they are documented or prescribed (work ‘as imagined’).<sup>10–13</sup> The Systems Engineering Initiative for

Patient Safety (SEIPS, [box 1](#)) is an important example of an HF/E model that has been developed specifically for healthcare settings. Analysis using the SEIPS is intended to identify the range of influences on how work is done, and thus help in guiding system redesign to support better performance and safety.

A second important and complementary methodological approach to understanding the systems where work takes place involves the social sciences, including studies using qualitative methods such as ethnographic observations and interviews. Work in this tradition has been especially important in affording deep insight into cultural and organisational features of work settings,<sup>14 15</sup> the role of technologies<sup>15–17</sup> and influences on professional work and practices.<sup>18 19</sup>

In this article, we combine these approaches—HF/E and social science—to characterise current work systems for EFM in hospital maternity units with the aim of understanding influences on risk.

**METHODS**

This article is an output of the IMMO (IMproving the practice of intrapartum electronic fetal heart rate MONitoring with cardiotocography for safer child-birth)<sup>20</sup> programme and its reporting is guided by the Standards for Reporting Qualitative Research.<sup>21</sup> We conducted a study combining observations and interviews in three maternity units, selected purposively based on their size (annual number of deliveries) and their location in different nations of the UK. We recruited one small unit (less than 2000 births/year), one medium unit (2000–5000 births/year) and one large unit (more than 5000 births/year) across rural, urban and metropolitan settings. Because HF/E system analyses require setting clear boundaries to the system being investigated, we limited the study focus to clinical situations where the decision to initiate EFM had already taken place, and up to the point of escalation, without examining clinical interventions in response to possible fetal deterioration.

Data were collected between April 2019 and March 2020. In each unit, the data were collected both by a human factors specialist and an experienced social scientist. Each researcher completed up to seven ~8-hour days in each maternity unit, conducting day and night observations and interviews with professional staff. Observations focused on the labour ward, with occasional time spent in the antenatal ward and antenatal assessment clinic to understand upstream processes. Observers sometimes spent long periods of time in one room, or moved through multiple areas, or shadowed a professional. To avoid imposition of predetermined categories, no structured observation form was used, but the human factor experts used the SEIPS 2.0 framework to guide their observations.<sup>22–24</sup>

Interview participants were selected purposively to represent diversity in professional background (midwives and doctors), and different seniority and

**Box 1 Description of the SEIPS model**

The Systems Engineering Initiative for Patient Safety (SEIPS)<sup>22</sup> takes a whole-system view that identifies the elements of systems using a structured framework and examines how they interact<sup>23</sup> to influence processes and, ultimately, outcomes. In version 2.0,<sup>23</sup> SEIPS contains three types of elements:

- ⇒ The *work system* is composed of interacting components: persons (eg, care providers or patients), tasks, tools and technology (eg, medical devices or information technology (IT) systems), organisation and internal environment (ie, architectural layout and ambient characteristics, such as noise and light). The work system is affected by influences from the external environment.
- ⇒ Components of the work system interact to produce *processes* of professional, patient and collaborative work at various levels: physical, cognitive and social/behavioural.
- ⇒ These processes affect the *outcomes* for patients, healthcare workers and organisations. The outcomes can be desirable or not, and close or distant in time.

This model integrates well-established work system modelling principles with Donabedian’s<sup>50</sup> classic structure-process-outcome framework for healthcare quality. Importantly, the work system and the processes are affected by previous outcomes through feedback loops, by which individuals and organisations adjust over time to maintain or improve performance.

professional experiences, though availability of staff limited the extent to which selection was possible. Interviews were tape-recorded and transcribed verbatim. Ethnographic data were captured in the form of brief, anonymised, in situ field notes and later written up in full. Debriefing sessions of the research team were recorded, transcribed and treated as data alongside the field notes.

We conducted two types of analysis in parallel. The first used the SEIPS 2.0 model to describe the work systems we observed and the interactions of people, tasks, technology, organisational structures, environment and external influences. Human factor specialists coded their observations as well as all interviews, and then related these codes to the SEIPS categories. Second, social scientists coded their observations and all interviews based on the constant comparative method,<sup>25</sup> involving initial open coding of a selection of interviews and observations to guide development of a coding frame that was then applied to the whole dataset. Data were analysed at the site level first, then across sites. QSR NVivo V.12 software was used to support coding, management and retrieval of data.

The team was multidisciplinary from the onset, and the design and conduct of the project involved obstetricians, midwives, HF/E specialists and social scientists. We held biweekly meetings between researchers from the HF/E and social science streams while planning and conducting the research to ensure integration of design and sharing of emerging findings. We also organised multidisciplinary debriefing sessions every time a researcher (HF/E or social scientists) had completed their visits in one site. After HF/E specialists and social scientists had completed their initial parallel analysis of the data, we discussed themes and findings collectively to arrive at a synthesis. This prompted further interrogation of the data, until we arrived at a shared understanding.

Written informed consent was obtained from participants to interviews. Oral consent was sought from healthcare professionals and those in labour and their partners before observations. Maternity service user representatives were involved in reviewing the study documents (protocol, leaflets, posters), which were revised following their comments.

## RESULTS

Across the three participating units, we conducted 325 hours of observation and completed 23 interviews (table 1). In presenting our findings below, we first identify and briefly describe the key SEIPS 2.0 elements (figure 1), with illustrative quotations presented in the online supplemental file. We then integrate our human factors and our social science analyses to explain how these system components interact, and how these interactions affect the functionality of the system as a whole.

### System description

In all three units, those in labour were cared for primarily by midwives on the labour suite. For those deemed ‘high risk’, EFM using CTG was used to track changes in the fetal heart rate. In all three units, midwives were expected to monitor the paper record produced by CTG machines (known as ‘the trace’) to assess changes and classify deviations from normal using prespecified criteria. Their interpretations were intended to be recorded on the traces, often using a dedicated ‘sticker’ template that was attached to the trace (figure 2). Midwives were also expected to seek second opinions (known as ‘fresh eyes’<sup>26</sup>) from professional colleagues, usually another midwife, every hour, and to make decisions about whether any potentially concerning fetal heart rate patterns required escalation (eg, to a more senior midwifery colleague or an obstetrician). Obstetricians, including doctors in training, specialist and associate specialist doctors, and consultants, were expected to review traces with other clinical information as part of routine rounds and on request by midwives, and to make decisions about whether further clinical intervention (eg, expedite birth) might be needed.


### Staffing levels and workload

SEIPS elements: organisation, people, internal and external environment  
 Staffing levels, including sufficient midwives, obstetricians, and other staff, and in the right role configurations, were of critical importance. However, we identified—for example, through posters indicating planned and actual staffing available, as well as through discussions with staff members and observations of

**Table 1** Data collected across three sites

|        | Hours of observation   | Interviewees (n)                     |
|--------|--|--------------------------------------|
| Site 1 | 58 hours 35 min (social science)<br>65 hours (human factors)                       | 11 midwives and 3 doctors: 14 total  |
| Site 2 | 70 hours 15 min (social science)<br>39 hours 30 min (human factors)                | 3 midwives and 2 doctors: 5 total    |
| Site 3 | 42 hours 30 min (social science)<br>49 hours 30 min (human factors)                | 2 midwives and 2 doctors: 4 total    |
| Total  | 171 hours 20 min (social science)<br>154 hours (human factors)<br>325 hours 20 min | 16 midwives<br>7 doctors<br>23 total |

## Why is electronic fetal monitoring so hard?



**Figure 1** Systems Engineering Initiative for Patient Safety (SEIPS) modelling of the work system, processes and outcomes for electronic fetal monitoring, with illustrative examples. The internal environment is the physical environment in the unit (its layout and architectural characteristics, including, for example, lighting and noise). The external environment covers contextual elements beyond the control of the organisation. Processes are aggregations of tasks into sequences of actions and decisions. Outcomes include those proximal (close in time) and distal (emerging later) that affect the people being cared for (here, the woman, the baby and possibly the partner), the professionals delivering care (here, midwives and doctors) and the organisation (the NHS trust). Adaptation describes how work system elements dynamically adjust to perceived gaps between actual and ideal outcomes. In this figure, the names of actors in the external environment reflect the situation of a site in England. Abbreviations: CQC: Care Quality Commission; CTG: cardiotocography; DHSC: Department of Health and Social Care; IT: information technology; NHS, National Health Service. See online supplemental file. This figure is based on Holden *et al*<sup>23</sup> and Holden and Carayon.<sup>51</sup>

handovers—that units were often understaffed or lacking sufficient numbers of staff in the appropriate roles with the appropriate proficiencies. Individual professionals were often dealing with very high workloads, which meant that it was not always possible for a midwife to be consistently present in the room with labouring women. Sometimes, staff who lacked specific experience or expertise were asked to cover vacancies or temporary absences.

Clinical workloads meant that midwives were sometimes unable to attend multidisciplinary CTG meetings, which was where traces from their unit were discussed with the medical team. Staff also reported difficulties in taking breaks, adding to the fatigue and pressure.

When you're actually there on the coalface it's really, really tough because you might have another patient, you might have answered the phone, you might have


a coordinator on who you don't really value their opinion, [...] you haven't eaten, you haven't drunk, you haven't been to the bathroom – they are all the extra factors that make the job really hard. (Midwife, Site 2)

### Availability of well-designed and well-functioning equipment in appropriate spaces

SEIPS elements: tools and technology, external and internal environment, organisation

Well-designed, well-functioning equipment is fundamental to effective EFM. In none of the three sites was this requirement consistently met. In two sites, there were too few CTG machines. In all three, machine design and maintenance were often suboptimal. Equipment malfunctions were common and appeared to be related to both defects inherent in the machines themselves and to poor maintenance (table 2). Routine

Documentation proforma for intrapartum CTG interpretation (based on FIGO 2015)

| Feature  | Reassuring (acceptable)   | Non-reassuring   | Abnormal  |
|--|---|--|---|
| <b>Baseline rate (bpm)</b>   | Baseline 110 – 160 bpm<br>Rate:   | Baseline 100 – 109 bpm for more than 10 minutes<br>Rate:<br>Baseline rate more than 160bpm<br>Rate:  | Baseline less than 100 bpm<br>Rate:   |
| <b>Variability (bpm)</b>   | Variability of 5 bpm or more (but less than 25 bpm)   | Variability less than 5 bpm for less than 50 minutes<br>Variability more than 25 bpm for less than 30 minutes<br>Sinusoidal pattern lasting for less than 30 minutes | Variability less than 5 bpm for more than 50 minutes<br>Variability more than 25 bpm for more than 30 minutes<br>Sinusoidal pattern lasting for more than 30 minutes                                |
| <b>Accelerations</b>   | Present   | Comments:  |   |
| <b>Decelerations</b>   | None<br>True early decelerations  | NON V-shaped/U-shaped (atypical) Variable decelerations (with reduced variability within deceleration) with more than 50% of contractions & for less than 30 minutes | NON V-shaped/U-shaped (atypical) Variable decelerations (with reduced variability within contraction) with more than 50% of contractions & for more than 30 minutes                                 |
|  | V-shaped (typical) Variable decelerations exhibiting a symmetrical rapid drop and rapid recovery to the baseline & all other features of the CTG are normal       | Late decelerations (U shaped and/or with reduced variability) with more than 50% of contractions for less than 30 minutes  | Late decelerations (U-shaped and/or with reduced variability) with more than 50% of contractions for more than 30 minutes   |
|  | NON V-shaped/U-shaped (atypical) Variable or Late decelerations (U-shaped and/or with reduced variability within deceleration) with less than 50% of contractions | Single prolonged deceleration lasting more than 3 minutes, but less than 5 minutes   | Single prolonged deceleration for more than 5 minutes<br>(A prolonged deceleration of less than 80 bpm with reduced variability & lasting more than 5 minutes is often associated with hypoxia)     |
| <b>Dilatation:</b>   | Liquor colour:  | Maternal pulse:  | Contractions :10 Gestation:   |
| <b>Reason for CTG:</b>   | Other risk factors:   |  |   |
| <b>Opinion</b>   | <b>Normal CTG</b><br>(All four features reassuring)<br>No intervention necessary  | <b>Suspicious CTG</b><br>(One non-reassuring feature)<br>Low probability of hypoxia<br>Correct reversible causes (refer to algorithm & EFM interpretation guidance)  | <b>Pathological CTG</b><br>(Two or more non-reassuring or one or more abnormal features)<br>High probability of hypoxia – urgent action required (refer to algorithm & EFM Interpretation guidance) |
| <b>Action taken:</b><br>(Always consider medical/clinical & obstetric circumstances when interpreting CTG & determining actions)   |   |  |   |
| Date:  | Time:   | Signature:   | Status:   |
| <p><b>Fresh Eyes Opinion</b>  <b>at least 2 hourly</b> <b>I agree with opinion? YES / NO</b> <b>If opinion different complete new sticker</b></p> <p>Date: ..... Time: ..... Signature: ..... Status: .....</p> |   |  |   |

**Figure 2** Example of 'sticker' for recording the interpretation of cardiocography (CTG) traces (not from one of the sites we visited, in the interests of confidentiality). FIGO, International Federation of Gynaecology and Obstetrics.

problems included paper running out or jamming, sometimes at key moments. At one site, we identified two occasions when monitors were not printing the right time or even the right date on the paper.

The midwife went back in at this point and when trying to print off the trace accidentally turned the machine off due to confusion over the options. So this

was turned off and the trace was lost and then they had to run the test again. (Observations, Site 2)

Challenges also occurred because of variations between machines, linked to how decisions were made externally by manufacturers and internally by procurement decisions, including legacy procurement. Brands,

**Table 2** Issues of availability and functionality of electronic fetal monitoring equipment

| Category                | Problem  | Consequence   |
|-------------------------|--|---|
| Technical features      | Difficulties in getting and keeping a good signal                    | Midwives holding the electrode in place (thus diverting resources from other tasks), or asking the mother to keep the electrode in place (limiting her movements) |
|                         | Wireless monitors picking up signals from other monitors             | Risk of inaccurate appreciation of the situation  |
|                         | Electrode battery not lasting long enough                            | Time-consuming efforts to find alternatives, risk of delays in monitoring   |
|                         | Showing up mother's heart rate instead of fetus's                    | Risk of inaccurate appreciation of the situation  |
| Human-machine interface | Inadvertently change printing speed                                  | Risk of misinterpreting the trace   |
|                         | Inadvertently switching off the monitor                              | Loss of time and data   |
|                         | Different grids on screen and on paper                               | Risk of misinterpreting the trace   |
| Supplies                | Missing parts (eg, leads for fetal scalp electrodes)                 | Time-consuming efforts to find alternatives, risk of delays in monitoring   |
|                         | Missing consumables (eg, paper)                                      | Time-consuming efforts to find consumables, risk of delays in monitoring  |
| Procurement             | Different machines in a unit printing in different colours or scales | Risk of misinterpreting the trace   |
|                         | Insufficient number of monitors                                      | Time-consuming efforts to find equipment, risk of delays in monitoring  |
| Maintenance and set-up  | Paper jamming  | Loss of time and data, delays or interruption to monitoring   |
|                         | Wrong time or date   | Wrong timeline when reviewing the trace   |
| Physical layout         | Bulky monitors in small rooms  | Cramped workspaces  |

makes and vintages of machines varied within and between sites. How they needed to be handled varied, as did their outputs.

Monitors were variably located, sometimes centrally and sometimes in the room with the person in labour. Central monitoring allowed professionals to keep an overview of the situation across the labour suite and to discuss potentially alarming traces efficiently. But having the CTG machine in the room with the woman brought potential benefits, including the ability to assess the whole picture more easily and to communicate effectively.

#### Clarity about criteria for assessing deterioration

SEIPS elements: external and internal environment, people, organisation, tasks

Maternity professionals need to have shared understanding about the criteria to be used to classify CTG traces, along with clear thresholds for making decisions about what might constitute a pattern of deterioration and its significance. Across the three sites, however, we identified lack of consensus on the relevant criteria and thresholds and how they should be interpreted and actioned. Different guidelines were used across the three units, including those of the National Institute for Health and Care Excellence<sup>27</sup> and the International Federation of Gynaecology and Obstetrics.<sup>28</sup> A 'physiological' approach<sup>29</sup> (without official standing, but promoted through external training events) was also being used by some staff. There was some evidence of hybridised use of criteria, where staff drew on elements of different approaches simultaneously.

Each unit had its own system aimed at supporting review of traces. One example was the mnemonic DR C BRAVADO (Define Risk; C: Contractions; BRA: Baseline RATE; V: Variability; A: Accelerations; D: Decelerations; O: Overall impression), which was used to structure thinking about trace interpretation. Some elements of the guidance were encoded in stickers that were attached to CTG traces and used by staff to record their interpretation of the trace, along with any actions taken or required. Guidance was also available on hospital intranets. However, ambiguities about the applicable criteria and thresholds were pervasive. Participants also reported struggling to keep up with changes to guidelines; they had to break old habits and form new ones, as well as deploy changes to their operational systems.

Interviewer: So how do you find these stickers?

Midwife: There are slightly newer ones now, so they are better; but CTG, the evaluation is always changing; and like you've heard on discussion there, they're planning on changing them again. (Midwife, Site 3)

Lack of standardisation contributed to variation and confusion within and between units, so different colleagues were not always working with the shared understanding of what constituted a concerning

pattern. Doctors in training frequently moved between units and were sometimes more familiar with a different conceptual model to interpret a trace from that used by their current colleagues. Further variation in the interpretation of traces arose from different individuals' approaches to reviewing traces and making judgements.

#### Accurate records of interpretations of CTG traces

SEIPS elements: organisation, people, tools and technology

Identifying possible fetal deterioration requires examining CTG traces over time, so making accurate clinical interpretations and recording them over the course of labour is a fundamental work system requirement. However, it was not reliably met in the units in the study. We identified different norms of recording between and within units, with different professionals and teams using different approaches in different ways. For instance, all three sites were using stickers that functioned both as a way of structuring interpretation by providing prompts, and as a way of documenting the interpretation (figure 2). Stickers were more likely to be used by midwives. Doctors were more likely to write in the notes or directly onto the trace, but even this was variable. Doctors did not always record their interpretation (or planned action) in real time when called for a review, sometimes creating difficulties in later decision-making. Though many professionals valued the stickers, they were also at risk of becoming something of a 'tick box' exercise in practice.

When it becomes complicated, when there are issues, is when it's least used the sticker for interpretation. So, when we should be using it the most we use it the least [...] It's all lovely for a nice, normal [CTG], we've got lovely notes for normal. When it all goes wrong is when the notes are the least used. (Senior midwife, Site 2)

#### Support for sound decision-making to recognise deterioration

SEIPS elements: people, organisation, task

Fatigue, familiarity, inaccurate heuristics, cognitive load and limited knowledge are all recognised contributors to flawed decision-making (eg, inaccurate interpretation of a CTG trace). We identified all of these factors to varying degrees across the units. In part to address these challenges, all three units were using a 'fresh eyes' system, which required midwives to ask another professional every hour to independently review and interpret the CTG trace and record this interpretation on a sticker. Many valued the system, seeing it as an important support.

Especially on a night I find when you're looking at trace and you start to lose that objectivity with it because it's darkish, you're tired. But at least we have someone coming in each hour to check, and if you're getting your proper break relief you again have

another person in for a little while just checking you're okay. (Midwife, Site 2)

'Fresh eyes' was not straightforward in practice, however. It was not always easy to find a colleague to review every hour, especially one who had not seen the trace before. It seemed also that two individuals might inadvertently collude to produce a consensual interpretation, rather than the second individual actively challenging the first interpretation. For example, the midwife in the room often began to present their interpretation before the 'fresh eyes' midwife had a chance to look at it, potentially creating a framing effect. One incident that we observed involved a trace that had no accelerations and variability below 5 bpm. It was classified as 'normal' by the midwife in the room, even though such a pattern should be seen as concerning. The first midwife gave their interpretation to the 'fresh eyes' midwife as soon as they entered the room, and the second midwife classified the trace as 'normal' too.

### Clinical competence in detecting fetal deterioration

SEIPS elements: person, task, organisation

Detection of fetal deterioration is a highly skilled activity requiring a very high level of clinical competence, including the ability to appreciate the whole clinical picture in context and to use clinical intuition appropriately. Both individual technical proficiency and collective competence (including clear role differentiation, location of authority, effective teamwork, coordination and distributed cognition) were necessary to achieve this. Informal learning opportunities were provided through role modelling, conversations between professionals and norm-setting behaviours. All three sites had mandatory annual online CTG training and annual in-person training sessions and dedicated meetings (including debriefing sessions), but attendance was impacted by understaffing.

The different roles and duties of midwives and obstetricians were important to the ability to function collectively. Midwives tended to be constantly present with the person in labour and attentive to the trace over long periods, allowing them to detect patterns over time and to gain deep knowledge of a woman's clinical condition and preferences to identify traces for escalation. Obstetricians' more episodic engagement enabled different perspectives and forms of clinical knowledge to be brought to bear. Doctors' behaviours included pulling out the whole trace to observe the wider picture and see trends or changes in the context of women's physiology and medical and obstetric history.

Regular meetings were held to review decisions that had been taken during previous shifts with the aim of enhancing organisational learning and improvement. However, their quality and openness relied on psychological safety,<sup>30</sup> and some participants indicated that new or junior members of the team could feel judged

or singled out, thus potentially undermining the benefits. It was also difficult for midwives to attend these meetings, despite their intended multidisciplinary, because there was often no cover for their clinical work.

If you've got a patient in labour or here, you know, you can't because the meeting can take a good hour and you can't really go and sit down. But they are good, I think they are very, very good. (Midwife, Site 1)

### Communicating concerns and escalation processes

SEIPS elements: person, task, organisation

Communicating concerns about possible fetal deterioration is essential to making decisions about whether any intervention may be needed. Communication practices depended on quality of decision-making at the point of care, shared understanding and ability to command attention. None of these were straightforward to achieve. In all units, midwives were expected to escalate concerns in the event of a trace classified as 'pathological' and overtly pathological traces were generally escalated without further ado. However, most traces were not obviously pathological and instead required complex interpretations and judgements that were not easy to make.

Well, when the CTG is normal you can be very reassured. When the CTG is abnormal, it can mean a myriad of different things. (...) The decision-making is so complex, (...) I think it only really works in the hands of expert interpretation. (Midwife, Site 2)

So there was a point in the night... a woman had a suspicious CTG, but the midwife said that it was OK okay for now, no action needed, because the results of the fetal blood samples were reassuring (...) But also, almost at the back of their minds, you could see that they were a bit more alert. A registrar and the midwife said, outside the woman's room, 'we all know something's going to happen there'... There was a sense that complications were about to happen. (Observation, Site 1)

Once a decision was made to escalate, the workflow was variable, but typically involved contacting the senior midwife or labour suite coordinator, who could then engage an experienced senior doctor. Sometimes, however, in particularly concerning situations, midwives escalated straight to a doctor and informed the coordinator subsequently. In taking escalation actions, judgements about the competence and decision-making ability of colleagues sometimes featured. For example, experienced midwives were sometimes reluctant to defer to the views of doctors whom they saw as inexperienced, worried that escalation would be frustrated by an incorrect interpretation.

Soon after the ward round the specialty trainee came to speak to me and he was a bit flustered or perhaps frustrated was the way to describe it. And he was

saying the midwives were putting pressure on him to perform a C-section on the woman in labour. He felt a C-section was unnecessary [...] but as the labour wasn't progressing the midwives wanted a section. (Observations, Site 3)

On the other hand, a major influence on escalation decisions was fear of a cascade of intervention that might be triggered by involving obstetric colleagues. Once a fetal trace began to be escalated, it became increasingly difficult to recover the possibility of a vaginal birth without intervention. Participants explained that there was a very high false positive rate for abnormal CTGs; expediting birth by caesarean in all situations where the CTG was deemed not normal would result in unnecessary intervention, and reduce resource for situations where intervention was needed.

### Clearly designed workflows, teamwork and situation awareness

SEIPS elements: task, organisation, internal environment

Crucial to being able to mobilise an appropriate response were clearly defined processes, clinical workflows, roles and tasks, for example, to define responsibilities for decision-making, ensuring action and maintaining situation awareness. However, they were variably evident in the units we studied. In all three, the labour ward coordinator played a critical role in monitoring situations, anticipating need and coordinating responses. This required ensuring that this individual and other key colleagues were kept briefed when concerns were developing in relation to a birth.

In these decisions and practices, professional hierarchies and experiences, relationships between different professions, personality and personal histories all mattered. Disagreements over the right course of action to take were common, although they varied in content.

When new people come along, I find the most stressful time, new doctors, the changeover. I loathe that time. Because when you know someone well and they say '[the trace] is OK', you're thinking, OK, they know. But when someone new comes and says it's okay I'm thinking, do I trust you? It takes a while to build that. (Midwife, Site 3)

## DISCUSSION

This study of everyday work on three maternity units affords important understanding of the problems to be solved to improve EFM.<sup>31 32</sup> Our account of 'work as done',<sup>11</sup> combining SEIPS and social science analysis, showed that EFM is a profoundly collective process, involving multiple interactions between people, tasks, tools and technology, organisation, culture and behaviour,<sup>23</sup> all constituting a single clinical microsystem.<sup>33</sup> We also found that the work systems for EFM appear to be poorly optimised for safety. A particular challenge is that the diverse elements of these work systems are

intimately inter-related and interdependent, involving multiple actors who must coordinate activity, achieve shared understanding, make sound decisions in the face of uncertainty and competing considerations, demonstrate respect for those in labour and their partners and take appropriate actions at the right time. Structural challenges are pervasive, for example, in relation to staffing, equipment design and supply, and buildings and facilities.<sup>34 35</sup> Further complexity is introduced by the contested nature of the evidence underpinning some practices. These findings confirm that EFM is a sociotechnical system involving multiple interdependent elements that may interact in complex ways,<sup>36</sup> and that improvement efforts focused solely on individual practice (eg, competence in CTG interpretation) are unlikely to succeed on their own. Improving electronic monitoring will require a multifaceted, integrated systems approach that draws on evidence-based strategies, including work systems that are purposefully designed, better and clearer guidance and processes, better tools and technologies, and better understanding of people, their roles and skills.

Some hazards we identified are structural in character and highly consequential, reflecting how work systems for fetal monitoring sit within larger subsystems that influence what can be achieved locally. For example, sufficient supplies of functioning equipment that reliably record the fetal heart rate are a basic work system requirement,<sup>37</sup> but one that was not consistently met in the units in our study and could in principle be addressed locally. Redesign of tools and technology so that machines are ergonomically designed, with standard interfaces (eg, printing and display format) across machines used in each site, should be a priority. However, it will require large-scale external as well as internal action, with procurement processes<sup>38</sup> and manufacturers each playing an important role—so will require large-scale coordination and leadership to resolve.<sup>39</sup>

Large-scale coordination will also be required to address many of the structural issues on maternity units relating to sufficient staff in appropriately configured roles, which we found mattered greatly for teamwork, coordination, decision-making and action. For example, detection and appropriate communication of a concern and mobilising an effective response may be difficult in conditions of excessive task loads, availability of staff, fatigue and cognitive burden. Yet, these structural issues were often beyond the scope of control of individual units to resolve easily, given challenges in payment models, staff recruitment and retention, and national policies on workforce.

One obvious target for improvement is the multiplicity of guidelines used between and within trusts, which would benefit from rationalisation and coherence at national level,<sup>40</sup> to avoid the confusion and inefficiency we found in units. Standardised formal risk assessment tools that have been codesigned with

staff and patients and offer clear triggers, thresholds and processes for calling for help are likely to support control of risk, situation awareness and multidisciplinary coordination.<sup>41</sup> Leadership at system level will be needed to achieve the necessary consensus.

Other hazards are, perhaps, more tractable to action in the short term. For example, though clinical skills are not, on their own, a comprehensive solution to safety risks in fetal monitoring, they remain fundamental as a key sociotechnical factor. Accordingly, high-quality multiprofessional training is critical,<sup>36 40 42–44</sup> as are regular multiprofessional meetings.<sup>40</sup> A further target for improvement concerns interprofessional relationships on labour wards, which have also been identified in previous research as highly consequential for practices relating to EFM.<sup>18 19</sup> Our analysis made clear that just as important as the technical skills of interpretation are ‘non-technical’ skills such as situation awareness, teamwork, communication, decision-making and so on. The available evidence suggests that improvement in clinical and litigation outcomes can be secured in all of these areas through high-quality, high-fidelity training.<sup>44–46</sup>

Our findings contribute to long-standing debates about EFM,<sup>47</sup> including questions about whether the equivocal nature of the evidence about whether it ‘works’ arises because of its inherent limitations as a technology or because of challenges in implementation. An important strength of our study is our use of combined HF/E and social science approaches, and its offering of non-clinical ‘outsider’ insights into the nature of the work systems underlying EFM.<sup>12 48</sup> The SEIPS 2.0 model proved a flexible framework in this diagnosis stage, compared with more structured HF/E approaches.

This study also has limitations. We did not have access to performance or clinical data, meaning that we could not quantify the relationship between the process variations we identified and clinical outcomes or risk control. We collected data in three maternity units. Although we sampled for diversity (size, location), we may have missed important hazards that could exist in other units. More interviews were carried in site 1 than in the others, which could bias our results (although observations were more equally distributed). Finally, we stopped our investigation at the point of escalation, without looking at timely response and ‘rescue’. These are also crucial influences on safety<sup>49</sup> and will be the focus of future work.

## CONCLUSIONS

Improving EFM is not simply a matter of producing better guidelines or more rigorous individual training. Instead, it requires understanding that CTG is a practice fraught with sociotechnical complexity and interdependencies, and is profoundly collective in character, underpinned by relationships, expertise and skill. A broad set of influences on safety and their

interdependencies must be considered in creating the conditions for improvement. While some issues may be tractable to local action, others will require large-scale coordination.

### Author affiliations

<sup>1</sup>Laboratoire Génie Industriel, CentraleSupélec, Gif-sur-Yvette, France

<sup>2</sup>Department of Public Health and Primary Care, THIS Institute (The Healthcare Improvement Studies Institute), University of Cambridge, Cambridge, UK

<sup>3</sup>Emergo by UL, Southampton, UK

<sup>4</sup>Nuffield Department of Primary Care Health Sciences, University of Oxford, Oxford, UK

<sup>5</sup>North Bristol NHS Trust, Bristol, UK

<sup>6</sup>University Division of Anaesthesia, Cambridge University Hospitals NHS Foundation Trust, Cambridge, UK

<sup>7</sup>Department of Health Sciences, University of Leicester, Leicester, UK

**Twitter** Lisa Hinton @LisaHinton4, Francesca Helen Dakin @dakinfrancesca and Mary Dixon-Woods @MaryDixonWoods

**Acknowledgements** We thank the staff and patients of the participating maternity units for their cooperation and support in the conduct of the study. We thank Matt Woodward of THIS Institute for his support of this work.

**Contributors** MDW is the guarantor for this study

**Funding** This study was funded by The Health Foundation (RHZF/001-RG88620).

**Competing interests** TD: research and innovation lead for PROMPT Maternity Foundation.

**Patient consent for publication** Not applicable.

**Ethics approval** This study involves human participants and was approved by the West Midlands-South Birmingham Research Ethics Committee (reference number: 18/WM/0292). Participants gave informed consent to participate in the study before taking part.

**Provenance and peer review** Not commissioned; externally peer reviewed.

**Data availability statement** No data are available. No data are available owing to the need to preserve confidentiality.

**Supplemental material** This content has been supplied by the author(s). It has not been vetted by BMJ Publishing Group Limited (BMJ) and may not have been peer-reviewed. Any opinions or recommendations discussed are solely those of the author(s) and are not endorsed by BMJ. BMJ disclaims all liability and responsibility arising from any reliance placed on the content. Where the content includes any translated material, BMJ does not warrant the accuracy and reliability of the translations (including but not limited to local regulations, clinical guidelines, terminology, drug names and drug dosages), and is not responsible for any error and/or omissions arising from translation and adaptation or otherwise.

**Open access** This is an open access article distributed in accordance with the Creative Commons Attribution Non Commercial (CC BY-NC 4.0) license, which permits others to distribute, remix, adapt, build upon this work non-commercially, and license their derivative works on different terms, provided the original work is properly cited, appropriate credit is given, any changes made indicated, and the use is non-commercial. See: <http://creativecommons.org/licenses/by-nc/4.0/>.

### ORCID iDs

Guillaume Lamé <http://orcid.org/0000-0001-9514-1890>

Elisa Giulia Liberati <http://orcid.org/0000-0003-4981-1210>

Jenni Burt <http://orcid.org/0000-0002-0037-274X>

Lisa Hinton <http://orcid.org/0000-0002-6082-3151>

Janet Willars <http://orcid.org/0000-0002-7886-3223>

Mary Dixon-Woods <http://orcid.org/0000-0002-5915-0041>

## REFERENCES

- 1 Rowe R, Draper ES, Kenyon S, *et al.* Intrapartum-related perinatal deaths in births planned in Midwifery-led settings in great Britain: findings and recommendations from the Esmie confidential enquiry. *BJOG* 2021;128:1712–3.
- 2 Robertson L, Knight H, Prosser-Snelling E, *et al.* Each baby counts in 2018 – lessons learned and future directions. *Obstetrics, Gynaecology & Reproductive Medicine* 2018;28:253–7.
- 3 Yau CWH, Leigh B, Liberati E, *et al.* Clinical negligence costs: taking action to safeguard NHS Sustainability. *BMJ* 2020;368:m552.
- 4 Brocklehurst P, Field D, Greene K, *et al.* Computerised interpretation of fetal heart rate during labour (INFANT): a randomised controlled trial. *The Lancet* 2017;389:1719–29.
- 5 Kelly S, Redmond P, King S, *et al.* Training in the use of Intrapartum electronic fetal monitoring with Cardiotocography: systematic review and meta-analysis. *BJOG* 2021;128:1408–19.
- 6 Ghaferi AA, Osborne NH, Birkmeyer JD, *et al.* Hospital characteristics associated with failure to rescue from complications after Pancreatectomy. *J Am Coll Surg* 2010;211:325–30.
- 7 WHO. Human factors, in technical series on safer primary care. Geneva, 2016.
- 8 National Research Council. *Health care comes home: the human factors*. Washington, D.C: National Academies Press. xi, 2011: 189.
- 9 Karsh B-T, Holden RJ, Alper SJ, *et al.* A human factors engineering paradigm for patient safety: designing to support the performance of the Healthcare professional. *Qual Saf Health Care* 2006;15 Suppl 1(Suppl 1):i59–65.
- 10 Woods DD. How to design a safety organization: test case for resilience engineering (Ch.19). In: Hollnagel E, Woods DD, Levenson N, eds. *Ashgate*. 2006: 315–26.
- 11 Blandford A, Furniss D, Vincent C. Patient safety and interactive medical devices: realigning work as imagined and work as done. *Clin Risk* 2014;20:107–10.
- 12 Smith AF, Plunkett E. People, systems and safety: resilience and excellence in Healthcare practice. *Anaesthesia* 2019;74:508–17.
- 13 Nemeth C, O'Connor M, Klock PA, *et al.* Discovering Healthcare cognition: the use of cognitive artifacts to reveal cognitive work. *Organization Studies* 2006;27:1011–35.
- 14 McKnight M. The information seeking of on-duty critical care nurses: evidence from participant observation and in-context interviews. *J Med Libr Assoc* 2006;94:145–51.
- 15 Waring JJ, Bishop S. Lean Healthcare: rhetoric, ritual and resistance. *Social Science & Medicine* 2010;71:1332–40.
- 16 Woolgar S, Neyland D. Mundane governance. In: *Mundane governance: Ontology and accountability*. 2013: OUP Oxford.
- 17 Timmermans S, Berg M. The practice of medical technology. *Sociology of Health & Illness* 2003;25:97–114.
- 18 Small KA, Sidebotham M, Fenwick J, *et al.* The social Organisation of decision-making about Intrapartum fetal monitoring: an institutional Ethnography. *Women Birth* 2023;36:281–9.
- 19 Brydges R, Nemoy L, Campbell DM, *et al.* We can't just have a casual conversation": an institutional Ethnography-informed study of work in labour and birth. *Social Science & Medicine* 2021;279:113975.
- 20 Lamé G, Liberati E, Burt J, *et al.* Improving the practice of Intrapartum electronic fetal heart rate monitoring with Cardiotocography for safer childbirth (the IMMO programme): protocol for a qualitative study. *BMJ Open* 2019;9:e030271.
- 21 O'Brien BC, Harris IB, Beckman TJ, *et al.* Standards for reporting qualitative research: A synthesis of recommendations. *Acad Med* 2014;89:1245–51.
- 22 Carayon P, Schoofs Hundt A, Karsh B-T, *et al.* Work system design for patient safety: the SEIPS model. *Qual Saf Health Care* 2006;15(suppl 1):i50–8.
- 23 Holden RJ, Carayon P, Gurses AP, *et al.* SEIPS 2.0: a human factors framework for studying and improving the work of Healthcare professionals and patients. *Ergonomics* 2013;56:1669–86.
- 24 Werner NE, Ponnala S, Doutcheva N, *et al.* Human factors/Ergonomics work system analysis of patient work: state of the science and future directions. *Int J Qual Health Care* 2021;33(Supplement\_1):60–71.
- 25 Charmaz K. *Constructing grounded theory: a practical guide through qualitative analysis*. London: Sage, 2006.
- 26 Donnelly L, Hamilton L. A 'fresh eyes approach' *Midwives* 2012;15:44–5.
- 27 NICE. Intrapartum care for healthy women and babies. In: *National Institute for Health and Care Excellence: London*. 2014: 89.
- 28 Ayres-de-Campos D, Spong CY, Chandraran E, *et al.* FIGO consensus guidelines on Intrapartum fetal monitoring: Cardiotocography. *Int J Gynaecol Obstet* 2015;131:13–24.
- 29 Chandraran E, ed. *Handbook of CTG interpretation: from patterns to physiology*. Cambridge: Cambridge University Press, 2017: 256.
- 30 Edmondson AC. Learning from failure in health care: frequent opportunities, pervasive barriers. *Qual Saf Health Care* 2004;13 Suppl 2(Suppl 2):ii3–9.
- 31 Rae A, Provan D, Aboelssad H, *et al.* A manifesto for reality-based safety science. *Safety Science* 2020;126:104654.
- 32 Altaf S, Oppenheimer C, Shaw R, *et al.* Practices and views on fetal heart monitoring: a structured observation and interview study. *BJOG* 2006;113:409–18.
- 33 Mohr JJ, Batalden PB. Improving safety on the front lines: the role of clinical Microsystems. *Qual Saf Health Care* 2002;11:45–50.
- 34 Greenhalgh T, Wherton J, Shaw S, *et al.* Infrastructure Revisited: an Ethnographic case study of how health information infrastructure shapes and constrains technological innovation. *J Med Internet Res* 2019;21:e16093.
- 35 Tsoukas H, Vladimirou E. What is Organisational knowledge *J Management Stud* 2001;38:973–93.
- 36 Magro M. Five years of cerebral palsy claims - A thematic review of NHS resolution data. In: *NHS Resolution*. London, UK, 2017: 92.
- 37 Each baby counts: 2018 progress report. Royal college of Obstetricians and Gynaecologists London RCOG; 2018.
- 38 Healthcare Safety Investigation Branch. *Suitability of equipment and technology used for continuous fetal heart rate monitoring*. HSIB: Reading, 2021: 55.
- 39 Dixon-Woods M, Pronovost PJ. Patient safety and the problem of many hands. *BMJ Qual Saf* 2016;25:485–8.
- 40 Ockenden D, Ockenden Report. Emerging findings and recommendations from the independent review of maternity services at the Shrewsbury and Telford hospital NHS trust: our first report following 250 clinical reviews; 2020. 48.
- 41 Lin DM, Ghaferi AA. Four failure-to-rescue design themes to improve rescue. *ASA Monitor* 2021;85:30–1.

- 42 NHS Resolution. The early notification scheme progress report: collaboration and improved experience for families - an overview of the scheme to date together with thematic analysis of a cohort of cases from year 1 of the scheme, 2017–2018. In: *NHS Resolution*. London, 2019.
- 43 Royal College of Obstetricians and Gynaecologists,. Each baby counts: 2015 summary report London RCOG; 2017.
- 44 NHS England. *Saving Babies' Lives version two - A care bundle for reducing perinatal mortality*. NHS England: London, 2019: 72.
- 45 Royal College of obstetricians and Gynaecologists,. Each baby counts: 2019 progress report. London RCOG; 2020.
- 46 Draycott T, Sibanda T, Owen L, *et al*. Does training in obstetric emergencies improve neonatal outcome *BJOG* 2006;113:177–82.
- 47 Macones GA, Hankins GDV, Spong CY, *et al*. The 2008 National Institute of child health and human development workshop report on electronic fetal monitoring: update on definitions, interpretation, and research guidelines. *Journal of Obstetric, Gynecologic & Neonatal Nursing* 2008;37:510–5.
- 48 Waterson P, Catchpole K. Human factors in Healthcare: welcome progress, but still scratching the surface. *BMJ Qual Saf* 2016;25:480–4.
- 49 Mackintosh N, Sandall J. The social practice of rescue: the safety implications of acute illness Trajectories and patient Categorisation in medical and maternity settings. *Sociol Health Illn* 2016;38:252–69.
- 50 Donabedian A. *The Definition of Quality and Approaches to Its Assessment [= Explorations in Quality Assessment and Monitoring*. Ann Arbor: Health Administration Press, 1980: 8–11.
- 51 Holden RJ, Carayon P. SEIPS 101 and seven simple SEIPS tools. *BMJ Qual Saf* 2021;30:901–10.