

Highlights

- 1000 probabilistic models created with varying unicompartmental surgical cuts
- Tibia fracture risk increased for excessive resection or posterior vertical cut depth
- Regression fitted finite element results with 90% correlation using 3 cut parameters
- Predicted instrumentation modifications may reduce the likelihood of fracture

Minimising Tibial Fracture after Unicompartmental Knee Replacement: A Probabilistic Finite Element Study

Elise C Pegg^{1*}, Jonathan Walter², Darryl D D’Lima³, Benjamin J Fregly⁴, Harinderjit S Gill¹, David W Murray⁵

¹ Department of Mechanical Engineering, University of Bath, Bath, UK

² CED Technologies, Inc. Jacksonville, Florida, USA

³ Shiley Center for Orthopaedic Research & Education, Scripps Clinic, La Jolla, California, USA.

⁴ Department of Mechanical Engineering, Rice University, Houston, Texas, USA

⁵ Nuffield Department of Orthopaedics, Rheumatology and Musculoskeletal Sciences, University of Oxford, Oxford, UK

*Corresponding author:

Dr Elise Pegg

Address: University of Bath,
Department of Mechanical Engineering,
Claverton Down,
Bath
BA2 7AY

Email: e.c.pegg@bath.ac.uk

Running title: Tibial fracture after knee replacement

Word count: 250 words abstract. 3718 words main text.

24 **Abstract**

25 *Background*

26 Periprosthetic tibial fracture after unicompartmental knee replacement is a challenging
27 post-operative complication. Patients have an increased risk of mortality after fracture, the
28 majority undergo further surgery, and the revision operations are less successful.
29 Inappropriate surgical technique increases the risk of fracture, but it is unclear which
30 technical aspects of the surgery are most problematic and no research has been performed on
31 how surgical factors interact.

32 *Methods*

33 Firstly, this study quantified the typical variance in surgical cuts made during
34 unicompartmental knee replacement (determined from bones prepared by surgeons during an
35 instructional course). Secondly, these measured distributions were used to create a
36 probabilistic finite element model of the tibia after replacement. A thousand finite element
37 models were created using the Monte Carlo method, representing 1000 virtual operations, and
38 the risk of tibial fracture was assessed.

39 *Findings*

40 Multivariate linear regression of the results showed that excessive resection depth and
41 making the vertical cut too deep posteriorly increased the risk of fracture. These two
42 parameters also had high variability in the prepared synthetic bones. The regression equation
43 calculated the risk of fracture from three cut parameters (resection depth, vertical and
44 horizontal posterior cuts) and fit the model results with 90% correlation.

45 *Interpretation*

46 This study introduces the application of a probabilistic approach to predict the aetiology of
47 fracture after unicompartmental knee replacement and has quantified the potential importance
48 of surgical saw cut variations for the first time. Targeted changes to operative technique can
49 now be considered to seek to reduce the risk of periprosthetic fracture.

50 Keywords: Knee, Bone, Fracture, Unicompartmental, Finite Element.

1 Introduction

Periprosthetic tibial fracture after unicompartmental knee replacement (UKR) is a severe complication which can be challenging to treat and manage [1]. Fracture is associated with increased mortality and significant morbidity, and is increasing in incidence [2]. Of the cases of tibial fracture after UKR reported in the literature [1; 3-11], approximately half of the fractures occurred during the operation, and half occurred within 6 weeks post-operatively. More than 50% of the reported case studies end with revision to total knee replacement, requiring removal of the cruciate ligament(s) and leading to reduced knee function [1; 3-11].

The reported incidence of tibial fracture after UKR ranges from 0.8% [1] to 5.0 % [11]. The absolute number of patients at risk of fracture is rising [2] as a result of increasing numbers of UKRs being performed each year [12], greater life expectancy [13], higher cases of osteoporosis [14], and increasing patient activity [15]. It is, therefore, important to identify which aspects of UKR surgery put patients at the greatest risk of fracture, so that the operative technique can be optimised to minimise the occurrence of this serious complication.

The issue of periprosthetic fracture has been reported in several different unicompartmental knee designs, so it appears the issue is not design-specific [1; 3-4; 6; 8; 11], though one study suggested cementless components are at greater risk [16]. There is uncertainty in the literature regarding the most important surgical risk factors for tibial fracture after UKR. The surgical errors that have been proposed to cause tibial plateau fracture include:

- excessive depth of surgical cuts made for the tray, tray keel, or pegs [1; 4; 9; 17; 18]
- too many holes in the cortex for alignment guides [6; 10]
- perforation of the tibial cortex [5]

- under-sizing of the tibial tray [1; 3]
- use of excessive force when impacting the plateau [1]
- excessive removal of bone [1].

However, of these studies, Clarius *et al.* were the only authors to base their conclusions on experimental evidence and showed that extended vertical cuts reduced the force required to cause tibial fracture by 30% [18].

Finite element analysis (FEA) is a useful tool for predicting bone fracture, and it has been applied most commonly to fractures of the femoral neck. Schileo *et al.* proposed a Risk Of Fracture (ROF) criterion (Equation 1) which has been validated for hip fracture cases [19]. The ROF is calculated from the maximum principal strain (ϵ) within the bone divided by elastic limit strain values. The criterion distinguishes between tensile and compressive loading states, and high ROF values in a localised region indicate a higher risk of fracture.

$$ROF = \begin{cases} \epsilon / 0.0073 & \text{if tensile} \\ \epsilon / 0.0104 & \text{if compressive} \end{cases} \quad (1)$$

An advantage of using FEA to examine risk factors for bone fracture is that the uncertainty resulting from confounding factors is removed, enabling the study to focus on the parameters of interest. The aim of this study was apply Schileo's fracture criterion and utilise probabilistic FEA methods to assess which surgical parameters increase the risk of periprosthetic fracture after unicompartmental knee replacement.

2 Methods

The study first quantified the surgical variability in the preparation of tibia for UKR, then used the Monte Carlo method to virtually implant 1000 UKRs, representing that variability. The risk of fracture for the finite element models was found and multivariate linear regression used to assess the influence of each surgical cut parameter.

2.1 *Quantification of variability in surgical cuts*

Twenty three right tibial Sawbones (custom anatomic design made for Zimmer-Biomet UK Ltd. by Sawbones, Pacific Research Laboratories Inc., Vashon Island, Washington, USA) were prepared for medial mobile UKR (Oxford Partial Knee, Biomet, Bridgend, UK) as part of an instructional course. The attendees were a mixture of experienced and inexperienced orthopaedic surgeons who each prepared a Sawbone tibia after receiving training in the operative technique. Measurements were then taken of the positions and depths of the surgical cuts (Figure 1). The parameters examined were:

- the resection depth (the superior-inferior distance from the tibial plateau on the lateral side to the resected medial horizontal cut, where the distance was parallel to the mechanical axis)
- the angle between the horizontal and vertical cuts
- the depth of the vertical cuts, both at the posterior and anterior cortex
- the depth of the horizontal cuts, both at the posterior and anterior cortex
- the depth of the pin hole (used to hold the cutting guide)

2.2 *Finite element model*

2.2.1 *Geometry*

The finite element model was based on a previously published UKR tibial model that was validated against cadaveric tests [20]. The tibial geometry was segmented from a CT scan of a cadaveric tibia obtained from a male donor aged 60 years with a body mass index of 22.5. The geometry was segmented using Mimics software (version 14.1, Materialise, Leuven, Belgium) and smoothed using the Scanto3D function in SolidWorks software (version 2012, Simulia, Waltham, MA, USA). The tibia was aligned so that the tibial mechanical axis was the Z-axis, anterior-posterior was the X-axis, and medial-lateral was the Y-axis. Previous work verified that use of a shortened tibia improves computational speed without affecting the strain in the periprosthetic region [21]. Therefore, the length of the tibia was shortened to 100 mm proximally.

The UKR was implanted virtually using Boolean functions within ABAQUS software (Version 6.12, Dassault-Systèmes, Rhode Island, USA). A Python script (version 2.6, Python Software Foundation) was created to automate the implantation for different surgical and loading parameters. The width of all saw cuts used was 1 mm, which is the width of the saw blade used during surgery [22]. The base of the Oxford Unicompartmental Knee tibial tray was fully fixed to the tibia, and frictionless contact was defined between the tray wall and the bone. Neither the effect of interference fit nor loosening was examined in this study.

2.2.2 *Mesh*

The finite element mesh was created using ABAQUS software. Quadratic tetrahedral elements (C3D10) were used to mesh the bone and the tibial tray was meshed with quadrilateral rigid elements (R3D4). A smaller element size (a third of the overall element

size) was assigned to; the muscle attachment sites, the edges created by the saw cuts, and the drilled pin-hole.

A mesh convergence study was performed to determine the optimal mesh density, where convergence was defined as when the output was within 5% of the next three finer element sizes (0.1 mm mesh size intervals). The model converged at an overall element size of 2.4 mm.

2.2.3 Material properties

The tray was modelled as a rigid cobalt chromium-molybdenum alloy with a density of 8.4 g cm⁻³ [23]. The tibia was modelled as a heterogeneous linear elastic material, where the modulus of each element was assigned based on the corresponding gray scale value of that element in the CT scan of the tibia. The bone material assignment was performed with Mimics software (400 material intervals with a modulus range of 1 to 22 GPa, consistent with previous work [20]).

2.2.4 Musculoskeletal model

The muscle and contact loads applied to the tibia throughout the gait cycle were estimated using data from an instrumented total knee replacement (TKR) implanted in a male subject (age: 83 years, BMI: 22.5, alignment: neutral) at the Shiley Center for Orthopaedic Research and Education at the Scripps Clinic in California [24]. The data were recorded while the patient performed overground walking trials at a self-selected speed [25; 26] and included the following quantities: contact forces on the tibial tray, ground reaction forces and moments, surface marker positions, and electromyographic (EMG) data. Medial and lateral tibial contact forces were calculated from the implant load cell data using an elastic foundation

contact model [27], and muscle force estimates were generated using static optimization of a subject-specific knee model [28] constructed in OpenSim [29]. The musculoskeletal knee model and muscle force estimation approach have been described in detail in a previous study [23].

2.2.5 Boundary conditions

The muscle and contact loads from the musculoskeletal model were applied to the FE model using distributed coupling to the tibial attachment site (Figure 2). On the lateral side the compartment loads were applied to the tibial articular surface in the same manner, while on the medial side the compartment load was applied to the upper surface of the tibial tray using an equation derived in a previous study to represent the pressure field [23]. The distal end of the tibia was fixed in all degrees of freedom.

The cadaveric tibia used for the finite element model in the present study was different from the instrumented knee subject tibia. Both tibias were from male subjects with a similar body mass index (instrumented tibia: 22.5 and cadaveric tibia: 25.9) and size (instrumented tibia: 75.0 mm tibial width, and cadaveric tibia: 76.5 mm tibial width) but different age (instrumented tibia: 83, cadaveric tibia: 60). An iterative closest point (ICP) algorithm was used to register the two tibias and determine the muscle attachment sites and vectors for the new geometry.

2.2.6 Post-processing

The risk of fracture parameter described by Schileo et al. [19] (Equation 1) is not automatically calculated by ABAQUS software, so a custom Python script was written to interact with ABAQUS and calculate the new field output. The two outputs used for the

analysis were: (1) the maximum ROF value (omitting artificially high results at muscle attachment sites), and (2) the total volume of elements exceeding an ROF of 1 (threshold for fracture defined by Schileo).

2.3 Application of the Monte Carlo method

The measurements taken from the tibia prepared during the surgical training course (Section 2.1, Figure 1) were used to define the envelope of surgical cut variation for the models. A thousand finite element models were then created to represent the variance in surgical technique.

The distribution of each surgical cut parameter was categorised from the measured data using the Kernel Density function from the ‘scikit-learn’ machine learning module implemented in Python [30]. A Gaussian kernel ($K(x;h)$) was applied with a bandwidth (h) of 0.75, to create the function representing the distribution of cut parameters measured from the sawbones. The kernel has the form given in Equation (2) where the density estimate at point y is found from the provided group of points $x_i; i = 1 \dots N$.

$$\rho_K(y) = \sum_{i=1}^N K\left(\frac{y - x_i}{h}\right) \quad (2)$$

An ABAQUS-python script was then used to automate the creation of each finite element model. The script involved the following steps:

1. Randomly select each surgical cut parameter from its calculated distribution, using Python ‘random’ and ‘scikit-learn’ packages.
2. Prepare tibia using Boolean operations

3. Assemble tibia and UKR components
4. Apply muscle loading, joint loading, constraints. and materials
5. Mesh and solve

To confirm that 1000 models were sufficient to achieve convergence of the Monte Carlo method, we used the method described by Fishman *et al.* [31]. Convergence was defined when the mean and coefficient of variance were within 3% of their values from the last 10% of valid instantiations [31; 32].

2.4 *Model verification*

The finite element model was verified two ways: (1) the location of elements at risk of fracture were compared to typical clinical fracture locations [1] and (2) the maximum ROF in the periprosthetic region was compared with failure loads reported by Clarius *et al.* [18]. To replicate the experiments performed by Clarius an increasing load (max 10 kN) was applied to the medial compartment while the two risk of fracture criteria were recorded. The tibia was analysed with, and without an extended vertical cut (cut angled at 10 degrees [18]). No muscle or lateral compartment loading was applied.

2.5 *Statistical analysis*

Which parameters influenced the risk of fracture was determined by performing an analysis of variance (ANOVA) test. The parameters which significantly ($p < 0.05$) influenced the risk of fracture were then input into a generalised linear regression (GLM) model. All statistical analyses were implemented in R (www.r-project.org). To ensure the dependent variables (maximum ROF and Volume of failed elements) were normally distributed for the ANOVA

and GLM model, we transformed the data by taking the logarithm of the maximum ROF and the cube root of the volume of failed elements.

3 Results

3.1 Quantification of variability in surgical cuts

The measurements of the prepared tibial Sawbones highlighted large variability in the vertical and horizontal cuts posteriorly (Table 1). The standard deviation in the anterior cut depths was half that of the posterior cuts. Furthermore, in 14 of the 23 Sawbone tibias, the pin hole had gone into the keel slot, greatly increasing the hole depth and producing a bi-modal distribution with a high standard deviation. The cut angle had very low variability (percent deviation 1.6%) and so was not included in the Monte Carlo models.

3.2 Application of the Monte Carlo method

A linear relationship was found between medial contact force and the maximum ROF when loaded through the whole gait cycle ($R^2 = 0.83$), despite the varying muscle loads and load vectors from the musculoskeletal model (Figure 3). The maximum ROF value and the maximum volume of failed elements occurred at 16% of the gait cycle, so these results were used for the regression analysis.

The ANOVA results (Table 2) found the extension of the vertical cut posteriorly (e), the resection depth (a), and extension of the horizontal cut posteriorly (f) to significantly influence both the maximum ROF value and the volume of failed elements. Consequently, these parameters were used to create the regression model. The correlation between both output variables and the posterior vertical and horizontal cuts and the resection depth was also confirmed visually (Figure 4).

The multivariate linear regression model found that the greater the resection depth and the more extended the posterior vertical cut, the greater the risk of fracture in terms of both the maximum ROF and the volume of failed elements. In contrast, extension of the horizontal cut posteriorly reduced the risk of fracture slightly. The parameters which most influenced the risk of fracture were the resection depth and extension of the vertical cut posteriorly, as can be seen from the 3-dimensional scatterplot shown in Figure 5.

From the known resection depth, posterior vertical cut, and posterior horizontal cut for each of the 1000 models, the regression equations were used to calculate the maximum ROF value (Equation (3)) and the volume of failed elements (Equation (4)). The equations were able to predict the finite element maximum ROF with a Pearson's correlation coefficient of 0.59 and the volume of failed elements with a 90% correlation, indicating a reasonable regression model fit.

$$Max\ ROF = 10^{(0.0152e + 0.0161a - 0.0052f + 0.102)} \quad (3)$$

$$Volume\ of\ failed\ elements = (0.0454e + 0.061a - 0.029f + 0.267)^3 \quad (4)$$

Where: (*a*) is the resection depth, (*e*) is the extension of the vertical cut posteriorly, and (*f*) is the extension of the horizontal cut posteriorly.

3.3 Model Verification

When the tibia was prepared and loaded in the same manner as described by Clarius *et al.* [18], regions of high ROF were observed in the corner between the horizontal and the vertical cut and in the region surrounding the keel. At high loads, these two regions combined to create a line of high fracture risk extending to the tibial cortex (Figure 6). The line matched

the path of fractures observed clinically [1], confirming that the ROF parameter is an indicator of tibial fracture risk.

The average failure load reported by Clarius *et al.* for a tibia with an excessive vertical cut was 2.6 kN (range 1.08-5.04), and 3.9 kN (range 2.35-8.50) for a the tibia with a perfect cut [18]. The finite element models when loaded under these conditions had corresponding maximum ROF values of 5.2 and 5.6, and volume of failed elements of 128 mm³ and 177 mm³, respectively. These results indicate that a maximum ROF value above 5, with a failure volume greater than 128 mm³, would represent a high fracture risk. From the 1000 models examined in the Monte Carlo simulation, 0.3% had a maximum ROF greater than 5; none reached the volume threshold.

4 Discussion

This study used a probabilistic finite element modelling approach to investigate the influence of different surgical cuts used to prepare the tibia for unicompartmental knee replacement on the risk of periprosthetic fracture. Of the surgical parameters investigated, excessive resection depth and an extended posterior vertical saw cut were found to significantly increase the risk of fracture according to the regression model. Furthermore, based on measurements of the Sawbone tibias prepared by surgeons as part of an instructional course, the depths of the vertical saw cuts posteriorly are highly variable. This combination of results is concerning, as high variability in a factor believed to increase the risk of fracture increases uncertainty in the surgical outcome.

The tibial saw guide is an important part of the surgical instrumentation for making the vertical saw cut. The guide comprises a rectangular block, which is pinned to the anterior

side of the tibia (causing the pin holes described) and provides a horizontal surface to stop the saw blade when making the vertical cut. Although the guide provides a stop anteriorly, there is no such stop posteriorly, and the surgeon is required to estimate the correct cut angle (7 degrees). The guide is also used to aid the horizontal cut, where the flat side of the reciprocating saw rests on the top of the block which ensures the cut is straight and has a 7 degree posterior slope [22]. If the surgeon under-estimates the slope of the vertical cut, the horizontal and vertical cuts will not meet and the vertical cut will need to be extended to enable the worn tibial plateau to be removed. If the surgeon over-estimates the slope, the vertical cut will be excessive, causing a posterior notch. It is, therefore, difficult for a surgeon to ensure that the vertical cut is not excessive with the current operative technique, and limited posterior visibility makes it hard to identify cut depth intra-operatively.

The resection depth is controlled by the height at which the tibial guide is pinned. The operative technique suggests the level should be 2 to 3 mm lower than the eroded bone [22]. Several studies have suggested that errors in the vertical cut increase the risk of fracture [1; 9; 17], and Clarius *et al.* demonstrated this relationship experimentally [18]. However, resection depth has been proposed by only one other publication as a critical parameter and is largely overlooked in the literature [1]. If clinicians were made aware that excessive resection can contribute to fracture, it would be simple for them to modify their surgical practice accordingly.

Regardless of the manufacturer or implant type, all UKR designs require an L-shaped space to be created for the tibial component, which requires a horizontal resection cut and a vertical cut to be made by the surgeon. This consistency in UKR surgical technique may explain why tibial plateau fracture is not restricted to only one device design [3; 5; 6; 8]. By knowing the

309 surgical factors which may increase tibial fracture risk, surgeons and orthopaedic
310 manufacturers can begin to propose solutions that can minimise the risk of fracture after
311 UKR.

312 Interestingly, the finite element model which simulated loading throughout a whole gait cycle
313 found a linear relationship between the risk of fracture and the medial load. Rudol *et al.*
314 suggested that peri-prosthetic fracture after UKR may be linked to patient weight [9], and our
315 results indicate it could be a risk factor. Whether high body mass index should be considered
316 a contraindication for UKR is a controversial topic, with evidence both for [33] and against
317 [34; 35]. Some case studies in the literature have mentioned limiting weight-bearing and
318 using medial unloading braces to offset the medial load in cases of peri-prosthetic fracture to
319 aid healing [3; 10], but not as a preventative measure. In patients considered at risk, bracing
320 could be used as a non-invasive treatment.

321 Periprosthetic tibial fractures after UKR can occur intraoperatively or post-operatively [1; 9].
322 Reports of intraoperative fracture describe a high strain-rate impact load which causes the
323 bone to fracture [3; 5]. Post-operative fractures are associated with a combination of intra-
324 operative damage and cumulative damage from cyclic loading of the bone [36]. Studies of
325 patient activities after knee replacement have shown that in a typical day a patient will stand
326 for 21% of the time, walk for 8%, and climb stairs for 1%; the remaining time is non-weight
327 bearing [37]. In terms of cyclic loading, gait is therefore the most likely activity to cause
328 post-operative peri-prosthetic fracture, though the largest medial contact forces occur for stair
329 ascent and descent [38]. Our finite model did not examine the development of cumulative
330 strains within the bone, but both static [19] and fatigue mechanisms of bone fracture [36]
331 have been related to strain.

It is important to consider the limitations of this work. The load data which were applied to the model were based on results from an instrumented total knee replacement, rather than from a unicompartmental knee replacement. As UKR forces have never been measured directly *in vivo*, it is not possible to know whether the load distribution between the condyles is equivalent. However, an anatomic approach was used to implant the instrumented TKR [38], and therefore alignment should have been similar to an implanted UKR with a similar load distribution between the condyles. In addition, our model assumed perfect fixation of the base of the tibial tray to the bone and so could not consider component loosening or interference fit. Incorporating loosening and interference fit adds significant complexity to the model and is planned for inclusion in future work.

In conclusion, the results of this study have highlighted the importance of careful surgical preparation of the tibial plateau prior to UKR implantation. This study suggests that the cause of fracture is multifactorial and that to minimise the risk of fracture, a surgeon should;

- ensure that the vertical cut does not go too deep posteriorly
- be conservative with resection of the tibia

It may be possible to reduce the likelihood of an excessively deep vertical cut by altering the surgical technique. If the horizontal cut were made before the vertical cut, a shim could be inserted into the horizontal saw cut to stop the vertical cut from going too deep. Surgeon training and better communication of the fracture risks could encourage surgeons to be more conservative when resecting the tibia. If orthopaedic manufacturers and surgeons worked to implement these changes in operative technique, our results suggest that the risk of tibial plateau fracture after UKR could be reduced.

Acknowledgements

The work was funded in part by NIH grant R01EB009351. Some of the authors have received funding from Biomet UK Healthcare Ltd. (the manufacturer of the implant examined in this study), but the funding was unrelated to the present study. Dr Pegg's salary was funded by the Oxford Orthopaedic Engineering Centre. We would like to thank the surgeons who attended the instructional course, and Kyung Tae Kim, M.D., Ph.D. for providing data regarding cases of tibial fracture after UKR in Seoul.

References

1. Pandit, H., Murray, D.W., Dodd, C.A., et al., 2007. Medial tibial plateau fracture and the Oxford unicompartmental knee. *Orthopedics* 30: 28-31.
2. Della Rocca, G.J., Leung, K.S., Pape, H.-C., 2011. Periprosthetic Fractures: Epidemiology and Future Projections. *Journal of Orthopaedic Trauma* 25: S66-S70. DOI: 10.1097/BOT.0b013e31821b8c28.
3. Van Loon, P., de Munnynck, B., Bellemans, J., 2006. Periprosthetic fracture of the tibial plateau after unicompartmental knee arthroplasty. *Acta Orthop Belg* 72: 369-374.
4. Lindstrand, A., Stenström, A., Ryd, L., et al., 2000. The introduction period of unicompartmental knee arthroplasty is critical. *The Journal of Arthroplasty* 15: 608-616. DOI: 10.1054/arth.2000.6619.
5. Sloper, P.J.H., Hing, C.B., Donell, S.T., et al., 2003. Intra-operative tibial plateau fracture during unicompartmental knee replacement: a case report. *The Knee* 10: 367-369. DOI: 10.1016/S0968-0160(03)00003-6.
6. Yang, K.-Y., Yeo, S.-J., Lo, N.-N., 2003. Stress fracture of the medial tibial plateau after minimally invasive unicompartmental knee arthroplasty: A Report of 2 Cases. *The Journal of Arthroplasty* 18: 801-803. DOI: 10.1016/S0883-5403(03)00332-2.
7. Kumar, A., Chambers, I., Wong, P., 2008. Periprosthetic Fracture of the Proximal Tibia After Lateral Unicompartmental Knee Arthroplasty. *The Journal of Arthroplasty* 23: 615-618. DOI: 10.1016/j.arth.2007.04.036.

381 8. Kumar, A., Fiddian, N.J., 1997. Fracture of the medial tibial plateau following
382 unicompartmental knee replacement. *The Knee* 4: 177-178. DOI: 10.1016/S0968-
383 0160(97)00257-3.

384 9. Rudol, G., Jackson, M.P., James, S.E., 2007. Medial Tibial Plateau Fracture Complicating
385 Unicompartmental Knee Arthroplasty. *The Journal of Arthroplasty* 22: 148-150. DOI:
386 10.1016/j.arth.2006.01.005.

387 10. Brumby, S.A., Carrington, R., Zayontz, S., et al., 2003. Tibial plateau stress fracture: A
388 complication of unicompartmental knee arthroplasty using 4 guide pinholes. *The Journal of*
389 *Arthroplasty* 18: 809-812. DOI: 10.1016/S0883-5403(03)00330-9.

390 11. Berger, R.A., Meneghini, R.M., Jacobs, J.J., et al., 2005. Results of Unicompartmental
391 Knee Arthroplasty at a Minimum of Ten Years of Follow-up. *JBJS* 87: 999-1006. DOI:
392 10.2106/JBJS.C.00568.

393 12. NJR, 2014. National Joint Registry 11th Annual Report.

394 13. Bennett, J.E., Li, G., Foreman, K., et al. 2015. The future of life expectancy and life
395 expectancy inequalities in England and Wales: Bayesian spatiotemporal forecasting. *The*
396 *Lancet* 386: 11-17. DOI: 10.1016/S0140-6736(15)60296-3

397 14. Gauthier, A., Kanis, J., Jiang, Y., et al., 2011. Epidemiological burden of postmenopausal
398 osteoporosis in the UK from 2010 to 2021: estimations from a disease model. *Archives of*
399 *Osteoporosis* 6: 179-188. DOI: 10.1007/s11657-011-0063-y.

400 15. Naudie, D.D., Ammeen, D.J., Engh, G.A., et al., 2007. Wear and osteolysis around total
401 knee arthroplasty. *J Am Acad Orthop Surg* 15: 53-64.

402 16. Seeger, J.B., Haas, D., Jager, S., et al., 2012. Extended sagittal saw cut significantly
403 reduces fracture load in cementless unicompartmental knee arthroplasty compared to
404 cemented tibia plateaus: an experimental cadaver study. *Knee Surg Sports Traumatol*
405 *Arthrosc* 20: 1087-1091. DOI: 10.1007/s00167-011-1698-3.

406 17. Clarius, M., Aldinger, P.R., Bruckner, T., et al., 2009. Saw cuts in unicompartmental knee
407 arthroplasty: An analysis of sawbone preparations. *The Knee* 16: 314-316. DOI:
408 10.1016/j.knee.2008.12.018.

- 409 18. Clarius, M., Haas, D., Aldinger, P.R., et al., 2010. Periprosthetic tibial fractures in
410 unicompartmental knee arthroplasty as a function of extended sagittal saw cuts: An
411 experimental study. *The Knee* 17: 57-60. DOI: 10.1016/j.knee.2009.05.004.
- 412 19. Schileo, E., Taddei, F., Cristofolini, L., et al., 2008. Subject-specific finite element
413 models implementing a maximum principal strain criterion are able to estimate failure risk
414 and fracture location on human femurs tested in vitro. *Journal of Biomechanics* 41: 356-367.
415 DOI: 10.1016/j.jbiomech.2007.09.009.
- 416 20. Gray, H.A., Taddei, F., Zavatsky, A.B., et al., 2008. Experimental validation of a finite
417 element model of a human cadaveric tibia. *J Biomech Eng* 130: 031016. DOI:
418 10.1115/1.2913335.
- 419 21. Simpson, D.J., Price, A.J., Gulati, A., et al., 2009. Elevated proximal tibial strains
420 following unicompartmental knee replacement—A possible cause of pain. *Medical*
421 *Engineering & Physics* 31: 752-757. DOI: 10.1016/j.medengphy.2009.02.004.
- 422 22. Biomet, 2011. Oxford® Cementless Partial Knee Supplementary Surgical Technique. In.
423 Biomet UK Ltd., Bridgend, UK.
- 424 23. Pegg, E.C., Walter, J., Mellon, S.J., et al., 2013. Evaluation of factors affecting tibial bone
425 strain after unicompartmental knee replacement. *Journal of Orthopaedic Research* 31: 821-
426 828. DOI: 10.1002/jor.22283.
- 427 24. D'Lima, D.D., Townsend, C.P., Arms, S.W., et al., 2005. An implantable telemetry
428 device to measure intra-articular tibial forces. *Journal of Biomechanics* 38: 299-304. DOI:
429 10.1016/j.jbiomech.2004.02.011.
- 430 25. Fregly, B.J., Besier, T.F., Lloyd, D.G., et al., 2012. Grand challenge competition to
431 predict in vivo knee loads. *Journal of Orthopaedic Research* 30: 503-513. DOI:
432 10.1002/jor.22023.
- 433 26. Fregly, B.J., D'Lima, D.D., Colwell, C.W., 2009. Effective gait patterns for offloading the
434 medial compartment of the knee. *Journal of Orthopaedic Research* 27: 1016-1021. DOI:
435 10.1002/jor.20843.
- 436 27. Bei, Y., Fregly, B.J., 2004. Multibody dynamic simulation of knee contact mechanics.
437 *Medical Engineering & Physics* 26: 777-789. DOI: 10.1016/j.medengphy.2004.07.004.

438 28. Lin, Y.-C., Walter, J.P., Banks, S.A., et al., 2010. Simultaneous prediction of muscle and
439 contact forces in the knee during gait. *Journal of Biomechanics* 43: 945-952. DOI:
440 10.1016/j.jbiomech.2009.10.048.

441 29. Delp, S.L., Anderson, F.C., Arnold, A.S., et al., 2007. OpenSim: Open-Source Software
442 to Create and Analyze Dynamic Simulations of Movement. *Biomedical Engineering, IEEE*
443 *Transactions on* 54: 1940-1950. DOI: 10.1109/TBME.2007.901024.

444 30. Pedregosa, F., Varoquaux, G., Gramfort, A., et al., 2011. Scikit-learn: Machine Learning
445 in Python. *Journal of Machine Learning Research* 12: 2825-2830.

446 31. Fishman, G.S., 1996. Monte Carlo. Springer-Verlag New York.

447 32. Reinbolt, J.A., Haftka, R.T., Chmielewski, T.L., et al., 2007. Are patient-specific joint
448 and inertial parameters necessary for accurate inverse dynamics analyses of gait? *IEEE Trans*
449 *Biomed Eng* 54: 782-793. DOI: 10.1109/TBME.2006.889187.

450 33. Berend, K.R., Lombardi, A.V.J., Mallory, T.H., et al., 2005. Early Failure of Minimally
451 Invasive Unicompartmental Knee Arthroplasty Is Associated with Obesity. *Clinical*
452 *Orthopaedics and Related Research* 440: 60-66. DOI: 10.1097/01.blo.0000187062.65691.e3.

453 34. Murray, D.W., Pandit, H., Weston-Simons, J.S., et al., 2013. Does body mass index affect
454 the outcome of unicompartmental knee replacement? *The Knee* 20: 461-465. DOI:
455 10.1016/j.knee.2012.09.017.


456 35. Naal, F., Neuerburg, C., Salzmann, G., et al., 2009. Association of body mass index and
457 clinical outcome 2 years after unicompartmental knee arthroplasty. *Archives of Orthopaedic*
458 *and Trauma Surgery* 129: 463-468. DOI: 10.1007/s00402-008-0633-7.

459 36. Carter, D.R., Caler, W.E., 1985. A cumulative damage model for bone fracture. *Journal*
460 *of Orthopaedic Research* 3: 84-90. DOI: 10.1002/jor.1100030110.

461 37. Huddleston, J., Scarborough, D., Goldvasser, D., et al., 2009. 2009 Marshall Urist Young
462 Investigator Award: How Often Do Patients with High-Flex Total Knee Arthroplasty Use
463 High Flexion? *Clinical Orthopaedics and Related Research®* 467: 1898-1906. DOI:
464 10.1007/s11999-009-0874-y.

465 38. Mündermann, A., Dyrby, C.O., D'Lima, D.D., et al., 2008. In vivo knee loading
466 characteristics during activities of daily living as measured by an instrumented total knee
467 replacement. *Journal of Orthopaedic Research* 26: 1167-1172. DOI: 10.1002/jor.20655.

468 **Tables**


Parameter	Mean	Standard Deviation	0%	25%	50%	75%	100%
							
a (resection depth, mm)	8.8	1.7	5.0	8.0	9.0	10.0	11.0
b (angle between cuts, deg)	90.6	1.4	88.0	90.0	90.0	91.0	95.0
c (vertical cut anterior, mm)	0.5	1.0	0.0	0.0	0.0	0.3	4.0
d (horizontal cut anterior, mm)	0.7	0.9	0.0	0.0	0.0	1.0	3.0
e (vertical cut posterior, mm)	4.2	3.9	0.0	0.0	4.0	7.0	12.0
f (horizontal cut posterior, mm)	1.3	2.1	0.0	0.0	1.0	1.3	7.5
g (pin depth, mm)	28.6	6.8	8.0	25.0	30.0	33.5	36.0

469 Table 1. The surgical cut parameters measured from 23 synthetic tibias prepared by surgeons
 470 during an instructional course. The mean value, standard deviation and distribution
 471 percentiles for each parameter are summarised.

Parameter	Maximum ROF value			Volume of failed elements		
	F	p	Sig	F	p	Sig
a (resection depth, mm)	183.4	2.2e-16	***	1295.5	2.2e-16	***
c (vertical cut anterior, mm)	0.1	0.028	*	0.04	0.843	NS
d (horizontal cut anterior, mm)	21.2	0.709	NS	21.3	4.3e-06	***
e (vertical cut posterior, mm)	4.9	2.2e-16	***	2859.7	2.2e-16	***
f (horizontal cut posterior, mm)	628.8	4.8e-06	***	315.4	2.2e-16	***
g (pin depth, mm)	0.8	0.365	NS	0.3	0.565	NS

Table 3. ANOVA test of the null hypotheses that the surgical cut parameters do not influence the maximum ROF value, and the volume of failed elements. The ANOVA F-value (F), p-value (p) and significance (Sig) results (*= $p < 0.05$, **= $p < 0.01$, ***= $p < 0.001$, NS= $p > 0.5$) are shown.


477 **Figure Legends**



478


479 Figure 1. The surgical cut parameters measured from synthetic sawbone tibia were: the
480 resection depth (a), the angle between the horizontal and vertical cuts (b), the extension of the
481 vertical and horizontal cuts posteriorly (e, f) and anteriorly (c, d), and the depth of the pin
482 hole required to hold the cutting guide (g).

483




484

485 Figure 2. The constraint (blue), load locations, and vectors (red) applied to the model at 16%
 486 of the gait cycle. The medial view shown includes the gracilis (Grac), sartorius (Sart),
 487 semiteninosus (Semiten), semimembranosus (Semimem), vastus medialis, vastus intermedius
 488 and vastus lateralis (Vastus) muscles forces; the tensor fasciae latae muscle forces were also
 489 applied on the lateral side.



490 (a)



491 (b)

492 Figure 3. The risk of fracture (ROF) varied through the gait cycle (a) and a linear correlation
 493 was observed with medial load (b).





Figure 4. Scatterplots of regression parameters which were found to significantly influence the maxim ROF value (log) and volume of failed elements (cube root). The lines in each plot represent the mean and the interquartile range.



500

501 Figure 5. Scatterplot illustrating the dependence of the volume of failed elements on the
 502 resection depth and the vertical cut. The multivariate regression model fit is represented by
 503 the plane and the red lines indicate the residuals.

504

505




Figure 6. Distribution of the risk of fracture through a perfectly cut tibia loaded at 3.9 kN (a), and a tibia with excessive vertical cut loaded at 2.6 kN (b). Both models represent conditions which caused tibial fracture in experiments performed by Clarius *et al.*. The region most at risk of fracture extends diagonally from the vertical cut to the tibial cortex, via the keel.

Conflict of interest statement

Some of the authors have received funding from Biomet UK Healthcare Ltd. (the manufacturer of the implant examined in this study), but the funding was unrelated to the present study.

Author contributions statement

Elise Pegg: Methodology, Software, Formal Analysis, Visualisation, Writing - Original Draft. **Jonathan Walter:** Methodology, Software, Formal Analysis, Writing - Review & Editing. **Darryl D'Lima:** Resources. **Benjamin Fregly:** Supervision, Resources, Writing - Review & Editing. **Harinderjit Gill:** Methodology, Writing - Review & Editing. **David Murray:** Ideas, Supervision, Writing - Review & Editing

CRediT author definitions

Term	Definition
Ideas	formulation or evolution of overarching research goals and aims
Methodology	Programming, software development; designing computer programs; implementation of the computer code and supporting algorithms; testing of existing code components
Software	Programming, software development; designing computer programs; implementation of the computer code and supporting algorithms; testing of existing code components
Validation	Verification whether as a part of the activity or separate, of the overall replication/ reproducibility of results/experiments and other research outputs
Formal Analysis	Application of statistical, mathematical, computational, or other formal techniques to analyze or synthesize study data
Investigation	Conducting a research and investigation process, specifically performing the experiments, or data/evidence collection
Resources	Provision of study materials, reagents, materials, patients, laboratory samples, animals, instrumentation, computing resources, or other analysis tools
Data Curation	Management activities to annotate (produce metadata), scrub data and maintain research data (including software code, where it is necessary for interpreting the data itself) for initial use and later reuse
Writing – Original Draft	Preparation, creation and/or presentation of the published work, specifically writing the initial draft (including substantive translation)
Writing – Review & Editing	Preparation, creation and/or presentation of the published work by those from the original research group, specifically critical review, commentary or revision - including pre-or postpublication stages
Visualization	Preparation, creation and/or presentation of the published work, specifically visualization/ data presentation
Supervision	Oversight and leadership responsibility for the research activity planning and execution, including mentorship external to the core team
Project Administration	Management and coordination responsibility for the research activity planning and execution
Funding Acquisition	Acquisition of the financial support for the project leading to this publication