








GUIDELINES**OSIPI Special Section****ASL lexicon and reporting recommendations: A consensus report from the ISMRM Open Science Initiative for Perfusion Imaging (OSIPI)**

Yuriko Suzuki¹  | **Patricia Clement**^{2,3} | **Weiyang Dai**⁴ | **Sudipto Dolui**⁵ | **Maria A. Fernández-Seara**^{6,7} | **Thomas Lindner**⁸  | **Henk J. M. M. Mutsaerts**⁹  | **Jan Petr**¹⁰  | **Xingfeng Shao**¹¹  | **Manuel Taso**¹²  | **David L. Thomas**¹³  | **ISMRM Perfusion Study Group**

Correspondence

Yuriko Suzuki, Wellcome Centre for Integrative Neuroimaging, FMRIB, Nuffield Department of Clinical Neurosciences, University of Oxford, Headley Way, Headington, Oxford, OX3 9DU, UK.
Email: yuriko.suzuki@ndcn.ox.ac.uk

Funding information

Deutsche Forschungsgemeinschaft, Grant/Award Number: LI-3030/2-1; EU Joint Program for Neurodegenerative Disease Research; European Union Horizon 2020 research and innovation programme, Grant/Award Number: ASPIRE E!113701; Eurostars-2 joint programme; Hartstichting, Grant/Award Number: 03-004-2020-T049; Netherlands Enterprise Agency (RvO); Netherlands Organisation for health Research and Development and Alzheimer Nederland, Grant/Award Number: DEBBIE JPND2020-568-106; National Institutes of Health, Grant/Award Numbers: R01 NS111115, R03 AG063213; Royal Academy of Engineering, Grant/Award Number: RF/201920/19/236; Spanish Ministry of Science, Innovation and Universities, Grant/Award Numbers: PI18/00084, PI21/00578; UCLH NIHR Biomedical Research Centre; Wellcome Trust, Grant/Award Numbers: 203139/Z/16/Z, Centre award 539208

Abstract

The 2015 consensus statement published by the International Society for Magnetic Resonance in Medicine (ISMRM) Perfusion Study Group and the European Cooperation in Science and Technology (COST) Action ASL in Dementia aimed to encourage the implementation of robust arterial spin labeling (ASL) perfusion MRI for clinical applications and promote consistency across scanner types, sites, and studies. Subsequently, the recommended 3D pseudo-continuous ASL sequence has been implemented by most major MRI manufacturers. However, ASL remains a rapidly and widely developing field, leading inevitably to further divergence of the technique and its associated terminology, which could cause confusion and hamper research reproducibility. On behalf of the ISMRM Perfusion Study Group, and as part of the ISMRM Open Science Initiative for Perfusion Imaging (OSIPI), the ASL Lexicon Task Force has been working on the development of an ASL Lexicon and Reporting Recommendations for perfusion imaging and analysis, aiming to (1) develop standardized, consensus nomenclature and terminology for the broad range of ASL imaging techniques and parameters, as well as for the physiological constants required for quantitative analysis; and (2) provide a community-endorsed recommendation of the imaging parameters that we encourage authors to include when describing ASL methods in scientific reports/papers. In this paper, the sequences and parameters in (pseudo-)continuous ASL, pulsed ASL, velocity-selective ASL, and multi-timepoint ASL for brain perfusion imaging are included. However, the content of the lexicon is not intended to be limited to these techniques, and this paper provides the foundation for a growing online inventory that will be extended by the community as further methods and improvements are developed and established.

KEYWORDS

arterial spin labeling, interoperability, noninvasive, perfusion imaging, reproducibility

For affiliations refer to page 1758

This is an open access article under the terms of the [Creative Commons Attribution](https://creativecommons.org/licenses/by/4.0/) License, which permits use, distribution and reproduction in any medium, provided the original work is properly cited.

© 2023 The Authors. *Magnetic Resonance in Medicine* published by Wiley Periodicals LLC on behalf of International Society for Magnetic Resonance in Medicine.

1 | INTRODUCTION

Following the consensus statement for the recommended implementation of arterial spin labeling (ASL) perfusion MRI for clinical application in the brain¹ by the Perfusion Study Group (SG) of the International Society for Magnetic Resonance in Medicine (ISMRM) and the European Consortium for ASL in Dementia (European Cooperation in Science and Technology (COST) Action BM1103) in 2014 (referred to hereafter as the *ASL White Paper*), standardized ASL perfusion imaging sequences have now been implemented by the majority of MRI manufacturers. Recommended acquisition protocols and the increased availability of ASL imaging sequences have encouraged the use of ASL in clinical applications.² However, ASL remains a rapidly and widely developing field, both in terms of improving the accuracy and precision of cerebral blood flow (CBF) quantification via advances in pulse sequence and post-processing methods, and providing other output derivatives in addition to CBF (e.g., arterial transit time). These advances have greatly expanded the scope of ASL but also bring further divergence of the technique, particularly in the terminology used, which can lead to confusion and hamper interoperability. In addition, motivated by the noninvasive nature of ASL, there is an increased number of large cohort studies that adopt ASL perfusion imaging, such as the Alzheimer's Disease Neuroimaging Initiative (<http://adni.loni.usc.edu/adni-3/>) and some branches of the Human Connectome Project,³ in which data are acquired from multiple sites using different MRI scanners. To maximize the usefulness of these data, guidelines for consistent reporting of image acquisition parameters are essential.

As part of the ISMRM Open Science Initiative for Perfusion Imaging (ISMRM OSIPi, referred to hereafter as *OSIPi*), an initiative and activity of the ISMRM Perfusion SG, the ASL Lexicon Task Force has been working on the development of an ASL Lexicon and Reporting Recommendation for Perfusion Imaging and Analysis. The purpose of the ASL lexicon is to develop standardized nomenclature and terminology for the broad range of ASL imaging techniques and parameters, as well as for the physiological constants required for quantitative analysis. However, this ASL lexicon does not provide recommended standard ASL implementations and optimal parameter values, which is the remit of the other parallel recommendations/guidelines, and the readers are directed to those, such as the ASL White Paper, for a standard ASL implementation and processing approach, and its recent extensions for up-to-date summaries of more specific ASL techniques and developments (see the following subsection “1.1 Previous efforts on ASL standardization relevant to the development of ASL Lexicon and Reporting

Recommendation”). Instead, ASL lexicon aims to provide harmonization across documentation, reports, and publications by standardizing the terminology and parameter definitions, which is beyond the scope of the ASL White Paper and the others. In addition, the developed ASL lexicon is intended to form a common, community-endorsed recommendation for reporting of ASL perfusion imaging, providing a list describing which parameters in acquisition protocols should be reported by investigators and how, aiming to improve the interoperability and comparability of reported studies.

In summary, this paper is primarily intended to provide:

- A lexicon for researchers/developers of ASL sequences and analysis tools to conform to the community-based consensus recommendation in order to avoid misunderstandings caused by diverse terminologies and inconsistent definitions.
- A reporting guideline for researchers using ASL sequences and analysis tools to find how their ASL studies and results should be documented and reported, which should make their studies more widely understandable and reproducible.

Within the OSIPi framework, an overarching aim of this paper is to enable researchers and developers to use openly available datasets (e.g., data repositories) without ambiguity relating to the acquisition parameters used.

1.1 | Previous efforts on ASL standardization relevant to the development of ASL Lexicon and Reporting Recommendation

- A consensus statement of recommended implementations of ASL perfusion imaging for clinical applications (ASL White Paper¹), published by an expert group of members of the ISMRM Perfusion SG and the European Consortium for ASL in Dementia (COST Action BM1103).
- Brain Image Data Structure (BIDS) extension for ASL⁴ (referred to as *ASL-BIDS* in this paper): A community effort to standardize how to organize and share neuroimaging datasets, which was extended to include ASL in 2021 (<https://bids-specification.readthedocs.io/en/stable/04-modality-specific-files/01-magnetic-resonance-imaging-data.html#arterial-spin-labeling-perfusion-data>).
- Technical recommendations for renal ASL⁵ from an international group of experts working under the

framework of “Magnetic Resonance Imaging Biomarkers for Chronic Kidney Diseases (PARENCHIMA)”, funded by the EU COST Action CA16103 (referred to here as *PARENCHIMA renal ASL*).

- A series of extensions to the ASL White Paper (referred to as the *ASL Gray Papers* in this paper) on the following topics:
 - “Velocity-Selective Arterial Spin Labeling Perfusion MRI: A Review of the State of the Art and Recommendations for Clinical Implementation.”⁶
 - “Recent Technical Developments in ASL: A Review of the State of the Art.”⁷
 - “Current State and Guidance on Arterial Spin Labeling Perfusion MRI in Clinical Neuroimaging.”²
 - “Update on State-of-the-Art for Arterial Spin Labeling (ASL) Human Perfusion Imaging Outside of the Brain.”⁸
 - Quantitative Cerebral Perfusion MRI Using Multi-Timepoint Arterial Spin Labeling: Recommendations and Clinical Applications (preprint⁹).

1.2 | Development process of ASL Lexicon and reporting recommendation

The ASL Lexicon task force consists of 11 Perfusion SG members with diverse expertise in ASL imaging, who attended the launch events of OSUPI during and after the annual ISMRM meeting in 2019 and expressed their interest to contribute. The developmental process of the ASL Lexicon and Reporting Recommendation was as follows:

Stage 1 (June 2020–May 2021): Previously published ASL papers were reviewed by the task force members, and comprehensive lists of ASL techniques, acquisition parameters, output derivatives, and physiological parameters required for quantification were compiled. Terminology was harmonized with other community efforts, as mentioned above. The Reporting Recommendation was also drafted based on the BIDS extension for ASL,⁴ consisting of two recommendation levels:

- **Required: essential for meaningful interpretation of the ASL data and for quantitative analysis.** These must be included in an ASL publication in order for its data set to be ‘OSUPI-compliant’.
- **Recommended: parameters that are useful for interpretation of the ASL data and could explain specific characteristics or systematic differences between data sets.** Authors are encouraged to include as many of these as possible in ASL publications.

Stage 2 (June 2021–July 2021): A separate and independent expert panel provided feedback and comments on the stage 1 draft. These experts were involved with the development of the ASL Gray Papers (please see the previous subsection “1.1 Previous efforts on ASL standardization relevant to the development of ASL Lexicon and Reporting Recommendation”). Based on their feedback, an updated stage 2 draft was generated.

Stage 3 (June 2021–October 2021): A manufacturer survey was carried out with the major MRI scanner manufacturers (in alphabetical order: Canon Medical Systems Corporation, Tochigi, Japan, FUJIFILM Healthcare, Tokyo, Japan, GE Healthcare, Waukesha, WI, Philips Healthcare, Best, The Netherlands, and Siemens Healthcare, Erlangen, Germany) to identify any potential conflicts or incompatible terminologies and definitions with their current ASL product implementation. In addition, information was requested relating to if/how the acquisition parameters listed in the reporting recommendation can be obtained via the graphical user interface (GUI) of the commercial MRI scanners.

Stage 4 (November 2021–January 2022): The draft document was shared with all members of the ISMRM Perfusion SG for general feedback and comments. Also, a survey was enclosed so that they could indicate if they agreed with the drafted Reporting Recommendation categories with regard to the ASL acquisition parameters. In the survey, the responders were provided with four options for each ASL acquisition parameter listed in the recommendation:

- Yes, I think *Required/Recommended* is the appropriate category for parameter *xxx*.
- No, the parameter *xxx* should be in another category (i.e., *Recommended* for *Required/Required* for *Recommended*).
- No, we should remove the parameter *xxx* from the recommendation.
- I am not familiar with this parameter.

On November 19, 2021, a virtual Q&A session was held with ISMRM Perfusion SG members in which the concept of this initiative was explained and any queries were addressed.

Stage 5 (February 2022–April 2022): A total of 38 responses to the survey were collected and are summarized in Figures S1–S3, which are available online. Based on those responses, the reporting recommendation was finalized and is provided in section 3: Reporting Recommendation. United Imaging Healthcare (Shanghai) also joined the manufacturer survey, and the summary of the survey responses from all six MRI manufacturers is provided in Figures S4 and S5, showing how the acquisition parameters listed in the Reporting Recommendation can

be obtained via the commercial MRI scanner GUIs. It was found that, when this specific sequence/technique is implemented as a product, all corresponding parameters in the Required category were either displayed in the GUI or available on request from the manufacturers. In the Recommended category, however, several parameters are not available for some manufacturers. Therefore, the recommendation level remains that we only “encourage” authors to include as many of the recommended parameters as possible in ASL publications. After all feedback/comments were addressed, the ASL Lexicon was divided into two groups: (a) techniques and their parameters that are widely used and mature enough to be standardized, which are mostly covered by the ASL White Paper and some (but not all) of the ASL Gray Papers; and (b) advanced and emerging techniques and their parameters. Only the former (i.e. [a]) is included in this paper, to avoid premature standardization of emerging techniques in a published paper. The final draft of this paper was shared with the ISMRM Perfusion SG members for endorsement.

May 2022–future: The online version of the ASL Lexicon and Reporting Recommendation will be managed and updated by the community as further methods and improvements are developed.

2 | ASL LEXICON

The ASL lexicon organizes comprehensive lists of terminology and definitions for ASL imaging techniques and acquisition parameters, as well as physiological constants and parameters required in quantitative analysis. As explained in subsection 1.2 (stage 5), this paper contains only techniques and parameters that are widely used and mature enough to be standardized, that is: (pseudo-)continuous ASL ((P)CASL), pulsed ASL (PASL), velocity-selective ASL (VSASL), and multi-timepoint ASL

for brain perfusion imaging. Other more advanced and emerging techniques (e.g., vessel-selective ASL, MR fingerprinting ASL, ASL angiography, modified ASL labeling methods that measure other physiological parameters [e.g., water extraction fraction], and corresponding image processing) and ASL applications in the body will be listed on the online version of ASL lexicon that is available at <https://osipi.ismrm.org/task-forces/tf4-1/>, with a view to standardization in future work.

2.1 | General definition of ASL

In this subsection, the basic structural elements of the standard ASL sequence are listed and defined. In this paper, the term *labeling* is used in the description throughout all techniques, which is, however, interchangeable with *tagging*.

- **ASL:** Any MRI technique in which contrast is generated by manipulation of the arterial blood magnetization using RF pulses prior to image acquisition, with the aim of isolating flow signal for angiography/perfusion imaging.
- **Labeling pulse:** RF pulse, or train of RF pulses, intended to change the magnetization state of blood in order to differentiate it from stationary tissue. In general, labeling pulses can be spatially selective, targeting the blood outside the imaging volume (upstream), or velocity-/acceleration-selective, targeting blood without special selectivity (i.e., including blood flowing within the imaging volume) according to its velocity or acceleration.
- **Control pulse:** RF pulse, or train of RF pulses, intended to match the static tissue magnetization transfer (MT), diffusion, eddy currents, or any other side effects of the labeling pulse, while causing minimal perturbation to arterial blood.




FIGURE 1 Schematic sequence diagram of QUIPSS-II and Q2TIPS. Abbreviation: Q2TIPS, QUIPSS II with thin-slice TI_1 periodic saturation QUIPSS-II, quantitative imaging of perfusion using a single subtraction II; TI , inversion time.

FIGURE 2 Schematic sequence diagram of VSASL (based on Figure 3 in Qin et al.⁶). Abbreviation: BS, background suppression; TI, inversion time; T_{sat} , saturation time; VSASL, velocity-selective arterial spin labeling; VSI, velocity-selective inversion; VSS, velocity-selective saturation.





FIGURE 3 Schematic sequence diagram of PCASL. The timing parameter for the BS pulses. In this example, two BS pulses are used. Currently, there are three different definitions of the timings implemented in commercial scanners: (A) time from the center of the first PCASL labeling pulse to the center of the Nth BS pulse, (B) time from the center of the first excitation pulse to the center of the Nth BS pulse, (C) BS_n Time: from the center of the last PCASL labeling pulse to the center of the Nth BS pulse; BS_n TI: from the center of the Nth BS pulse to the center of the first excitation pulse (Figure [c] courtesy: Canon Medical Systems Corporation) Abbreviation: BS, background suppression; LD, labeling duration; PCASL, pseudo-continuous arterial spin labeling; PLD, post-labeling delay; TI, inversion time; TR, repetition time.

- Labeled image: The image acquired after preparation by a labeling pulse.
- Control image: The image acquired after preparation by a control pulse.
- Delay time: The time interval between the labeling and the image readout that allows the labeled arterial blood bolus to reach the tissue of interest. In general, it is called the *post-labeling delay* (PLD) in (P)CASL and *inversion time* (TI) in PASL. In VSASL, both TI and PLD are used to define different temporal parameters. See subsection 2.3, Parameters in ASL labeling method, Table 5, and Figures 1–3 below for more details.
- Single PLD/TI: ASL protocol in which images are acquired with a single delay time.
- Multiple PLD/TI: ASL protocol in which images are acquired with multiple (more than one) delay times.
- Background suppression (also often abbreviated as BS): The strategy for reduction of static tissue signal intensity using a train of RF pulses applied prior to image readout. The aim of background suppression is to improve SNR of the ASL image by reducing signal fluctuations in the labeled and control images. There are many background suppression schemes available, involving saturation and inversion pulses.
- Saturation: The saturation of the imaging volume is performed just before and/or after the labeling and control pulses to set its longitudinal magnetization to zero and thereby eliminate any label/control MT or slice profile mismatches. Water suppression enhanced through T₁ effects (WET) pulses¹⁰ are commonly used.¹¹
- Vascular suppression (also known as *vascular crushing*): The reduction of signal present in larger arterial vessels at the time of imaging. Generally, it is achieved by applying vascular crusher gradients after the excitation pulse, which selectively dephases signal based on the velocity profile of the spins in the direction of the gradient.^{12–14} In VSASL, vascular suppression is achieved using velocity selective saturation pulses.⁶
- M0 image (also known as *proton density image*): The additional calibration image required for blood flow quantification, used to estimate the fully relaxed magnetization (M0) of blood (M0b) and tissue (M0t), which are necessary to calculate perfusion from ASL images (see subsection 2.6 and Table 8). M0 image is commonly obtained as a proton density image by turning off all preparation pulses before acquisition while using a relatively long TR. When background suppression pulses are not applied, the average (mean) control image can be used as the M0 image by correcting for T₁ relaxation.

2.2 | ASL labeling methods

This subsection focuses on the name of the techniques and their notations and descriptions with regard to ASL labeling methods. In general, ASL labeling methods are divided into three labeling types: (P)CASL, PASL, and VSASL. In addition to these labeling methods, this subsection also covers the techniques/sequences for multi-timepoint ASL.

2.2.1 | (Pseudo-) continuous ASL ((P)CASL)

CASL^{15–17} and *PCASL*¹⁸ are general terms for the ASL labeling methods in which labeling is performed by applying RF pulses for long duration (typically 1–3 s) in combination with a magnetic field gradient. Flowing blood spins are inverted as they flow through a thin labeling plane by means of flow-driven adiabatic inversion.¹⁹ In the ASL White Paper, PCASL is the recommended ASL labeling method for clinical use due to its high SNR efficiency compared to PASL.¹ Several techniques related to (P)CASL are listed in Table 1.

2.2.2 | Pulsed ASL (PASL)

PASL^{23–25} is a general term for the ASL labeling method in which the labeling is performed by applying a single (or a limited number of) short RF pulse(s) that instantaneously invert the blood magnetization. In general, PASL labeling methods are grouped into two types: (i) asymmetric PASL, in which a spatially selective RF slab (typically 10–20 ms) labels the spins outside of the imaging volume on the upstream side (i.e., neck for imaging of brain), for example, Echo-planar imaging and signal targeting with alternating radiofrequency (EPISTAR)²⁶; and (ii) symmetric PASL, in which the label is performed by a nonselective global inversion pulse (e.g., Flow-sensitive alternating inversion recovery (FAIR)²⁴) and spins outside of the imaging volume are labeled, regardless of whether they are upstream or downstream, symmetrically. In PASL, the bolus duration of labeled blood is unknown a priori and depends on the labeling slab thickness and blood flow velocity. To achieve accurate quantification of CBF with a single TI acquisition, a bolus cutoff technique, such as quantitative imaging of perfusion using a single subtraction II (QUIPPS-II)²⁷ and QUIPPS-II with thin-slice T₁ periodic saturation (Q2TIPS),²⁸ is applied to define the bolus duration. Figure 1 shows schematic sequence diagrams of QUIPPS-II and Q2TIPS. Several labeling methods related to PASL are listed in Table 2.

2.2.3 | Velocity-selective ASL (VSASL)

VSASL^{6,36} is a general term for the ASL labeling method in which the magnetization is labeled by saturation or

TABLE 1 (P)CASL sequences

Name	Notation	Description
Continuous ASL ¹⁵⁻¹⁷	CASL	A single, continuous-wave, RF pulse applied over a long period, typically 1–3 s, in combination with a constant magnetic field gradient. Arterial blood is continuously inverted as it flows through a specified labeling plane by means of flow-driven adiabatic inversion. ¹⁹
Pseudo-continuous ASL ¹⁸	PCASL	Similar to CASL, labeling occurs over a long period, typically 1.5–2 s, and inverts flowing arterial blood. In PCASL, however, a train of short RF pulses applied at a rate of approximately 1 per ms replaces the single, continuous pulse of CASL. For the control scan, phase modulation of the RF pulse train is applied such that MT effects are identical to the labeling scan, but the arterial blood magnetization is unperturbed. Note that the mechanism of arterial blood inversion is equivalent for CASL and PCASL and, consequently, the quantification model is the same.
Balanced PCASL	bPCASL	PCASL implementation that uses the same gradient waveform for the label and control pulses.
Unbalanced PCASL	ubPCASL	PCASL implementation that uses different gradient waveforms for the label and control pulses, so that the G_{av} becomes zero for the control. If optimized, improved robustness to off-resonance effects at the labeling plane can be achieved compared to bPCASL. ²⁰
Separate RF labeling coils ^{21,22}		Using separate, dedicated RF transmission coils (i.e., in addition to the RF transmit coils for imaging) positioned over the artery/arteries of interest, to reduce power deposition and avoid MT effects in the perfused organ.
Flow-encoding arterial spin tagging ¹³	FEAST	A technique based on (P)CASL that acquires a pair of ASL subtraction images with and without crusher gradients for vascular suppression. The ATT is calculated using the ratio between (P)CASL images with and without vascular suppression.

Abbreviation: ASL, arterial spin labeling; ATT, arterial transit time; G_{av} , average gradient; MT, magnetization transfer.

inversion based on its velocity. See Figure 2 for a general schematic diagram. In the original implementation, the saturation of flowing blood signal is achieved using a double-refocused hyperbolic secant/tangent (DRHS/T)³⁷ or B_1 -insensitive rotation-8 (BIR-8)³⁸ pulse train in combination with velocity-encoding gradients in subsequent implementations. Control images are acquired without velocity-encoding gradients. Several variant approaches in VSASL are listed in Table 3. Detailed descriptions of the VSASL technique and recommendations for clinical application can be found in the recent ASL Gray Paper.⁶

2.2.4 | Multi-timepoint ASL

*Multi-timepoint ASL*⁴³⁻⁴⁶ is a general term for ASL techniques in which data are acquired repeatedly with several time parameters (delay time, and/or labeling duration for PCASL) to observe the kinetic ASL signal. These approaches are often called *multi-delay ASL*, particularly when only the delay time is varied. *Multi-phase ASL* is also sometimes used as a synonym of multi-delay ASL; however, this should not be confused with the PCASL approach that uses a range of RF phase offsets in the PCASL pulse train to reduce the sensitivity of the

CBF estimation to B_0 inhomogeneity.⁴⁷ Table 4 shows several approaches to achieve multi-timepoint ASL. Detailed descriptions and recommendations for the use of multi-timepoint ASL can be found in the relevant recent ASL Gray Paper.⁹

2.3 | Parameters in ASL labeling method

The parameters related to ASL labeling methods are provided in Table 5. In general, the names and definitions of parameters comply with the ASL White Paper,¹ as well as ASL-BIDS⁴; otherwise, the difference is provided in the description. In general, for timing parameters there is no preference between the use of “ms” and “s.” In ASL-BIDS, however, the values in JSON sidecars are entered without specifying the units; therefore, the use of correct units, as specified in the ASL-BIDS definition, should be strictly followed.

In ASL perfusion imaging, the application of background suppression is recommended. It should be noted that, currently, there are three different ways in which the background suppression pulse timings are defined by six major MRI manufacturers (Canon, Fujifilm, GE, Philips, Siemens, and United Imaging; in alphabetical order), which are shown in Figure 3.

TABLE 2 PASL sequences

Name	Notation	Description
Pulsed ASL	PASL	A general term for ASL methods with a single (or a limited number of) short RF pulse(s) (typically 10–20 ms) applied to “instantaneously” invert a slab of arterial blood magnetization.
Echo-planar imaging and signal targeting with alternating radiofrequency ²⁶	EPISTAR, (also known as STAR for non-EPI readout)	A variation of PASL in which the label is performed by a slab-selective adiabatic inversion pulse applied proximal to the imaging volume/slices. In the original multi-slice implementation, the control preparation was achieved by applying slab-selective RF pulses over the same region as the label, with total power matched to the labeling inversion pulse (to generate the same MT effects across the imaging volume) but which resulted in minimal perturbation of the arterial blood magnetization, for example, two consecutive inversion pulses with half power. In the product implementation by Philips Healthcare, however, the control preparation is achieved by mirroring the frequency modulation for the second part of the adiabatic pulse. ²⁹
Flow-sensitive alternating inversion recovery ²⁴	FAIR	A variation of PASL in which the label is performed by a non-slice-selective global inversion pulse, whereas the control image is obtained using a slice-selective inversion pulse applied to the imaging slab. Because of this symmetric nature, FAIR allows the inflow of the labeled blood from both sides of the imaging volume.
Proximal inversion with control for off-resonance effects ³⁰	PICORE	A variation of PASL, in which the label is the same as in EPISTAR, whereas the control image is obtained using an off-resonance inversion pulse that is applied with the same frequency offset as the label but without a slab-selective gradient.
Double inversions with proximal labeling of both tag and control images ³¹	DIPLOMA	A variation of PASL designed to reduce the residual MT mismatch between the label and control images observed in EPISTAR and PICORE. In both label and control, two consecutive adiabatic inversion pulses are applied; in the labeling preparation, application of an off-resonance inversion pulse (similar to the one applied in PICORE for control preparation) is followed by a slab-selective inversion pulse. In the control preparation, two slab-selective inversion pulses are applied.
Transfer-insensitive labeling technique ³²	TILT	A variation of PASL in which labeling is achieved by two successive 90° RF pulses. For the control, the phase of the second pulse is shifted by 180°, thereby yielding no net effect on blood water magnetization.
Bolus cutoff technique		In PASL, the bolus duration (see Table 5 for the definition) of labeled blood is unknown a priori and depends on the labeling slab thickness and blood flow velocity. To achieve accurate quantification of CBF with a single TI acquisition, several techniques have been proposed to define the bolus duration, as described below.
Quantitative imaging of perfusion using a single subtraction ²⁷	QUIPSS	QUIPSS aims to eliminate arterial transit time effects in PASL, to enable reliable quantification of CBF with a single TI acquisition. This is achieved by applying a saturation RF pulse to the imaging volume at a time TI1 after labeling, when TI1 is greater than the arterial transit time, followed by image acquisition at time TI. N.B. this approach has not been widely adopted due to the prevalence of intravascular signal in the ASL difference images.
Quantitative imaging of perfusion using a single subtraction II ²⁸	QUIPSS-II	QUIPSS-II aims to control the bolus duration in PASL and allow reliable quantification of CBF when using PASL with a single TI. This is achieved by applying a saturation RF slab to the area in which the labeling RF slab is applied, thereby cutting off the “tail” of the labeled bolus. See “Bolus duration” in Table 5 for the definition.
QUIPPS II with thin-slice TI ₁ periodic saturation ²⁸	Q2TIPS	Modified version of QUIPSS-II, aiming to improve the saturation efficiency by replacing the QUIPSS-II saturation pulse with multiple thin RF saturation pulses applied at the distal edge of the labeling slab.
QUIPSS II with window-sliding saturation sequence ³³	Q2WISE	A hybrid technique between Q2TIPS and QUIPSS II. In Q2WISE, saturation is achieved by using two thin saturation pulses and one thick slab saturation pulse to reduce the RF power deposition.
Wedge-shaped PASL ³⁴	WS-PASL	A variation of PASL in which a wedge-shaped adiabatic inversion pulse is used to directly control the bolus duration in different vessels based on the flow velocity.
Attenuating the static signal in arterial spin tagging ³⁵	ASSIST	FAIR ASL approach with multiple inversion pulses during the TI to suppress static tissue signal (first implementation of background suppression with ASL).

Abbreviation: CBF, cerebral blood flow; MT, magnetization transfer; TI, inversion time.

TABLE 3 VSASL sequences

Name	Notation	Description
Velocity-selective ASL ^{6,36}	VSASL	A general term for ASL techniques in which the magnetization is labeled by saturation or inversion based on its velocity. See Figure 2 for a general schematic diagram. In the original implementation, the saturation of flowing blood signal is achieved using a DRHS/T ³⁷ or BIR-8 ³⁸ pulse train in combination with velocity-encoding gradients. Control images are acquired without velocity-encoding gradients.
Fourier-transform-based velocity-selective saturation ASL ^{39,40}	FT-VSS ASL	A variation of VSASL in which the magnetization within a certain velocity band is saturated by a train of composite pulses incorporating velocity-sensitive bipolar gradients and refocusing 180° pulses. In contrast to the above-mentioned original implementation of VSASL, in FT-VSS-ASL, the static magnetization is saturated while preserving the magnetization flowing above the velocity threshold. In the control image acquisition, all magnetization is saturated.
Fourier-transform-based velocity-selective inversion ASL ³⁷	FT-VSI ASL	Analogous to the FT-VSS ASL method described above, FT-VSI ASL uses composite velocity-selective inversion pulses to invert both flowing and static tissue magnetization (label) or only the static tissue magnetization (control). SNR is improved compared with saturation-based VSASL.
Multi-module velocity-selective ASL ⁴¹	mm-VSASL	A strategy to measure ASL signal with multiple VS labeling modules to increase labeling bolus duration and reduce T ₁ relaxation of the ASL signal. This method provides improved SNR compared to conventional single-module VSASL with VS saturation.
Acceleration-selective ASL ⁴²	AccASL	An extension of VSASL that labels (saturates) based on the acceleration/deceleration of blood spins rather than their velocity. Because arterial blood exhibits stronger acceleration/deceleration, it labels predominantly arterial (as opposed to venous) blood. AccASL includes both CBF and CBV weighting.

Abbreviation: BIR-8, B₁-insensitive rotation-8; CBF, cerebral blood flow; CBV, cerebral blood volume; DRHS/T, double-refocused hyperbolic secant/tangent.

TABLE 4 Multi-timepoint ASL

Name	Notation	Description
Multi-timepoint ASL		A general term for ASL techniques in which ASL data are acquired repeatedly with varied time parameters (delay time, and/or labeling duration for PCASL) to observe the kinetic ASL signal. Also often called <i>multi-delay ASL</i> , particularly when only the delay time is varied.
Multi-timepoint sequential ASL ^{44,48}		Multi-time point ASL acquisition that acquires images with multiple timepoints as successive single-TI/PLD scans, as opposed to LL-ASL or time-encoded PCASL.
Look-Locker ASL ⁴⁹	LL-ASL	ASL acquisitions in which several images are acquired at multiple time points after a single labeling module. Readouts with low flip angle are used to reduce saturation of the labeled blood by the first readouts.
Quantitative STAR labeling of arterial regions ¹²	QUASAR	PASL-based sequence that consists of several Look-Locker readouts: (i) with and without vascular suppression to obtain local arterial input function; (ii) two different readout flip angles. A Q2TIPS-like saturation is used to define the bolus duration. This sequence allows measurement of the local AIF and quantification of CBF by deconvolution.
Reduced resolution transit delay prescan ⁵⁰		A fast multi-timepoint ASL implementation that is specifically used to acquire low spatial resolution arterial transit time maps. Typically used as an ancillary scan to enhance the quantification accuracy of a standard-resolution single-PLD ASL acquisition.
Time-encoded PCASL (also commonly referred to as <i>Hadamard-encoded</i>) ⁵¹	te-PCASL	Segmenting the PCASL labeling/control module into varying control and label sub-periods according to an encoding matrix. This improves the temporal efficiency of multi-timepoint ASL; that is, it reduces the number of acquisitions required for a multi-PLD data set. The most typical implementation is Hadamard encoding. Modifications include, for example, Walsh-ordering. ⁵²

Abbreviation: AIF, arterial input function; PLD, post-labeling delay; TI, inversion time.

TABLE 5 Parameters in ASL labeling method

Name	Notation	Unit	Description
Labeling duration	LD; also known as τ	ms or s	For CASL/PCASL. Duration of the constant CASL labeling RF or PCASL labeling pulse train (see Figure 3).
Bolus duration	BD	ms or s	For PASL. Temporal width of the labeled bolus. For QUIPSS-II/Q2TIPS, this is defined as the time from the labeling pulse to the center of the first saturation pulse (and is equal to TI_1 , see below). If QUIPSS-II/Q2TIPS saturation is not used, this parameter is not known a priori but is determined by the arterial blood velocity and inversion slab thickness, or the RF transmit coil length for FAIR.
Post-labeling delay	PLD	ms or s	For CASL/PCASL. Time from the end of the labeling pulse to the center of the imaging excitation pulse (see Figure 3). In a 2D multi-slice acquisition, the PLD is defined by the time of the first slice acquisition; however, it is important to note that the effective PLD for each slice is different and is determined by the PLD and the interslice time (see Table 6).
Inversion time	TI	ms or s	For general PASL. Time from the center of the labeling pulse to the center of the imaging excitation pulse. In 2D multi-slice acquisition, this relates to the first acquired slice. ASL-BIDS uses the term <i>post-labeling delay</i> for this parameter in PASL.
Inflow time		ms or s	In some post-processing tools (e.g., FSL), inflow time is used to define the time from the start of labeling to the center of the imaging excitation pulse. For PASL, it is equivalent to inversion time. For PCASL, however, inflow time will be equivalent to PLD + LD.
	TI_1	ms or s	For QUIPSS (-II)/Q2TIPS. Time from the center of the labeling pulse to the center of the bolus saturation pulse (QUIPSS [-II]) or center of the first saturation pulse (Q2TIPS) (see Figure 1). BIDS uses the term <i>BolusCutOffDelayTime</i> (1).
	TI_2	ms or s	For QUIPSS (-II)/Q2TIPS. Time from the center of the labeling pulse to the center of the excitation pulse of the image acquisition. This value is equivalent to TI of the conventional (non-QUIPSS-II/Q2TIPS) PASL (see Figure 1).
	ΔTI	ms or s	For QUIPSS (-II), defined as $TI_2 - TI_1$ (see Figure 1).
TI_1 stop	TI_{1s}	ms or s	For Q2TIPS. Time from the center of the labeling pulse to the center of the last bolus saturation pulse (see Figure 1). BIDS uses the term <i>BolusCutOffDelayTime</i> (2).
Background suppression (pulse) timing	BS_1 to BS_n	ms or s	The timing parameters for the background suppression RF pulses. Currently, there are three different definitions for these timings implemented in commercial scanners (please see Figure 3): (A) time from the center of either the labeling pulse (for PASL) or the first labeling pulse (PCASL) to the center of the Nth background suppression pulse; (B) time from the center of the Nth background suppression to the center of the first readout excitation pulse; and (C) BS_n time from the center of the last PCASL labeling pulse to the center of the Nth BS pulse, and BS_n TI from the center of the Nth BS pulse to the center of the first excitation pulse.
Vascular crusher gradient strength	V_{enc} B	cm/s s/mm ²	Crusher gradients have an amplitude sufficient to cause a 180° phase shift for blood moving with a velocity of V_{enc} in the direction of the gradients. Crusher gradients are equivalent to diffusion-weighting gradients with this b value. ⁵³
Labeling plane			Plane at which flowing blood is labeled in (P)CASL. <i>Fixed labeling plane</i> means the labeling plane is parallel to the image slice orientation and angulation with a specified distance relative to the lowest slice. <i>Free labeling plane</i> means the labeling plane can be moved and angulated independently from the image volume.
Labeling plane offset/distance		mm	For PCASL. This is the distance between the center of the imaging volume and the center of the labeling plane.
Labeling pulse average gradient	G_{av}	mT/m	For PCASL. The non-zero mean gradient applied concurrently with the RF labeling pulses, which combines to produce flow-driven adiabatic inversion (see Figure 4).

TABLE 5 (Continued)

Name	Notation	Unit	Description
Labeling pulse maximum gradient	G_{\max}	mT/m	For PCASL. The amplitude of the slice-selection gradient applied during the labeling pulses in the PCASL labeling pulse train (see Figure 4).
Labeling pulse average B1	$B_{1\text{av}}$	μT	For (P)CASL. The average B_1 -field strength of the RF labeling pulses over the entire pulse train (see Figure 4).
Labeling pulse flip angle		degree ($^\circ$)	For PCASL. The flip angle of a single labeling pulse in the pCASL labeling pulse train.
Labeling pulse interval		ms	For PCASL. The interval between the centers of two successive PCASL labeling pulses (see Figure 4).
Labeling pulse duration		ms	For PCASL. The duration of each PCASL labeling pulse (see Figure 4).
PCASL control type			For PCASL. Type of the gradient scheme used in pCASL control condition: either balanced or unbalanced. <i>Balanced</i> : Identical G_{av} (non-zero) for label and control. <i>Unbalanced</i> : G_{av} in label is nonzero but zero in control (refocusing gradient lobes are increased in amplitude such that the mean gradient is zero).
Labeling slab			For PASL, the volume over which the labeling RF pulse is applied.
Labeling slab thickness		mm	For PASL. The nominal thickness of the labeling slab.
Labeling slab gap		mm	For PASL, this is the nominal gap between the leading edge of the labeling slab and the closest edge of the imaging volume.
Cutoff velocity	V_{cut}		In VSASL, spins moving above a chosen velocity, referred to as the <i>cutoff velocity</i> (V_{cut}), is labeled. V_{cut} determines how deep into the arterial tree the blood is labeled.

Abbreviation: BIDS, Brain Image Data Structure; FSL, FMRIB software library.

2.4 | Readout sequences and parameters

In this subsection, the basic readout sequences and parameters that appeared in the ASL White Paper¹ are listed (see Table 6). More advanced readout strategies can be found in the advanced ASL Gray Paper.⁷

2.5 | Derived parameters

Table 7 provides a list of the derivative parameters of standard ASL, namely that commonly appear in the perfusion imaging using single PLD. In general, the names and definitions of parameters comply with the ASL White Paper,¹ as well as ASL-BIDS⁴; otherwise, the difference is provided in the description.

2.6 | Ancillary parameters for quantification

This subsection focuses on the name, notations, and descriptions of the physiological constants and ancillary




FIGURE 4 Schematic diagram of PCASL labeling pulse train.

parameters used in ASL quantification, including equations for quantification.

2.6.1 | One-compartment model for single-PLD

The general kinetic model is used to derive the CBF quantification equations below. Several assumptions need to be

TABLE 6 Readout sequences and parameters

Name	Notation	Unit	Description
Echo-planar imaging	EPI		A 2D rapid imaging technique in which an excitation pulse is followed by acquisition of multiple k-space lines by switching the readout gradient polarity rapidly and applying phase-encoding blips. In single-shot EPI, all k-space lines are collected after a single excitation pulse, making it robust to motion.
Gradient and spin echo	GRASE		Rapid imaging technique in which the excitation pulse is followed by several refocusing pulses (similar to fast/turbo spin echo), and after each refocusing pulse, a series of gradient echoes are collected by rapidly switching the readout gradient polarity (similar to EPI). The use of refocusing RF pulses prolongs the lifetime of the transverse magnetization compared to EPI. Typically acquired as a multi-shot 3D acquisition in ASL applications.
Stack-of-spirals RARE or FSE	SoS		Non-Cartesian 3D fast spin-echo (also known as <i>turbo spin-echo</i> , or <i>RARE</i>) acquisition technique, in which the readout is performed using a spiral trajectory to efficiently sample kx-ky, with each spin echo being assigned to a different kz partition. Typically acquired as a multi-shot 3D acquisition in ASL applications.
Segmented 3D sequence			3D acquisition scheme (e.g., 3D GRASE or SoS) in which k-space is acquired over multiple TRs to keep each individual readout to a reasonable duration. It should be noted that, as compared to the single-shot sequence, this approach is more sensitive to motion. Also known as <i>multi-shot 3D</i> sequences.
Number of segments/shots	N_{seg}		In a segmented 3D sequence, this is the number of acquisition repeats required to sample the full k-space data set.
Repetition time	TR	ms	The time from the beginning of a labeling/control pulse to the beginning of the next control/labeling pulse. Note that, when a readout sequence with multiple excitation pulses (e.g., balanced SSFP) is used, this TR is different to the TR of the readout sequence. RepetitionTimePreparation is used in BIDS to differentiate this from RepetitionTimeExcitation.
Total acquired pair			The number of paired labeled and control images acquired for improving SNR (averaging) in single-delay ASL, or for fitting in multi-time point ASL. Note that, if online averaging is performed, this number will be greater than the number of reconstructed image pairs; in the extreme, the latter may be a single image pair, representing the average over all acquisitions. NB for te-PCASL, images are not acquired in label-control pairs; therefore, in this situation it is appropriate to specify the number of repeats of the full encoding matrix.
Interslice time		ms	For a 2D multi-slice acquisition scheme, the time between the excitation pulses of successive slices. This is needed in order to calculate the effective PLD/TI for each slice, which is required for accurate quantification.

Abbreviation: RARE, rapid acquisition with relaxation enhancement. SSFP, steady state free precession

fulfilled to ensure its validity—for example, delivery of the entire bolus to the tissue and that label relaxation is governed by blood T_1 during the entire measurement. This is the basic quantification model recommended by the ASL White Paper.¹ A list of the parameters used in these equations is provided in Table 8.

PCASL^{1,43}

$$\text{CBF} = \frac{6000 \cdot \lambda \cdot \Delta M \cdot e^{\frac{\text{PLD}}{T_{1b}}}}{2 \cdot \alpha \cdot T_{1b} \cdot M_{0t} \cdot \left(1 - e^{-\frac{\tau}{T_{1b}}}\right)} [\text{mL/min/100 g}]$$

PASL^{1,27}

$$\text{CBF} = \frac{6000 \cdot \lambda \cdot \Delta M \cdot e^{\frac{\text{TI}}{T_{1b}}}}{2 \cdot \alpha \cdot T_{1b} \cdot M_{0t}} [\text{mL/min/100 g}]$$

3 | REPORTING RECOMMENDATION

The Reporting Recommendation is provided in Table 9 and consists of two recommendation levels:

TABLE 7 Derivative parameters

Name	Notation	Unit	Description
ASL difference image	ΔM	Arbitrary unit	Image obtained by subtracting the labeled image from the control image, which subtracts out the static tissue signal and consequently shows the perfusion-weighted signal produced by ASL preparation.
Also known as <i>DeltaM</i>			
Also known as <i>Perfusion weighted image</i>			
Normalized perfusion weighted image	$\Delta M/M_0$	%	ASL difference image normalized by the M_0 , with a unit in %.
Cerebral blood flow	CBF	mL/100 g/min	Quantity of blood (mL) reaching 100 g of brain tissue per unit of time (min).
Arterial transit time	ATT	ms	Time between when blood is labeled and when it first arrives in the imaging voxel/slice.
Also known as <i>bolus arrival time</i>	BAT		Note that ATT, AAT, and BAT are dependent on the positioning of the labeling plane/slab relative to the imaging volume, and are therefore not generally comparable across studies.
Also known as <i>Arterial arrival time</i>	AAT		
Partial volume ^{54,55}	PV		The typical voxel size of ASL perfusion image is much larger than the cortical thickness, and individual voxels are likely to contain a mixture of GM, WM, and CSF, which is known as the <i>PV effect</i> .
Tissue partial volume	P_{GM}, P_{WM}	Fraction (0–1)	Partial volume of different tissue types (GM, WM, CSF) as a fraction of the total voxel volume.
Also known as <i>fractional tissue volume</i>			
Tissue specific perfusion	CBF_{GM}, CBF_{WM}		Perfusion of specific tissue types within a voxel, estimated either by (a) including only voxels with a tissue probability value higher than a stated threshold, or (b) explicitly correcting for the partial volume effects of GM/WM/CSF within voxels.

Abbreviation: AAT, Arterial arrival time; BAT, bolus arrival time; CSF, cerebrospinal fluid; GM, gray matter; ; WM, white matter.

- **Required: essential for meaningful interpretation of the ASL data and for quantitative analysis.** These must be included for describing ASL methods in reports/articles in order for its data set to be *OSIPI-compliant*.
- **Recommended: parameters that are useful for interpretation of the ASL data and could explain specific characteristics or systematic differences between data sets.** Authors are encouraged to include as many of these as possible in ASL publications.

4 | SUMMARY AND CONCLUSION

On behalf of the ISMRM Perfusion SG, this paper is intended to form a community-endorsed lexicon and recommendation for reporting of ASL perfusion imaging, detailing which parameters in acquisition protocols and analysis should be reported and how, with the aim of improving the reproducibility and consistency of the reported studies. In the future, this lexicon could also be used to improve the Digital Imaging and Communications in Medicine (DICOM) standard for the purposes of

TABLE 8 Parameters for the one-compartment model for single-PLD

Name	Notation	Unit	Description
T ₁ relaxation time of blood Also known as <i>blood</i> T ₁	T _{1b}	ms	The longitudinal relaxation time of arterial blood
Equilibrium magnetization of blood Also known as <i>M₀ of blood</i>	M _{0b}	Arbitrary unit	Fully relaxed longitudinal magnetization of arterial blood, which is required to scale the subtracted ASL signal and obtain absolute CBF units. In the ASL White paper, it is recommended to estimate M _{0b} from a voxel-by-voxel M _{0t} measured by an M ₀ image. The blood–brain partition coefficient λ scales M _{0t} to M _{0b} .
Equilibrium magnetization of tissue Also known as <i>M₀ of tissue</i>	M _{0t}	Arbitrary unit	Fully relaxed longitudinal magnetization of tissue. This value might be different in different organs or tissue types within an organ.
Blood–brain partition coefficient	λ	mL/g	The ratio between blood and tissue water concentration at equilibrium in mL of blood, per g of tissue. When used in ASL quantification, instantaneous equilibrium between tissue and veins is assumed
Labeling efficiency	α	Fraction (0–1)	Combines the inversion efficiency of the labeling pulse itself and the loss of label caused by background suppression (dependent on the number and type of BS pulses). A value of 1 corresponds to full inversion of blood magnetization.

TABLE 9 Reporting recommendation

Name	Abbreviation	Style	Condition
Required parameters: General			
Arterial spin labeling type		PASL, (P)CASL, Velocity-selective, etc.	
Background suppression	BS	Yes/No	
Method for M _{0b} estimation		The description of how M _{0b} is estimated, e.g., how M ₀ image is acquired, or any special method to estimate M _{0b} directly.	When CBF estimation is performed
Total acquired pairs		The number of paired labeled and control images, before online averaging is performed (if applicable)	
Acquisition voxel size		Value in mm	
Required parameters: (P)CASL			
Labeling duration	LD or τ	Value in ms or s	When (P)CASL is used
Post-labeling delay	PLD	Value in ms or s	When (P)CASL is used
Required parameters: PASL			
Inversion time/inflow time	TI	Value in ms or s	When PASL is used
Bolus cutoff techniques		Name of technique, or None	When PASL is used
	TI ₁	Value in ms or s	When QUIPSS-II or Q2TIPS are used
	TI ₂	Value in ms or s	When QUIPSS-II or Q2TIPS are used
	TI _{1s}	Value in ms or s	When Q2TIPS is used

TABLE 9 (Continued)

Name	Abbreviation	Style	Condition
Recommended parameters: General			
Number of background suppression pulse		Value	If BS is used and details are available to the user
Background suppression (pulse) timing	BS ₁ to BS _n or BS time/TI	Value in ms	If BS is used and details are available to the user
Background suppression timing definition		Description of how timing is defined. See “Background suppression (pulse) timing” in Table 5, or Figure 3.	If BS is used and details are available to the user
Labeling location description		Description of the labeling plane/slab location (other factor than offset/gap), such as the planning of the labeling plane/slab with respect to the imaging slices	
Shim volume		Description of shim volume used, e.g., imaging volume only, both imaging volume and labeling region, labeling region during labeling pulse and imaging volume during acquisition, or other (please specify)	
Vascular crushing	V _{enc}	Value in cm/s	When vascular crushing is performed. Ideally, both V _{enc} and b should be specified
	b	Value in s/mm ²	
Recommended parameters: (P)CASL			
PCASL control type		Balanced or unbalanced	When PCASL is used
CASL type		If a separate coil is used for labeling	When CASL is used
Labeling plane offset/distance		Value in mm	
Labeling pulse average gradient	G _{av}	Value in mT/m	When details are available to the user
Labeling pulse maximum gradient	G _{max}	Value in mT/m	When details are available to the user
Labeling pulse average B ₁	B _{1av}	Value in μT	When details are available to the user
Labeling pulse flip angle		Value in degree (°)	When details are available to the user
Labeling pulse interval		Value in ms	When details are available to the user
Labeling pulse duration		Value in ms	When details are available to the user
Recommended parameters: PASL			
PASL type		EPSTAR, FAIR, PICORE, etc.	
Labeling slab thickness		Value in mm	
Labeling slab gap		Value in mm	

communicating raw images and parametric maps of ASL perfusion MRI.

AFFILIATIONS

¹Wellcome Centre for Integrative Neuroimaging, FMRIB, Nuffield Department of Clinical Neurosciences, University of Oxford, Oxford, UK

²Department of Medical Imaging, Ghent University Hospital, Ghent, Belgium

³Department of Diagnostic Sciences, Ghent University, Ghent, Belgium

⁴State University of New York at Binghamton, Binghamton, New York, USA

⁵Department of Radiology, University of Pennsylvania, Philadelphia, Pennsylvania, USA

⁶Department of Radiology, Clínica Universidad de Navarra, Pamplona, Spain

⁷IdiSNA, Instituto de Investigación Sanitaria de Navarra, Pamplona, Spain

Department of Diagnostic and Interventional Neuroradiology, ⁸University Hospital Hamburg-Eppendorf, Hamburg, Germany

⁹Department of Radiology and Nuclear Medicine, Amsterdam Neuroscience, Amsterdam University Medical Center, Amsterdam, The Netherlands

¹⁰Helmholtz-Zentrum Dresden-Rossendorf, Institute of Radiopharmaceutical Cancer Research, Dresden, Germany

¹¹Laboratory of FMRI Technology (LOFT), Mark & Mary Stevens Neuroimaging and Informatics Institute, Keck School of Medicine, University of Southern California, Los Angeles, California, USA

¹²Division of MRI Research, Department of Radiology, Beth Israel Deaconess Medical Center and Harvard Medical School, Boston, Massachusetts, USA

¹³Department of Brain Repair and Rehabilitation, UCL Queen Square Institute of Neurology, University College London, London, UK

ACKNOWLEDGMENTS

The authors would like to thank the expert panel members (in alphabetical order): Michael Chappell (University of Nottingham), Jia Guo (University of California Riverside), Luis Hernandez-Garcia (University of Michigan), Thomas Okell (University of Oxford), Qin Qin (Johns Hopkins University), Eric Wong (University of California San Diego), Joseph Woods (University of Oxford), for their feedback—and the following representatives of MRI manufacturers in alphabetical order): Valentin Prevost (Canon Medical Systems Corporation), Takashi Tsuneki (FUJIFILM Healthcare), David Shin, Marc Lebel and Suchandrima Banerjee (GE Healthcare), Kim van de Ven (Philips Healthcare), Josef Pfeuffer and Marta Vidorreta (Siemens Healthcare), Jun Xie (United Imaging Healthcare). The authors also thank the ISMRM Perfusion SG members, especially those who responded to the online survey and poll, as well as those who provided feedback during the public presentation. We also thank the ISMRM Publication Committee for reviewing the work and

coordinating the ISMRM Board of Trustees approval process.

CONFLICT OF INTEREST STATEMENT

One of our co-author, Manuel Taso, has started to work for Siemens Medical Solutions USA, after the submission of our manuscript in March.

DATA AVAILABILITY STATEMENT

The online full version of the ASL Lexicon and Reporting Recommendation is available on the OSIPi website (<https://osipi.ismrm.org/task-forces/tf4-1/>). LaTeX equations used in this manuscript are provided together with their source code on the OSIPi website as well. Material can be freely reused in publications and educational material.

FUNDING INFORMATION

Y.S. is supported by the Royal Academy of Engineering Research Fellowship scheme (RF/201920/19/236) and core funding from the Wellcome Trust (203139/Z/16/Z). D.L.T. is supported by the University College London Hospitals (UCLH) National Institute for Health and Care Research (NIHR) Biomedical Research Centre and the Wellcome Trust (Centre award 539208). S.D. is supported by National Institutes of Health (NIH) (grants R03 AG063213 and R01 NS111115). M.F.S. is supported by the Spanish Ministry of Science, Innovation and Universities (grants PI18/00084 and PI21/00578). T.L. is supported by the German Research Foundation (DFG) (grant LI-3030/2-1). H.J.M.M.M. is supported by the Dutch Heart Foundation (03-004-2020-T049); the Eurostars-2 joint programme with co-funding from the European Union Horizon 2020 research and innovation programme (grant ASPIRE E!113701), provided by the Netherlands Enterprise Agency (RVO); and the EU Joint Program for Neurodegenerative Disease Research, provided by the Netherlands Organization for Health Research and Development and Alzheimer Nederland (DEBBIE) (JPND2020-568-106).

ORCID

Yuriko Suzuki  <https://orcid.org/0000-0002-4851-7872>

Thomas Lindner  <https://orcid.org/0000-0002-4408-2429>

Henk J. M. M. Mutsaerts  <https://orcid.org/0000-0003-0894-0307>

Jan Petr  <https://orcid.org/0000-0002-3201-6002>

Xingfeng Shao  <https://orcid.org/0000-0002-4130-6204>

Manuel Taso  <https://orcid.org/0000-0003-2405-9165>

David L. Thomas  <https://orcid.org/0000-0003-1491-1641>

REFERENCES

1. Alsop DC, Detre JA, Golay X, et al. Recommended implementation of arterial spin-labeled perfusion MRI for clinical applications: a consensus of the ISMRM perfusion study group and the European consortium for ASL in dementia. *Magn Reson Med.* 2015;73:102-116.
2. Lindner T, Bolar DS, Achten E, et al. Current state and guidance on arterial spin labeling perfusion MRI in clinical neuroimaging. *Magn Reson Med.* 2023;89:2024-2047.
3. Harms MP, Somerville LH, Ances BM, et al. Extending the human connectome project across ages: imaging protocols for the lifespan development and aging projects. *Neuroimage.* 2018;183:972-984.
4. Clement P, Castellaro M, Okell TW, et al. ASL-BIDS, the brain imaging data structure extension for arterial spin labeling. *Sci Data.* 2022;9:1-8.
5. Nery F, Buchanan CE, Hartevelde AA, et al. Consensus-based technical recommendations for clinical translation of renal ASL MRI. *MAGMA.* 2020;33:141-161.
6. Qin Q, Alsop DC, Bolar DS, et al. Velocity-selective arterial spin labeling perfusion MRI: a review of the state of the art and recommendations for clinical implementation. *Magn Reson Med.* 2022;88:1528-1547.
7. Hernandez-Garcia L, Aramendía-Vidaurreta V, Bolar DS, et al. Recent technical developments in ASL: a review of the state of the art. *Magn Reson Med.* 2022;88:2021-2042.
8. Taso M, Aramendía-Vidaurreta V, Englund EK, et al. Update on state-of-the-art for arterial spin labeling (ASL) human perfusion imaging outside of the brain. *Magn Reson Med.* 2023;89:1754-1776.
9. Woods JG, Achten E, Asllani I, et al. Recommendations for Quantitative Cerebral Perfusion MRI using Multi-Timepoint Arterial Spin Labeling: Acquisition, Quantification, and Clinical Applications. 2023. doi:10.31219/osf.io/4tskr
10. Ogg RJ, Kingsley PB, Taylor JS. WET, a T1- and B1-insensitive water-suppression method for in vivo localized 1H NMR spectroscopy. *J Magn Reson B.* 1994;104:1-10.
11. Golay X, Petersen ET, Hui F. Pulsed star labeling of arterial regions (PULSAR): a robust regional perfusion technique for high field imaging. *Magn Reson Med.* 2005;53:15-21.
12. Petersen ET, Lim T, Golay X. Model-free arterial spin labeling quantification approach for perfusion MRI. *Magn Reson Med.* 2006;55:219-232.
13. Wang J, Alsop DC, Song HK, et al. Arterial transit time imaging with flow encoding arterial spin tagging (FEAST). *Magn Reson Med.* 2003;50:599-607.
14. Ye FQ, Mattay VS, Jezzard P, Frank JA, Weinberger DR, McLaughlin AC. Correction for vascular artifacts in cerebral blood flow values measured by using arterial spin tagging techniques. *Magn Reson Med.* 1997;37:226-235.
15. Williams DS, Detre JA, Leigh JS, Koretsky AP. Magnetic resonance imaging of perfusion using spin inversion of arterial water. *Proc Natl Acad Sci U S A.* 1992;89:212-216.
16. Alsop DC, Detre JA. Reduced transit-time sensitivity in noninvasive magnetic resonance imaging of human cerebral blood flow. *J Cereb Blood Flow Metab.* 1996;16:1236-1249.
17. Alsop DC, Detre JA. Multisection cerebral blood flow MR imaging with continuous arterial spin labeling. *Radiology.* 1998;208:410-416.
18. Dai W, Garcia D, De Bazelaire C, Alsop DC. Continuous flow-driven inversion for arterial spin labeling using pulsed radio frequency and gradient fields. *Magn Reson Med.* 2008;60:1488-1497.
19. Dixon WT, Du LN, Faul DD, Gado M, Rosnick S. Projection angiograms of blood labeled by adiabatic fast passage. *Magn Reson Med.* 1986;3:454-462.
20. Zhao L, Vidorreta M, Soman S, Detre JA, Alsop DC. Improving the robustness of pseudo-continuous arterial spin labeling to off-resonance and pulsatile flow velocity. *Magn Reson Med.* 2017;78:1342-1351.
21. Talagala SL, Ye FQ, Ledden PJ, Chesnick S. Whole-brain 3D perfusion MRI at 3.0 T using CASL with a separate labeling coil. *Magn Reson Med.* 2004;52:131-140. doi:10.1002/mrm.20124
22. Mora Álvarez, MG, Stobbe, RW, Beaulieu, C. High resolution continuous arterial spin labeling of human cerebral perfusion using a separate neck tagging RF coil. *PLOS ONE.* 2019;14(4):e0215998. doi:10.1371/journal.pone.0215998
23. Kwong KK, Belliveau JW, Chesler DA, et al. Dynamic magnetic resonance imaging of human brain activity during primary sensory stimulation. *Proc Natl Acad Sci U S A.* 1992;89:5675-5679.
24. Kim S-G. Quantification of relative cerebral blood flow change by flow-sensitive alternating inversion recovery (FAIR) technique: application to functional mapping. *Magn Reson Med.* 1995;34:293-301.
25. Edelman RR, Siewert B, Darby DG, et al. Qualitative mapping of cerebral blood flow and functional localization with echo-planar MR imaging and signal targeting with alternating radio frequency. *Radiology.* 1994;192:513-520.
26. Edelman RR, Chen Q. EPISTAR MRI: multislice mapping of cerebral blood flow. *Magn Reson Med.* 1998;40:800-805.
27. Wong EC, Buxton RB, Frank LR. Quantitative imaging of perfusion using a single subtraction (QUIPSS and QUIPSS II). *Magn Reson Med.* 1998;39:702-708.
28. Luh WM, Wong EC, Bandettini PA, Hyde JS. QUIPSS II with thin-slice TI1 periodic saturation: a method for improving accuracy of quantitative perfusion imaging using pulsed arterial spin labeling. *Magn Reson Med.* 1999;41:1246-1254.
29. Petersen ET, Golay X. Improved inversion efficiency in arterial spin labeling using adiabatic null pulses. In Proceedings of the 15th Annual Meeting of ISMRM, Berlin, Germany. 2007. Poster 1408
30. Wong EC, Buxton RB, Frank LR. Implementation of quantitative perfusion imaging techniques for functional brain mapping using pulsed arterial spin labeling. *NMR Biomed.* 1997;10:237-249.
31. Jahng GH, Zhu XP, Matson GB, Weiner MW, Schuff N. Improved perfusion-weighted MRI by a novel double inversion with proximal labeling of both tagged and control acquisitions. *Magn Reson Med.* 2003;49:307-314.
32. Golay X, Stuber M, Pruessmann KP, Meier D, Boesiger P. Transfer insensitive labeling technique (TILT): application to multislice functional perfusion imaging. *J Magn Reson Imaging.* 1999;9:454-461.
33. Song R, Loeffler RB, Hillenbrand CM. QUIPSS II with window-sliding saturation sequence (Q2WISE). *Magn Reson Med.* 2012;67:1127-1132.
34. Guo J, Buxton RB, Wong EC. Wedge-shaped slice-selective adiabatic inversion pulse for controlling temporal width of bolus in pulsed arterial spin labeling. *Magn Reson Med.* 2016;76:838-847.

35. Ye FQ, Frank JA, Weinberger DR, McLaughlin AC. Noise reduction in 3D perfusion imaging by attenuating the static signal in arterial spin tagging (ASSIST). *Magn Reson Med.* 2000;44:92-100.
36. Wong EC, Cronin M, Wu WC, Inglis B, Frank LR, Liu TT. Velocity-selective arterial spin labeling. *Magn Reson Med.* 2006;55:1334-1341.
37. Qin Q, van Zijl PCM. Velocity-selective-inversion prepared arterial spin labeling. *Magn Reson Med.* 2016;76:1136-1148.
38. Guo J, Meakin JA, Jezard P, Wong EC. An optimized design to reduce eddy current sensitivity in velocity-selective arterial spin labeling using symmetric BIR-8 pulses. *Magn Reson Med.* 2015;73:1085-1094.
39. Qin Q, Shin T, Schär M, Guo H, Chen H, Qiao Y. Velocity-selective magnetization-prepared non-contrast-enhanced cerebral MR angiography at 3 Tesla: improved immunity to B0/B1 inhomogeneity. *Magn Reson Med.* 2016;75:1232-1241.
40. Shin T, Hu BS, Nishimura DG. Off-resonance-robust velocity-selective magnetization preparation for non-contrast-enhanced peripheral MR angiography. *Magn Reson Med.* 2013;70:1229-1240.
41. Guo J, Wong EC. Increased SNR efficiency in velocity selective arterial spin labeling using multiple velocity selective saturation modules (mm-VSASL). *Magn Reson Med.* 2015;74:694-705.
42. Schmid S, Ghariq E, Teeuwisse WM, Webb A, Van Osch MJP. Acceleration-selective arterial spin labeling. *Magn Reson Med.* 2014;71:191-199.
43. Buxton RB, Frank LR, Wong EC, Siewert B, Warach S, Edelman RR. A general kinetic model for quantitative perfusion imaging with arterial spin labeling. *Magn Reson Med.* 1998;40:383-396.
44. Gonzalez-At JB, Alsop DC, Detre JA. Cerebral perfusion and arterial transit time changes during task activation determined with continuous arterial spin labeling. *Magn Reson Med.* 2000;43:739-746.
45. Mildner T, Möller HE, Driesel W, Norris DG, Trampel R. Continuous arterial spin labeling at the human common carotid artery: the influence of transit times. *NMR Biomed.* 2005;18:19-23.
46. Yang Y, Engelen W, Xu S, Gu H, Silbersweig DA, Stern E. Transit time, trailing time, and cerebral blood flow during brain activation: measurement using multislice, pulsed spin-labeling perfusion imaging. *Magn Reson Med.* 2000;44:680-685.
47. Jung Y, Wong EC, Liu TT. Multiphase pseudocontinuous arterial spin labeling (MP-PCASL) for robust quantification of cerebral blood flow. *Magn Reson Med.* 2010;64:799-810.
48. Thomas DL, Lythgoe MF, Van Der Weerd L, Ordidge RJ, Gadian DG. Regional variation of cerebral blood flow and arterial transit time in the normal and hypoperfused rat brain measured using continuous arterial spin labeling MRI. *J Cereb Blood Flow Metab.* 2006;26:274-282.
49. Gunther M, Bock M, Schad LR. Arterial spin labeling in combination with a look-locker sampling strategy: inflow turbo-sampling EPI-FAIR (ITS-FAIR). *Magn Reson Med.* 2001;46:974-984.
50. Dai W, Robson PM, Shankaranarayanan A, Alsop DC. Reduced resolution transit delay prescan for quantitative continuous arterial spin labeling perfusion imaging. *Magn Reson Med.* 2012;67:1252-1265.
51. Dai W, Shankaranarayanan A, Alsop DC. Volumetric measurement of perfusion and arterial transit delay using hadamard encoded continuous arterial spin labeling. *Magn Reson Med.* 2013;69:1014-1022.
52. von Samson-Himmelstjerna F, Madai VI, Sobesky J, Guenther M. Walsh-ordered hadamard time-encoded pseudocontinuous ASL (WH pCASL). *Magn Reson Med.* 2016;76:1814-1824.
53. Thomas DL, Lythgoe MF, Pell GS, Calamante F, Ordidge RJ. The measurement of diffusion and perfusion in biological systems using magnetic resonance imaging. *Phys Med Biol.* 2000;45:R97-R138.
54. Asllani I, Borogovac A, Brown TR. Regression algorithm correcting for partial volume effects in arterial spin labeling MRI. *Magn Reson Med.* 2008;60:1362-1371.
55. Chappell MA, Groves AR, MacIntosh BJ, Donahue MJ, Jezard P, Woolrich MW. Partial volume correction of multiple inversion time arterial spin labeling MRI data. *Magn Reson Med.* 2011;65:1173-1183.

SUPPORTING INFORMATION

Additional supporting information may be found in the online version of the article at the publisher's website.

Figure S1. The demographic information of the survey responders.

Figure S2. The responses to the questions "Is 'Required' an appropriate category for the parameter xxx?" The numbers of votes for each response option with corresponding color scheme are shown next to the plots.

Figure S3. The responses to the questions "Is 'Recommended' an appropriate category for the parameter xxx?" The numbers of votes for each response option with corresponding color scheme are shown next to the plots.

Figure S4. Parameters in 'Required' category.

Figure S5. Parameters in 'Recommended' category.

How to cite this article: Suzuki Y, Clement P, Dai W, et al. ASL lexicon and reporting recommendations: A consensus report from the ISMRM Open Science Initiative for Perfusion Imaging (OSIPI). *Magn Reson Med.* 2024;91:1743-1760. doi: 10.1002/mrm.29815