

Reducing Slab Boundary Artifacts in Three-Dimensional Multislab Diffusion MRI Using Nonlinear Inversion for Slab Profile Encoding (NPEN)

Wenchuan Wu,* Peter J. Koopmans, Robert Frost, and Karla L. Miller

Purpose: To propose a method to reduce the slab boundary artifacts in three-dimensional multislab diffusion MRI.

Methods: Bloch simulation is used to investigate the effects of multiple factors on slab boundary artifacts, including characterization of residual errors on diffusion quantification. A nonlinear inversion method is proposed to simultaneously estimate the slab profile and the underlying (corrected) image.

Results: Correction results of numerical phantom and in vivo data demonstrate that the method can effectively remove slab boundary artifacts for diffusion data. Notably, the nonlinear inversion is also successful at short TR, a regimen where previously proposed methods (slab profile encoding and weighted average) retain residual artifacts in both diffusion-weighted images and diffusion metrics (mean diffusion coefficient and fractional anisotropy).

Conclusion: The nonlinear inversion for removing slab boundary artifacts provides improvements over existing methods, particularly at the short TRs required to maximize SNR efficiency. **Magn Reson Med 000:000–000, 2015. © 2015 The Authors. Magnetic Resonance in Medicine published by Wiley Periodicals, Inc. on behalf of International Society for Magnetic Resonance in Medicine. This is an open access article under the terms of the Creative Commons Attribution License, which permits use, distribution and reproduction in any medium, provided the original work is properly cited.**

Key words: multi-slab; boundary artifacts; 3D diffusion MRI; nonlinear inversion

INTRODUCTION

Improvements in spatial resolution for diffusion MRI (dMRI) of the brain can provide the ability to resolve small structures, enabling investigation of finer neuroanatomical features and detailed white matter fiber architecture. Although two-dimensional (2D) acquisition can

increase the in-plane resolution significantly (1–8), this approach faces limitations when small isotropic voxels are required. Most importantly, 2D methods couple the repetition time (TR) to the number of slices, such that whole brain coverage necessitates long TR (typically 6–10 s). In the brain, signal-to-noise ratio (SNR) efficiency of spin echo sequences is maximized when $TR = 1\text{--}2$ s, meaning that 2D acquisitions are far from optimal for even modest resolution (9). This effect becomes acute at high resolutions due to the competing needs for a large number of slices to cover the brain (requiring longer TR) and high SNR efficiency to support reduced voxel volume (favoring shorter TR). Although simultaneous multislice methods mitigate these issues to some degree, the coupling between the number of slices and the TR will remain an issue, particularly at high resolutions.

Three-dimensional (3D) acquisitions remove this coupling by exciting the entire volume and defining voxels based on 3D k-space, acquiring the entire image over multiple readouts (multiple shots). Nevertheless, the use of 3D acquisitions for dMRI faces several limitations. Most importantly, in order to produce image volumes reasonably quickly, 3D dMRI requires very short TR (~100 ms), which is also very SNR-inefficient (9). Another challenge comes from the motion-induced phase errors, which can degrade image quality in multishot acquisitions if not addressed properly. Applying a full 3D nonlinear phase correction method (4) requires time-consuming 3D phase navigators, which would significantly diminish the scan efficiency. Acquisition of 3D phase navigators could be avoided by using approaches such as driven equilibrium diffusion preparation (10), acquisition with a small field of view (FOV) (11–13), cardiac gating (11,14,15), and estimation of 3D phase errors using data from multiple shots (16,17). However, these methods still retain residual artifacts, especially when diffusion encoding is along the superior–inferior direction, the direction in which brain pulsation is most significant (18–20).

To overcome the limitations faced by 2D and 3D acquisition, 3D multislab acquisition has been proposed for high-resolution dMRI (9,15,21). In 3D multislab acquisition, the whole volume is divided into multiple smaller slabs, each of which is excited and encoded separately (i.e., each slab is in effect a reduced-FOV volume along the slice direction). As a 3D imaging method, it can easily achieve thin slices by increasing phase encoding lines in the slice direction. With proper choice of slab thickness, motion-induced 3D phase errors in each slab can be well approximated by a 2D navigator (21,22). This hybrid method partly removes the coupling between number of slices and TR: although TR is

FMRIB Centre, Nuffield Department of Clinical Neurosciences, University of Oxford, Oxford, United Kingdom

Grant sponsor: Marie Curie Initial Training Network program and the Wellcome Trust.

*Correspondence to: Wenchuan Wu, M.Sc., FMRIB Centre, Nuffield Department of Clinical Neurosciences, John Radcliffe Hospital, University of Oxford, Headington, Oxford OX3 9DU, UK. E-mail address: wenchuan.wu@ndcn.ox.ac.uk

Received 17 July 2015; revised 5 October 2015; accepted 5 October 2015

DOI 10.1002/mrm.26027

Published online 00 Month 2015 in Wiley Online Library (wileyonlinelibrary.com).

© 2015 The Authors. Magnetic Resonance in Medicine published by Wiley Periodicals, Inc. on behalf of International Society for Magnetic Resonance in Medicine. This is an open access article under the terms of the Creative Commons Attribution License, which permits use, distribution and reproduction in any medium, provided the original work is properly cited.




FIG. 1. (a) Illustration of slab aliasing. The multislab acquisition shown here excites three slabs with 10 k_z phase-encodes for each slab. Extension of the slab profile is wider than the phase encoded FOV (also the predefined slab thickness), leading to aliasing artifacts. (b) Illustration of slab crosstalk effect. The slab profile extends to its adjacent slabs, exacerbating T1 saturation effects at slab boundaries where adjacent slabs overlap. In this example with three-slab acquisition, magnetization at the slab center (region I) is excited every TR, whereas magnetization at slab boundary (region II) is excited with intervals of $2/3TR$, $1/3TR$, $2/3TR$, $1/3TR$..., leading to decreased signal at slab boundaries and asymmetry in the shapes. Slab profiles were generated using Bloch simulation with $TR=20$ s (a) and 2 s (b), respectively, and interleaved acquisition order. Details of other simulation parameters are described in the Methods section.

proportional to the number of slabs, the slice definition within each slab is accomplished using 3D k-space encoding. The number of slabs required to cover a given volume is much less than the number of slices and is primarily dictated by the thickest slab that can be phase-corrected with a 2D navigator. In practice, the whole brain can be covered in 12–14 slabs, for which the TR is highly compatible with maximizing SNR efficiency.

A remaining challenge for 3D multislab acquisition is the slab boundary artifact (23), which stems from the shape of the RF profile and appears as periodic signal modulation in the slice direction. Ideally, the RF profile of a slab would have rectangular shape and the width of the slab thickness, but this requires RF pulses to be infinitely long. In reality, RF pulses are truncated, resulting in a nonrectangular RF profile, including magnitude variation inside the slab, transition bands, and side lobes. Therefore, the slab edges have lower signal magnitude than the slab center. The transition bands and side lobes can extend to the neighboring slabs, introducing slab crosstalk and aliasing (Fig. 1).

Several methods have been proposed to reduce the slab boundary artifacts. The most widely used methods include oversampling in the slice direction, deliberate overlapping of the adjacent slabs (23–25), separating acquisitions for odd/even slabs (15), and applying a (weighted) slab combination

at the slab boundary region (9,15,21,23–26). These methods work well in dMRI if long TRs (~ 4 s) are used together with large amounts of slab overlap, but they exhibit residual artifacts in shorter TR acquisitions (27). A variation of the overlapping method shifts the slab by one slice position once all k_z phase encoding steps for a designated subset of k_y phase encoding steps are acquired to transform the signal modulation in the slice direction to the k_y phase encoding direction, using signal demodulation to reduce ghosting artifacts (28,29). This method has also been demonstrated in non-Cartesian sequences (30–32). However, if the signal modulation is not corrected properly, this method may retain ghosting artifacts. Another drawback to these methods is the reduction of scan efficiency due to oversampling, overlapping, or discarding some slices. A histogram-matching method was proposed to correct the slab boundary artifacts without increasing scan time, which assumes the slab boundary artifact is a global intensity variation in the slice direction (33). However, when this assumption is not valid, residual artifacts remain after correction. Recently, a slab profile encoding (PEN) method was proposed to minimize slab boundary artifacts with a minor increase in scan time (34). PEN treats multislab acquisition as a linear encoding problem, which can be solved by linear inversion approaches (e.g., least-squares/pseudo-inverse) as used in SENSE (35), with the slab profiles effectively substituting coil sensitivity profiles. With slab profiles estimated from calibration data and long TRs (~ 4 s), PEN can effectively correct aliasing, but some residual artifacts remain due to slab crosstalk, especially at shorter TRs required to maximize SNR efficiency.

We present a method to minimize the slab boundary artifacts for 3D multislab dMRI with SNR-optimal TRs (~ 2 s). We first present detailed simulations demonstrating that the slab boundary artifact is a combination of slab crosstalk and aliasing. Whereas aliasing has been explained and addressed in PEN (34), slab crosstalk effects have not been studied in detail. We characterize the effect of multiple factors (acquisition scheme, TR, T1, B0 inhomogeneity) on slab crosstalk and how it may alter the measurement of diffusion metrics, despite signal normalization by $b=0$ images as part of quantification. We then introduce a nonlinear inversion extension of PEN, which we call NPEN, where the slab profile encoding is formulated as a nonlinear optimization. Two constraints are added, which enforce in-plane smoothness on the slab profile and suppress frequency components corresponding to interslab distance. Results from simulation and in vivo data demonstrate the proposed method can effectively reduce the slab boundary artifacts for 3D multislab dMRI with a short TR of 2 s, a regimen in which weighted average (WA) and PEN reconstructions contain significant residual slab boundary artifacts.

METHODS

Bloch Simulations

This section is focused on simulation of slab crosstalk effect, including its causes and how it affects diffusion measurements. These results are crucial motivation for the correction method presented in the following section, where aliasing effects are also discussed.




FIG. 2. Results of Bloch simulation. (a–d) Comparison of signal profiles of WM generated from Bloch simulation using different TRs, acquisition schemes, and number of slabs. (a) 11 slabs, interleaved acquisition with TR=2 s (blue) and TR=4 s (red). (b) TR=2 s, 11 slabs with interleaved acquisition (blue) and sequential acquisition (red). (c) TR=2 s, interleaved acquisition with 11 slabs; the signal profiles of slab 5 (blue) and slab 6 (red) are overlaid for better comparison. (d) TR=2 s, interleaved acquisition with 12 slabs; the signal profiles of slab 5 (blue) and slab 6 (red) are overlaid. The effects of TR, acquisition scheme, and number of slabs on GM and CSF are similar to WM and are not shown here. (e) Slab profiles for WM, GM, and CSF, which were generated by normalizing the MR signal profile with the signal magnitude of the center slice. (f) Centered signal profiles simulated with 0 Hz and 100 Hz off-resonance frequencies.

Slab Crosstalk in Multislab Imaging

Figure 1b depicts a Bloch simulation of a multislab acquisition, including both RF profile and T1 saturation effects. The central part of each slab is not overlapped with adjacent slabs, so the magnetization in this region experiences an RF pulse every TR. At slab boundaries, the magnetization is excited every $\sim TR/2$, but with both the timing and flip angle varying for the target and adjacent slab excitations, which exacerbates T1 saturation and results in lower signal. This effect is called slab crosstalk.

The precise nature of slab crosstalk is affected by a range of factors, including RF profiles, excitation order, TR, T1, and B0 inhomogeneity. To obtain a quantitative understanding about how these factors affect the slab crosstalk, we performed Bloch simulations of the spin echo sequence used in this work (7). The simulations used 600 isochromats in each voxel, which are summed to estimate the MR signal. We simulated a range of TRs and acquisition schemes (sequential and interleaved slab ordering), the three major types of tissue in human brain (white matter [WM], gray matter [GM], and cerebrospinal fluid [CSF]), and a range of off-resonance frequencies. Simulations were matched to the pulse sequence design used for the in vivo experiments described in later sec-

tions. The sequence used 90° and 180° RF pulses designed with the SLR algorithm (36) with a time-bandwidth product of 20 and 8, respectively, which were chosen to achieve a sharp composite RF profile. The RF pulse waveforms are plotted in Supporting Figure S1. The duration for excitation pulse and refocusing pulse were 7.18 ms and 10.24 ms. The echo times for the first and second echo (phase navigator) were 78 ms and 122 ms. T1 and T2 values for 3T were assumed: WM, 840/70 ms; GM, 1320/110 ms; and CSF, 3000/2000 ms (37–39). In all simulations, a 15-mm slab thickness was chosen, because this thickness allows whole brain coverage with TR=2 s while the slabs remain thin enough to be corrected with 2D navigators (21). Each slab contained 10 slices, 20% of which were overlapped with slices from adjacent slabs to favor PEN and NPEN reconstruction (i.e., two slices on each side).

The effects of TR, acquisition schemes, and B0 inhomogeneity on slab crosstalk are demonstrated in Figure 2. Figure 2a compares the simulated signal profile with different TR, showing reduced saturation effect at slab boundaries with longer TR, although in practice this comes at the significant cost of SNR efficiency. Figure 2b compares acquisitions with different slab ordering schemes. Sequential acquisition suffers from substantial signal loss at one side of the slab, while interleaved




FIG. 3. Simulation of FA errors caused by slab crosstalk. (a, b) Simulations based on a WM/CSF partial volume model: (a) the absolute FA errors (differences of FA values between slab boundary and slab center) as a function of WM percentage and b value and (b) the absolute FA values at slab center. (c, d) Simulations from a WM/GM partial volume model: (c) the absolute FA errors as a function of WM percentage and b value and (d) the absolute FA values at slab center.

acquisition shows high signal intensity at both sides of the slab. Interleaved acquisition with an odd number of slabs produces almost identical signal profile for the odd and even slabs (Fig. 2c), whereas in the acquisition with an even number of slabs, the signal profile for the odd slab varies from the even slab (Fig. 2d). In the following simulations and in vivo scans, interleaved acquisition with an odd number of slabs was used to achieve homogeneous slab profiles with a well-defined periodicity, which facilitates NPEN reconstruction (described below). Figure 2e compares the slab profiles for WM, GM, and CSF, each normalized by the signal magnitude of the center slice to emphasize differences in shape due to T1 saturation effects.

To simulate the effects of B0 inhomogeneity, Bloch simulation was performed with off-resonance frequencies starting from -250 Hz to 250 Hz in 5-Hz increments. The signal profile of one slab in WM is shown in Supporting Figure S2. The profile not only shifts but also becomes distorted, resulting in more signal loss at the slab boundaries (Fig. 2f). The distortion is caused by the difference in bandwidths of the excitation and refocusing

pulses. It should be noted that a constant off-resonance frequency for each slab is assumed in the B0 inhomogeneity simulation, but in reality, the off-resonance frequencies are spatially varying, which can distort the signal profile even with matched bandwidths of the excitation and refocusing pulses.

Propagation of Slab Crosstalk into Diffusion Quantification

The practical importance of variations in signal intensity at slab boundaries will depend on whether these artifacts alter diffusion quantification, based on signal attenuation relative to a non-diffusion-weighted ($b=0$) reference. Multiplicative slab artifacts that attenuate diffusion-weighted and unweighted signals equally would be removed when normalizing diffusion-weighted images by the $b=0$ reference. In this case, the only effect on diffusion quantification would be SNR reduction at slab boundaries. Conversely, non-multiplicative effects could create a more significant artifact by introducing bias in diffusion metrics at different regions of slab profiles. For

example, mean diffusivity (MD) or fractional anisotropy (FA) could exhibit a consistent bias at slab boundaries. One potential cause of non-multiplicative attenuation is “partial volume” effects, in which both relaxation and diffusion alter the signal to create a more complicated signal dependence than either effect on its own.

To demonstrate non-multiplicative effects in diffusion metrics due to partial volume, we simulated a two-compartment diffusion-weighted signal corresponding to WM/GM and WM/CSF, respectively (40):

$$M(b, v) = M_0 \sum_{i=1}^2 f_i \exp(-bv^T D_i v) \left(\sum_{i=1}^2 f_i = 1 \right). \quad [1]$$

Here, f_i and D_i ($i=1,2$) are the volume fractions and diffusion tensors for the two compartments and v is the diffusion gradient vector. The $b=0$ signal M_0 is calculated using Bloch simulations and the diffusion-weighted signal M is subsequently calculated from the partial volume model. The diffusion tensor eigenvalues $[\lambda_1, \lambda_2, \lambda_3]$ for WM [1.4,0.35,0.35] (FA = 0.707), GM [0.7,0.7,0.7] and CSF [3.0,3.0,3.0] were used (40). A set of six noncollinear directions was assumed for v . A wide range of b value (1 – 10,000 s/mm²) was investigated. We calculated MD and FA for a simulated voxel with a given partial volume, considering these as broadly representative of bias in diffusion metrics (although specific model parameters will vary in their sensitivity to slab boundary artifacts). The goal of these simulations was to identify any consistent bias in diffusion metrics for the slab boundary relative to the slab center. A relative error ε_r was calculated to analyze the slab crosstalk effect on MD ($\varepsilon_r = |\theta_c - \theta_b|/\theta_c$, where θ_c and θ_b are the MD values at slab center and boundary). As the FA values of CSF and GM approach zero when their fractions go to 1, leading to inflated measure of relative errors, we calculated the absolute FA errors between slab center and boundary instead.

Figure 3 presents the results of FA simulation from the partial volume mode. The absolute errors depend on the partial volume ratio of the tissues and the b values used in the simulation. For WM/CSF partial volume (Fig. 3a and 3b), the absolute FA error approaches its maximum value of ~ 0.045 with WM percentage of $\sim 7\%$ and b value smaller than 1000 s/mm². For WM/GM partial volume (Fig. 3c and 3d), the maximal absolute FA error is ~ 0.02 . In both cases, the FA errors decrease when the b value goes up, as more isotropic-diffusion compartments (GM and CSF) are suppressed due to their higher diffusion coefficients. Figure 4 shows the relative errors for MD from WM/CSF partial volume, which are on the order of a few percent with a maximum value of $\sim 8\%$. For WM/GM partial volume, the relative errors for MD are negligible (image not shown). Importantly, these simulations of partial volume effects demonstrate that slab profile effects in voxels with partial volume cannot be trivially removed by normalization during the calculation of diffusion parameters.

Proposed Correction

The simulations presented in the previous sections demonstrate that slab-boundary artifacts depend on the




FIG. 4. Relative MD errors as a function of WM percentage and b value simulated with WM/CSF partial volume.

details of the tissue environment, including relaxation times and field inhomogeneity. Furthermore, these effects cannot be assumed to be removed by normalization with a $b=0$ image in the likely case of tissue partial volume. Finally, these effects are considerably worse in the SNR-optimal regimen of 1–2 s. The accuracy of 3D multislab dMRI thus depends on finding a method for reducing the impact of these artifacts on quantified diffusion values. It is useful to begin by reviewing slab aliasing and the previously proposed PEN method (34). We then present a new method based on PEN that aims to deal with the additional slab crosstalk issues raised by the simulations above.

The PEN method treats the 3D multislab acquisition as a linear encoding process, which can be expressed as

$$PFCSu = d, \quad [2]$$

where u is the unknown 3D image, $S = (s_1, s_2, \dots, s_{N_{slab}})$ describes the slab profiles, C describes the coil sensitivities, F is the Fourier transform operator, P captures the k -space sampling trajectory, and d is the acquired multislab k -space data. This is very similar to the general parallel imaging model, except that here the image is weighted by both coil sensitivity and slab profile. Aliasing artifacts arise as a result of undersampling in the slice direction, where we here consider the effect of magnetization excited by a given pulse throughout the entire brain volume.

The quality of PEN reconstruction is critically dependent on the ability to estimate the slab profiles accurately. As shown in Figure 2, the slab crosstalk is affected by multiple factors, making it difficult to estimate slab profiles accurately using Bloch simulation. An alternative approach is to measure slab profiles from a calibration scan, which is a multislab scan at the same slab locations but oversampled in the slice direction to minimize aliasing. The slab images were first zero-padded to the full FOV, and then a combined volume was generated by sum-of-squares of all the zero-padded slabs. Finally, the slab profile was estimated by normalizing each of the

zero-padded slab images by this combined single-slab image (34). This method will not increase the scan time significantly, but the reconstructed images will be modulated by the sum-of-squares of the slab profiles (34,35), which could be relatively flat in the slice direction with a long TR (~ 4 s) but would result in significant signal variation with a short TR (~ 2 s).

As originally proposed, the PEN method relies on the assumption that slab profile estimates are accurate, which is increasingly difficult to ensure at short TR. To overcome this problem, we propose a new nonlinear reconstruction that jointly estimates the slab profile and the image. In this method, both slab profile and image are treated as unknowns $x = [u, S]^T$, which should satisfy

$$E(x) = (PFCs_1 u, PFCs_2 u, \dots, PFCs_{N_{slab}} u), \quad [3]$$

where E is a nonlinear operator mapping image, slab profile, and coil sensitivity to the acquired data. This poses the 3D multislab reconstruction as a nonlinear inversion problem, which we refer as nonlinear inversion for slab profile encoding (NPEN). As shown previously (41–44), this problem can be solved using an iteratively regularized Gauss–Newton algorithm (45–47). In step n , Equation [3] is first linearized around x_n :

$$E'(x_n)\Delta x_n = d - E(x_n), \quad [4]$$

where $E'(x_n)$ is the Fréchet derivative of E at the current guess x_n , Δx_n is the update, and the new guess can be obtained by $x_{n+1} = x_n + \Delta x_n$. The linear problem described by Equation [4] is then solved by the following Tikhonov regularized minimization:

$$\min \| E'(x_n)\Delta x_n - (d - E(x_n)) \|^2 + \alpha_n \| x_n + \Delta x - x_0 \|^2, \quad [5]$$

where the first term is the data fidelity term and the second term is the regularization term. x_0 is an initial guess and α_n is a regularization parameter, which is reduced in each step to benefit from the robustness at the initial steps of iterations (i.e., gradient–descent-like) and the fast convergence when x is close to the solution (i.e., Gauss–Newton-like). Equation [5] can be solved with the conjugate gradient algorithm.

As reported previously (41,43), more accurate estimation of the solution to Equation [3] can be achieved by adding further constraints in the reconstruction. In this study, we considered two constraints. First, the slab profile was constrained to be smooth (in-plane), which is similar to the polynomial fitting process in the slab profile estimation in PEN. The justification for this constraint is the observation that the slab profiles for WM, GM, and CSF are very similar (Fig. 2e). To realize this constraint, the preconditioning method described by Uecker et al (41) was applied, where $x = [u, S]^T$ is replaced by $x = [u, W_s F S]^T$ and W_s is a matrix used to penalize high spatial frequency k_x - k_y components of S in each k_z plane. The inverse of $W_s F$ is incorporated into the preconditioned E accordingly.

The second constraint exploits the fact that slab boundary artifacts are approximately periodic in image space. Therefore, residual artifacts would correspond to spikes in k_z at known frequencies. By suppressing these spurious k -space signals, images with reduced slab boundary artifacts can be reconstructed. This constraint was transformed as an additional penalty term in the reconstruction

$$\min \| E'(x_n)\Delta x_n - (d - E(x_n)) \|^2 + \alpha_n \| x_n + \Delta x - x_0 \|^2 + \beta_n \| W_u F u_n \|^2, \quad [6]$$

where W_u is a weighting matrix with large values at the spike frequencies and small values at other frequencies such that $\| W_u F u_n \|^2$ penalizes residual slab boundary artifacts. β_n is a regularization parameter controlling the strength of this constraint. For pure frequencies, W_u would be a binary matrix; however, because the slab boundary artifacts are not pure frequencies, W_u is instead formulated as a Gaussian weighting centered at the primary frequencies dictated by slab separation.

Numerical Simulation

To test the reconstruction algorithm under controlled conditions, we generated realistic simulations of spin echo images based on the MNI-Colin27 high-resolution multimodal brain atlas (very high SNR T1- and T2-weighted scans of a single subject) (48). Artificial 3D multislab data were generated based on the following pipeline. Tissue partial volume maps were estimated using FAST (49) based on the T1-weighted images. The tissue-specific slab profiles calculated from Bloch simulations were weighted by the partial volume maps and summed to generate the integrated slab profile. The T2-weighted images were weighted by the integrated slab profile and Fourier-transformed to produce k -space data. Finally, the k -space data were sampled in a way that is consistent with the acquisition of 11 slabs with 10 slices per slab to simulate aliasing effects. These simulated 3D multislab data were then reconstructed using the candidate methods (WA, PEN, and NPEN; details below) and compared with the ground truth (the original T2-weighted atlas).

To compare the performance of PEN and NPEN at different TRs, we applied Bloch simulation at TR of 2 s, 4 s, and 7 s, respectively using SLR pulses designed with the same TBWP as used in the PEN work (34). The calculated slab profiles were used to generate simulation data in the way described above, and then the data was reconstructed using PEN and NPEN.

In Vivo Studies

A diffusion-weighted readout-segmented echo planar imaging (EPI) sequence (6,7,50) was modified to acquire 3D multislab data. This sequence was further modified to oversample the central segment of $b=0$ data by a factor of two. This oversampled $b=0$ data could be used to estimate slab profiles for PEN or calculate initial guesses for NPEN, alleviating the need for a separate set of calibration data.

Data were acquired from four healthy subjects with informed consent in accordance with local ethics. For each subject, nine slabs with 10 slices per slab were

acquired, and adjacent slabs were overlapped by two slices (20% overlapping), which resulted in 74 slices in the final reconstruction. k-Space coverage of each slab was achieved using a 3D adaptation of a readout-segmented EPI approach (9). The matrix size in each k_z plane was 146×146 , covered across five readout segments using 3/5 partial Fourier acquisition (i.e., only three segments were acquired) (51). Isotropic spatial resolution $1.5 \times 1.5 \times 1.5 \text{ mm}^3$ was acquired with a FOV of $220 \times 220 \times 111 \text{ mm}^3$. The same echo time (78 ms, 122 ms) and RF pulses were used as in the Bloch simulation, and TR was 2 s. The scan time for one diffusion direction was 1.3 min with central segment oversampling (the first $b=0$ scan only) and 1 min without central segment oversampling (all other volumes).

Diffusion-weighted images were acquired with $b = 1000 \text{ s/mm}^2$ using a Stejskal–Tanner diffusion preparation and 20 isotropically distributed diffusion encoding directions, and two $b=0$ images. Only the first $b=0$ data were oversampled for estimating the slab profile, which was then applied to the reconstruction of all other data (such that calibration data represents a constant overhead equivalent to a single volume scan). The total acquisition time was 22.3 min.

Reconstruction

The reconstruction of readout-segmented EPI data followed procedures described previously (6,9), except that the oversampled central segments were reconstructed separately. Two sets of images were obtained from this reconstruction: high-resolution 3D multislab images and low-resolution 3D multislab images with doubled FOV in the slice direction, which were used as calibration images. Three slab-boundary artifact correction methods—WA, PEN, and NPEN—were compared on simulation data and in vivo data.

In NPEN reconstruction, the initial guess for the image was set to zero. The calibration images were used to generate the initial guess for slab profiles as follows. These images contained no significant aliasing artifacts due to oversampling in the slice direction, such that the variation of signal intensity along the slice direction was dominated by slab profile, including side lobes that led to saturation effects. First, each slab image was zero-padded to the full FOV, resulting in image I . The averaged signal for each slice was calculated as

$$I_{ave}(z) = \sum_r |I(r, z)M(r, z)| / \sum_r |M(r, z)|, \quad [7]$$

where M is a brain mask generated by FSL’s Brain Extraction Tool (52), r indicates in-plane locations, and z indicates the locations in the slice direction. The averaged signal was then normalized by the signal magnitude at the slab center

$$I'_{ave}(z) = I_{ave}(z) / I_{ave}(z_c), \quad [8]$$

where z_c denotes the location of the center slice in the slab. Finally, the 3D slab profile estimate S_{est} was generated by repeating I'_{ave} for all the in-plane locations of S_{est} .

The initial regularization parameters α_0 and β_0 were set to ensure that after the first iteration, the residual $\|E(x) - d\|$ was approximately 3/4 of the initial residual (43). This resulted in $\alpha_0 = 0.2$ and $\beta_0 = 0.4$ for both simulated and in vivo data. The decay rates of α and β were set to 1.5, which performed well in the preliminary testing. To achieve better suppression of slab boundary artifacts, the decrease of β was stopped at 0.03.

PEN reconstruction was implemented as described by Van et al (34). The slab profile was estimated from the calibration images, in which the single slab image was divided by the sum-of-squares of all slab images. A 2D low-pass Hamming filter was applied in k-space to reduce noise in the estimated slab profiles. In WA reconstruction, a one-dimensional (1D) Fermi filter function was applied to each slab by multiplication in the spatial domain to reduce the signal from the edge slices with aliasing, and then the weighted slab images were averaged to generate a 3D single slab image (21,34).

Performance of different methods on the simulation data was evaluated using root-mean-square-error (RMSE) and difference maps from the ground truth image. To assess the reconstruction performance on different tissue types, RMSE and difference maps were also calculated for WM, GM, and CSF separately using binary segmentation maps generated from partial volume maps with a threshold of 0.8.

For the in vivo data, eddy current correction was performed in FSL using the first $b=0$ volume as a reference, after which DTI fitting was applied to generate diffusion parameter maps using FSL’s diffusion toolbox (53).

RESULTS

Reconstruction with Simulation Data

Ideally, we would like to use the minimal RMSE as the stopping criteria of the iteration. However, for in vivo scans, there is no ground truth image for RMSE calculation, so we need to find another stopping criteria. The RMSEs and image update ($\|\Delta u\|$) from the atlas-based simulation reconstructions are shown in Figure 5. After the first three iterations, $\|\Delta u\|$ decreases toward its minimum and increases afterward (Fig. 5a). The iteration number corresponding to the minimum $\|\Delta u\|$ also approximately corresponds to the lowest RMSE for the full image (Fig. 5b). This suggests that a stopping criterion based on $\|\Delta u\|$ would be appropriate in practice (when the ground truth image is not available to calculate RMSE). However, as shown in Figure 5b, the iteration numbers corresponding to the minimal RMSE of the full image, WM, and GM are different, which probably stems from the T1 dependence of the slab profile. To achieve the best reconstruction of WM, the iteration should be stopped before the minimal RMSE of the full image, or the minimal $\|\Delta u\|$ in practice. Therefore, in this study, the iteration was stopped with the smallest image update $\|\Delta u\|$, and the best reconstruction for WM was retrospectively chosen based on visual inspection. Although the final reconstruction is determined by visual inspection, the stopping criteria based on $\|\Delta u\|$ can limit the number of iterations and reduce the computation time.




FIG. 5. (a) Image update at each iteration. (b) RMSE for the full image (black), WM (red), and GM (blue) calculated between the reconstructed image and the ground truth at each iteration. Solid lines show the RMSE of NPEN reconstruction; dashed lines indicate the RMSE of PEN reconstruction.

Figure 6 shows the correction results on simulated data. WA and PEN results still retained strong artifacts at slab boundaries, whereas the NPEN result shows significantly reduced artifact level, revealed by smaller RMSE and minor residual artifacts in the difference map. It is




FIG. 6. Reconstruction results from simulated data. Upper row: ground truth image and images reconstructed by WA, PEN, and NPEN. Lower row: difference maps between the reconstructed images and the ground truth; RMSE is shown in the upper right-hand corner. The difference map is rescaled for better depiction of the error distributions.

worth noting that, according to the literature, PEN has a much lower artifact level (34), which is due to the use of longer TR corresponding to less T1 saturation but which results in lower SNR efficiency and longer scan time.

Figure 7 shows the tissue-specific comparison between three reconstruction methods. As shown in Figure 7a, the artifacts in WM have been mostly removed by NPEN, which results in an RMSE of 0.03, whereas WA and PEN results retain high artifact levels, with an RMSE of 0.1 and 0.09, respectively. The GM results (Fig. 7b) also suggest that NPEN performs better than WA and PEN. Figure 7c shows the CSF results, where none of the three methods provided a satisfactory correction, but NPEN still yielded the lowest RMSE. The tissue-specific comparison shows that most residual artifacts in NPEN are located in CSF region, which is not problematic for dMRI.

For a clearer comparison, the WM signal was averaged in each slice and is shown in Figure 8a. NPEN reduced the signal drops at slab boundaries and generated a signal profile similar to the ground truth, whereas signal variation along the slice direction was still strong in WA and PEN. A 1D signal profile in the slice direction passing through the left putamen is shown in Figure 8b; NPEN reconstruction provided the best approximation to the ground truth. Figure 8c shows another 1D signal profile in the slice direction passing through the left ventricle, which was chosen as a worst case scenario for NPEN reconstruction. NPEN exhibited large errors around slice 10, slice 40, and slice 58 (Fig. 8c, black arrows), which corresponded to CSF regions; however, in non-CSF regions, NPEN matched the ground truth nicely. In comparison, the WA and PEN results differed more from the ground truth in either CSF region or non-CSF region.

Figure 9 shows results of PEN and NPEN at different TR. It can be seen that the residual artifacts in PEN reconstruction at short TR (2 s) are very strong but can be reduced at longer TRs (4 and 7 s). NPEN can effectively reduce slab boundary artifacts regardless of TR values.

Reconstruction with In Vivo Data

Figure 10 compares the reconstruction methods on in vivo data. As shown in $b=0$ images (Fig. 10a and 10e)




FIG. 7. Tissue-specific comparison of the reconstruction results from simulation data for (a) WM, (b) GM, and (c) CSF. For each tissue type, the upper row shows the reconstruction results, whereas the lower row shows the difference maps between the reconstructed images and the ground truth. RMSE is shown in the upper right-hand corner of the difference map. The difference map is rescaled for better depiction of the error distributions.

and diffusion-weighted images (Fig. 10b), WA and PEN still contained strong slab boundary artifacts, whereas NPEN effectively reduced the artifact level with only minor residual artifacts visible in CSF regions. As shown in the FA maps (Fig. 10c) and color-coded maps of the principal eigenvector (Fig. 10d), WA still had strong artifacts at slab boundaries even after normalization by the $b=0$ image. The PEN result showed reduced artifacts in the FA map compared with WA but retained some residual errors, as indicated by yellow arrows in the zoomed-in FA map (Fig. 10f). Given that the primary goal of PEN was to remove aliasing, this suggests that the effect of aliasing artifacts in WA reconstruction is significant. NPEN achieves the best results, with almost no visible artifact, suggesting that NPEN is effective at removing slab boundary artifacts associated with the RF profile, aliasing and saturation. Reconstructions based on data acquired from a different

subject are shown in Supporting Figure S3, demonstrating similar performance of the different methods.

DISCUSSION

Several methods have been proposed to correct slab boundary artifacts in dMRI, including the weighted average and the PEN methods considered here. Weighted average avoids aliasing at slab boundaries by oversampling along the slab direction. The goal of PEN is to remove aliasing artifacts, with no attempt to address the saturation effects. Thus, PEN works well under the condition that the sum-of-squares image used in slab profile estimation has a relatively flat signal profile. If not, residual signal variation remains in the reconstructed images and can propagate into diffusion parameter maps such as MD and FA. Our NPEN method extends PEN to deal with both aliasing and slab crosstalk, which enables




FIG. 8. (a) Averaged WM signal at each slice. (b) Signal profile along a line passing through left putamen in the slice direction. (c) Signal profile along a line passing through the left ventricle in the slice direction. Black arrows indicate residual slab boundary artefacts in CSF regions.

correction of slab boundary artifacts under conditions of strong T1 saturation due to short TR (1–2 s) associated with optimal SNR efficiency.

The conventional coupling between FOV and TR is lessened in multislab acquisition. Encoding of slices through a combination of slab excitation and within-slab phase encoding confers flexibility in achieving a desired coverage and TR without major impact to scan time. With short TR, we can only accommodate a small number of excitation periods. At longer TR, we can achieve the same coverage by increasing the number of slabs while reducing the within-slab phase encoding, avoiding significant increase in total scan time. Moreover, one could also deploy simultaneous multislab acceleration (54) to increase coverage without requiring more excitations per TR. As such, the need for longer TR to improve reconstruction in PEN and WA does not in itself require longer scan time; and conversely, the key benefit conferred by NPEN in enabling short TR is increased SNR efficiency, not shorter scan times.

Nevertheless, one drawback to all slab boundary corrections is a reduction in scan time efficiency. At its most efficient, the number of k_z phase encodings would simply be equal to the number of reconstructed z slices (although this would require a perfect rectangular slab profile, which cannot be achieved in practice). We can define scan time efficiency in terms of the fractional increase in scan time beyond this idealized case. All slab boundary corrections have reduced scan time efficiency due to additional measurements associated with either 1) deliberate overlapping of adjacent slab FOVs in the transition bands or 2) oversampling of excited slab FOVs into their transition regions. WA uses both overlapping and oversampling, with the original implementation requiring 0.4 times more k_z encodings than reconstructed slices, or 71.4% scan time efficiency (21). Both PEN and NPEN overlap slab FOVs by $\sim 20\%$ and also incur a fixed overhead to acquire calibration data (here, equivalent to acquiring one additional volume). Hence, our protocol with two $b=0$ and 20 diffusion-weighted volumes had a 79.7% scan time efficiency.

Two constraints were applied in NPEN. The in-plane smoothness constraint imposed on the slab profile was used to improve the conditioning of the problem. This constraint seemed reasonable given that Bloch simulations of the slab profiles for WM, GM, and CSF are quite similar and B_0/B_1 profiles vary smoothly over space, although WM and CSF profiles do exhibit consistent differences (Fig. 2e). Nevertheless, the tissue-specific comparison (Fig. 7) suggests the artifacts in WM and GM components are nicely corrected, with greatest residual artifacts being in CSF, which is only of interest in areas of large partial volume. The second constraint was used to suppress the spurious frequency components corresponding to slab boundary artifacts. A similar approach was adopted by Engstrom et al (27), who used a band-pass filter in the estimation of a 1D banding function, which was applied to correct slab boundary artifacts. The main difference between the two approaches is that NPEN allows the spatial variation of the slab profile within each slice plane, capturing the effects caused by different tissue types and B_0 inhomogeneity, whereas the 1D banding function used in the previous method can only provide a global correction. The effects of these constraints on NPEN reconstruction are demonstrated in

FIG. 9. PEN (a) and NPEN (b) reconstruction at different TR.





FIG. 10. Comparison of three reconstruction methods based on in vivo data. (a) $b=0$ image. (b) DWI image ($b=1000 \text{ s/mm}^2$). (c) FA maps. (d) Color-coded maps of the principal eigenvector (red, right-left; green, anterior-posterior; blue, superior-inferior). (e, f) Zoomed-in regions specified by the yellow rectangles in panels a and c, respectively.



Supporting Figure S4. The reconstruction RMSE increases after a certain number of iterations. This likely stems from the overfitting, which may occur due to the high complexity of the nonlinear reconstruction model. In the reconstruction, the algorithm tries to improve the fitting of the unknowns based on the reconstruction model. However, after a certain point, the improvement on the fitting of the unknowns comes at the expense of increased errors, leading to inflated RMSE. As the RMSE curves for WM and GM are different, in some cases it is possible the error in GM decreases while the error in WM increases; however, as shown in Figure 5, this is not a significant effect, which is not clearly visible in our results.

In the PEN method, subject motion could be an outstanding problem as the slab profile estimated from the calibration scan might be misaligned with the images acquired later, especially for DTI where a large number of volumes are required and scan time is long. In comparison, NPEN is a data-driven method, in which the calibration image is only used for initialization and its effect is reduced during the iterative reconstruction, so NPEN is expected to be less sensitive to subject motion.

The long scan times (1 min per volume) used in the data acquisition are dictated by the use of readout segmentation and are the trade-off associated with reduced distortion and T_2^* blurring. It should be noted that acquisition with single-shot per k_z plane (with equivalent distortion and T_2^* blurring to 2D single-shot) can achieve ~ 20 s per volume, which is about 70% longer than a conventional 2D single-shot EPI due to the need for navigator echoes and slab overlap, but would still have about 50% higher SNR efficiency.

A drawback of NPEN is the computational cost. This is in part due to the operations on large matrix and iterative nature of the method. The computation time for all diffusion directions was ~ 1 h using MATLAB (2014a; MathWorks, Natick, Massachusetts, USA) on a distributed computing cluster. The cluster consisted of 19 computers, eight of which had AMD Opteron Processors 6328 (32 cores, 3.2GHz) and 512 GB RAM, the others of which had Intel Xeon Processors E5640 (16 cores, 2.67GHz) and 64 GB RAM. It should be noted that the computation time could be definitely reduced with optimized C/C++ implementation.

The NPEN method was investigated here in the context of diffusion measurements. However, this method may be applicable to other MRI applications using 3D multislabs acquisition, such as 3D fast spin echo imaging and time-of-flight angiography.

CONCLUSIONS

We present a new approach to reduce the slab boundary artifacts in 3D multislabs dMRI. Slab boundary artifacts were evaluated using Bloch simulations for a range of tissue properties and acquisition parameters, with particular interest on the potential for these artifacts to bias estimates of diffusion-based quantification (e.g., FA and MD). The presented reconstruction method, NPEN, can correct both the aliasing artifacts and signal variation caused by slab crosstalk even at short TR (~ 2 s). Recon-

struction results on simulated and in vivo data demonstrate that the proposed method improves upon previous methods, including the original PEN technique.

ACKNOWLEDGMENT

We thank David Porter for continued advice and support with the readout-segmented EPI sequence and Thorsten Feiweier for the diffusion preparation code used in this study.

REFERENCES

- Butts K, Pauly J, De Crespigny A, Moseley M. Isotropic diffusion-weighted and spiral-navigated interleaved EPI for routine imaging of acute stroke. *Magn Reson Med* 1997;38:741–749.
- Atkinson D, Porter Da, Hill DLG, Calamante F, Connelly A. Sampling and reconstruction effects due to motion in diffusion-weighted interleaved echo planar imaging. *Magn Reson Med* 2000;44:101–109.
- Pipe JG, Farthing VG, Forbes KP. Multishot diffusion-weighted FSE using PROPELLER MRI. *Magn Reson Med* 2002;47:42–52.
- Miller KL, Pauly JM. Nonlinear phase correction for navigated diffusion imaging. *Magn Reson Med* 2003;50:343–353.
- Liu C, Bammer R, Kim DH, Moseley ME. Self-navigated interleaved spiral (SNAILS): application to high-resolution diffusion tensor imaging. *Magn Reson Med* 2004;52:1388–1396.
- Holdsworth SJ, Skare S, Newbould RD, Bammer R. Robust GRAPPA-accelerated diffusion-weighted readout-segmented (RS)-EPI. *Magn Reson Med* 2009;62:1629–1640.
- Porter DA, Heidemann RM. High resolution diffusion-weighted imaging using readout-segmented echo-planar imaging, parallel imaging and a two-dimensional navigator-based reacquisition. *Magn Reson Med* 2009;62:468–475.
- Heidemann RM, Anwender A, Feiweier T, Knösche TR, Turner R. k-space and q-space: combining ultra-high spatial and angular resolution in diffusion imaging using ZOOPPA at 7T. *Neuroimage* 2012;60:967–978.
- Frost R, Miller KL, Tijssen RH, Porter DA, Jezzard P. 3D multi-slab diffusion-weighted readout-segmented EPI with real-time cardiac-reordered K-space acquisition. *Magn Reson Med* 2014;72:1565–1579.
- Jeong EK, Kim SE, Parker DL. High-resolution diffusion-weighted 3D MRI, using diffusion-weighted driven-equilibrium (DW-DE) and multishot segmented 3D-SSFP without navigator echoes. *Magn Reson Med* 2003;50:821–829.
- Golay X, Jiang H, van Zijl PC, Mori S. High-resolution isotropic 3D diffusion tensor imaging of the human brain. *Magn Reson Med* 2002;47:837–843.
- Wang JJ, Deichmann R, Turner R, Ordidge R. 3D DT-MRI using a reduced-FOV approach and saturation pulses. *Magn Reson Med* 2004;51:853–857.
- Jeong EK, Kim SE, Kholmovski EG, Parker DL. High-resolution DTI of a localized volume using 3D single-shot diffusion-weighted stimulated echo-planar imaging (3D ss-DWSTEPI). *Magn Reson Med* 2006;56:1173–1181.
- Jung Y, Samsonov AA, Block WF, Lazar M, Lu A, Liu J, Alexander AL. 3D diffusion tensor MRI with isotropic resolution using a steady-state radial acquisition. *J Magn Reson Imaging* 2009;29:1175–1184.
- Van AT, Hernando D, Sutton BP. Motion-induced phase error estimation and correction in 3D diffusion tensor imaging. *IEEE Trans Med Imaging* 2011;30:1933–1940.
- McNab JA, Gallichan D, Miller KL. 3D steady-state diffusion-weighted imaging with trajectory using radially batched internal navigator echoes (TURBINE). *Magn Reson Med* 2010;63:235–242.
- O'Halloran R, Aksoy M, Van A, Bammer R. 3D isotropic high-resolution diffusion-weighted MRI of the whole brain with a motion-corrected steady-state free precession sequence. *Magn Reson Med* 2013;70:466–478.
- Enzmann D, Pelc N. Brain motion: measurement with phase-contrast MR imaging. *Radiology* 1992;185:653–660.
- Greitz D, Wirestam R, Franck A, Nordell B, Thomsen C, Ståhlberg F. Pulsatile brain movement and associated hydrodynamics studied by magnetic resonance phase imaging. *Neuroradiology* 1992;34:370–380.

20. Poncet B, Wedeen V, Weisskoff R, Cohen M. Brain parenchyma motion: measurement with cine echo-planar MR imaging. *Radiology* 1992;185:645–651.
21. Engstrom M, Skare S. Diffusion-weighted 3D multislabs echo planar imaging for high signal-to-noise ratio efficiency and isotropic image resolution. *Magn Reson Med* 2013;70:1507–1514.
22. Frank LR, Jung Y, Inati S, Tyszka JM, Wong EC. High efficiency, low distortion 3D diffusion tensor imaging with variable density spiral fast spin echoes (3D DW VDS RARE). *NeuroImage* 2010;49:1510–1523.
23. Parker DL, Yuan C, Blatter DD. MR angiography by multiple thin slab 3D acquisition. *Magn Reson Med* 1991;17:434–451.
24. Oshio K, Jolesz FA, Melki PS, Mulkern RV. T2-weighted thin-section imaging with the multislabs three-dimensional RARE technique. *J Magn Reson Imaging* 1991;1:695–700.
25. Murakami JW, Weinberger E, Tsuruda JS, Mitchell JD, Yuan C. Multislabs three-dimensional T2-weighted fast spin-echo imaging of the hippocampus: sequence optimization. *J Magn Reson Imaging* 1995;5:309–315.
26. Chang H-C, Sundman M, Petit L, Guhaniyogi S, Chu M-L, Petty C, Song AW, Chen N-k. Human brain diffusion tensor imaging at sub-millimeter isotropic resolution on a 3 Tesla clinical MRI scanner. *NeuroImage* 2015;118:667–675.
27. Engstrom M, Martensson M, Avventi E, Skare S. On the signal-to-noise ratio efficiency and slab-banding artifacts in three-dimensional multislabs diffusion-weighted echo-planar imaging. *Magn Reson Med* 2015;73:718–725.
28. Liu K, Rutt BK. Sliding interleaved ky (SLINKY) acquisition: a novel 3D MRA technique with suppressed slab boundary artifact. *J Magn Reson Imaging* 1998;8:903–911.
29. Liu K, Xu Y, Loncar M. Reduced slab boundary artifact in multi-slab 3D fast spin-echo imaging. *Magn Reson Med* 2000;44:269–276.
30. Parker DL, Roberts JA, Alexander AL, Goodrich KC, Tsuruda J. Magnetic resonance angiography with sliding interleaved projection reconstruction (SLIPR) acquisition. *J Magn Reson Imaging* 1999;10:569–575.
31. Kwon KT, Kerr AB, Wu HH, Hu BS, Brittain JH, Nishimura DG. Non-contrast-enhanced peripheral angiography using a sliding interleaved cylinder acquisition. *Magn Reson Med* 2015;74:727–738.
32. Li Z, Wang D, Robison RK, Zwart NR, Schar M, Karis JP, Pipe JG. Sliding-slab three-dimensional TSE imaging with a spiral-in/out readout. *Magn Reson Med* 2015. doi: 10.1002/mrm.25660.
33. Kholmovski EG, Alexander AL, Parker DL. Correction of slab boundary artifact using histogram matching. *J Magn Reson Imaging* 2002;15:610–617.
34. Van AT, Aksoy M, Holdsworth SJ, Kopeinigg D, Vos SB, Bammer R. Slab profile encoding (PEN) for minimizing slab boundary artifact in three-dimensional diffusion-weighted multislabs acquisition. *Magn Reson Med* 2015;73:605–613.
35. Pruessmann KP, Weiger M, Scheidegger MB, Boesiger P. SENSE: sensitivity encoding for fast MRI. *Magn Reson Med* 1999;42:952–962.
36. Pauly J, Le Roux P, Nishimura D, Macovski A. Parameter relations for the Shinnar-Le Roux selective excitation pulse design algorithm. *IEEE Trans Med Imaging* 1991;10:53–65.
37. Wansapura JP, Holland SK, Dunn RS, Ball WS. NMR relaxation times in the human brain at 3.0 Tesla. *J Magn Reson Imaging* 1999;9:531–538.
38. Gelman N, Ewing JR, Gorell JM, Spickler EM, Solomon EG. Interregional variation of longitudinal relaxation rates in human brain at 3.0 T: relation to estimated iron and water contents. *Magn Reson Med* 2001;45:71–79.
39. Cox EF, Gowland PA. Simultaneous quantification of T2 and T2* using a combined gradient echo-spin echo sequence at ultrahigh field. *Magn Reson Med* 2010;64:1441–1446.
40. Alexander AL, Hasan KM, Lazar M, Tsuruda JS, Parker DL. Analysis of partial volume effects in diffusion-tensor MRI. *Magn Reson Med* 2001;45:770–780.
41. Uecker M, Hohage T, Block KT, Frahm J. Image reconstruction by regularized nonlinear inversion—joint estimation of coil sensitivities and image content. *Magn Reson Med* 2008;60:674–682.
42. Uecker M, Zhang S, Voit D, Karaus A, Merboldt KD, Frahm J. Real-time MRI at a resolution of 20 ms. *NMR Biomed* 2010;23:986–994.
43. Knoll F, Clason C, Bredies K, Uecker M, Stollberger R. Parallel imaging with nonlinear reconstruction using variational penalties. *Magn Reson Med* 2012;67:34–41.
44. Xu B, Spincemaille P, Chen G, Agrawal M, Nguyen TD, Prince MR, Wang Y. Fast 3D contrast enhanced MRI of the liver using temporal resolution acceleration with constrained evolution reconstruction. *Magn Reson Med* 2013;69:370–381.
45. Engl HW, Hanke M, Neubauer A. Regularization of inverse problems. Vol. 375 of *Mathematics and its Applications*. Dordrecht, Netherlands: Kluwer Academic Publishers; 1996.
46. Hohage T. Logarithmic convergence rates of the iteratively regularized Gauss-Newton method for an inverse potential and an inverse scattering problem. *Inverse problems* 1997;13:1279.
47. Bakushinsky AB, Kokurin MY. Iterative methods for approximate solution of inverse problems. Vol. 577 *Mathematics and its Applications*. Dordrecht, Netherlands: Springer; 2004.
48. Aubert-Broche B, Evans AC, Collins L. A new improved version of the realistic digital brain phantom. *NeuroImage* 2006;32:138–145.
49. Zhang Y, Brady M, Smith S. Segmentation of brain MR images through a hidden Markov random field model and the expectation-maximization algorithm. *IEEE Trans Med Imaging* 2001;20:45–57.
50. Robson MD, Anderson AW, Gore JC. Diffusion-weighted multiple shot echo planar imaging of humans without navigation. *Magn Reson Med* 1997;38:82–88.
51. Frost R, Porter DA, Miller KL, Jezzard P. Implementation and assessment of diffusion-weighted partial Fourier readout-segmented echo-planar imaging. *Magn Reson Med* 2012;68:441–451.
52. Smith SM. Fast robust automated brain extraction. *Hum Brain Mapp* 2002;17:143–155.
53. Smith SM, Jenkinson M, Woolrich MW, et al. Advances in functional and structural MR image analysis and implementation as FSL. *NeuroImage* 2004;23:S208–S219.
54. Frost R, Jezzard P, Porter DA, Tijssen RHN, Miller KL. Simultaneous Multi-slab Acquisition in 3D Multi-slab Diffusion-Weighted Readout-Segmented Echo-Planar Imaging. In *Proceedings of the 21st Annual Meeting of ISMRM*, Salt Lake City, Utah, USA, 2013. Abstract 3176.

SUPPORTING INFORMATION

Additional Supporting Information may be found in the online version of this article.

Supporting Figure S1. (a, b) Waveforms of the excitation pulse (a) and refocusing pulse (b) used in the simulations and in vivo scans. (c) Simulated signal profile with one slab (TR = 2 s).

Supporting Figure S2. WM signal profile of one slab simulated at different off-resonance frequencies.

Supporting Figure S3. Coronal slice of the reconstruction results based on the data acquired from a different subject. (a) b = 0 image. (b) DWI image (b = 1000 s/mm²). (c) FA maps. (d) Color-coded maps of the principal eigenvector (red, right-left; green, anterior-posterior; blue, superior-inferior). (e, f) Zoomed-in regions specified by the yellow rectangles in panels a and c, respectively.

Supporting Figure S4. Reconstruction results using iteratively regularized Gauss-Newton method. (a) Without additional constraints. (b) With in-plane smoothness constraints on slab profile. (c) With frequency constraints on image. (d) With both in-plane smoothness constraints on slab profile and frequency constraints on image. All other reconstruction parameters (e.g., number of iterations, regularization parameters) are the same in the reconstructions.